

Scientific Notebook No. 464: Proposed  
Experiment to Understand the Deliquescence  
Behavior of Salt Mixtures from Yucca  
Mountain Waters (07/12/2002 through  
04/07/2006)

# CENTER FOR NUCLEAR WASTE REGULATORY ANALYSES

Salt

Experiments

Book # 464

---

CNWRA  
CONTROLLED  
COPY 464

Project # 20.1402.561

Justin Landry Justin Landry JPL

L: etai yang L: etai yang L: y.

(1/2)

20-01-2002 of dn-1997

476 files



476 files

(1/2)

Notebook 464

SAIT - upto - 2002

10-02

Project # 20.1402.561

Initial Entry

**Proposed Experiment to Understand the Deliquescence Behavior of the Salt Mixtures From YM Waters**

April 19, 2001

All data files are stored in  
NoteBooks / Notebook 464-Salt, J. V. J. 11/17/03

**BACKGROUND**

Critical relative humidity (CRH) is an important parameter in predicting the life of the WP or DS. If the drift humidity is lower than the CRH, only dry air oxidation occurs; if the drift humidity is higher than the CRH, aqueous corrosion would occur. According to DOE's analysis, NaCl, NaNO<sub>3</sub> and KNO<sub>3</sub> would be the most dominant salts present in the evaporated water in the drift and, within the temperature range predicted for the drift, NaNO<sub>3</sub> is the most hygroscopic salt and has the lowest deliquescence relative humidity (50%). Therefore, DOE has considered 50% to be the CRH. This consideration may not be valid because it is based on the properties of pure salts. A mixture of the salts would have an even lower relative humidity. For instance, the hand book value for the NaCl-NaNO<sub>3</sub>-KNO<sub>3</sub> saturated solution is only 30% at 16.4°C. If this is true for all the temperature ranges predicted for the drift, the dry air oxidation period would be significantly shorter and the aqueous corrosion period would be significantly higher than what the DOE has considered. Therefore, there is a need to determine the deliquescence behavior of the mixture of salts that are expected to be dominant in the drift.

**OBJECTIVES**

Experimental measurement will be conducted in this study to determine the deliquescence humidity of the mixtures of the dominant salts that would be present in the drift after evaporation. This work will also provide useful data for the modeling analysis for more complicated salts

**EXPERIMENT**

**Methods:**

A gravimetric method is a widely used method for studying the deliquescence behavior of salts. However, the gravimetric method requires a microbalance which is not available at CNWRA at the present time. An electrochemical method will be used in this study. Figure 1 shows the proposed experimental set-up. The experiment will be conducted in a humidity chamber and a conductivity probe will be employed to determine the deliquescence humidity. Figure 2 shows the modeling result for the relative humidity of saturated NaCl-NaNO<sub>3</sub> mixtures. Figure 3 shows the modeling results of the electrical conductivity and the weight of a NaCl-NaNO<sub>3</sub> composition (NaCl fraction being 0.3061) at different relative humidities. Clearly both the gravimetric and the conductivity methods are highly sensitive to the changes of relative humidity near the mutual relative humidity (0.6646, see also Figure 2). At the mutual relative humidity, the changes are from 756g to 1562g for weight and from infinity to 0.2323 1/cm-ohm for conductivity respectively. Figure 3 also shows that both the weight vs RH and the conductivity vs RH curves can be used to determine the saturation relative humidity (0.6730, also see Figure 2).

The conductivity will be measured using a Solartron AC Impedance System in conjunction with an in-house developed Visual Basic software. All equipment and software are presently available in the corrosion laboratory at CNWRA.

J. V. J. 7/12/02

**Test Conditions:**

Subject to the limitations of the humidity chamber in the corrosion laboratory, The test ranges will be from 20% to 90% for relative humidity and 45 to 75°C for temperature.


**Salts and Salt Mixtures:**

Initial tests will be conducted with several pure salts of known relative humidities for the purpose of method calibration. These salts include: NaCl, KNO<sub>3</sub>, CaCl<sub>2</sub> and MgCl<sub>2</sub>


Upon completion of the calibration tests, effort will focused on the NaCl-NaNO<sub>3</sub>-KNO<sub>3</sub> system. The effects of Ca, HCO<sub>3</sub>, and MgCl<sub>2</sub> salts may also be added to the test matrix if resource is available.

**Schedule:**

May through August, 2001



*J. Yang*  
05/24/01



**Figure 1** Proposed experimental setup to measure the deliquescence relative humidity of salt mixtures

*J. Yang* 05/24/01

Equipment.

Conductivity measurement:

Solartron Electrochemical system

~~1286~~ + 1280.  
1287 J. Yang. 10/21/01

Software: Zplot 2.0

Humidity chamber: see page 35.

J. Yang  
7/15/02

~~Electrochemical cell.~~ J. Yang. 06/12/01.

Software:


Solartron 2.0 Zplot only allows manual operation. That is, the measurement can only be made for one humidity point (varying frequency).

We need to make multiple or continuous measurement to detect the change in  $|Z|$  as a function of Relative humidity.

An Automation Program: "AUTO.ZPLOT" is written to automatically start the Zplot program at a predetermined time interval. The automation program also save the data obtained after each run with Zplot into a file. It also display the  $|Z|$  calculated from  $\sqrt{Z_{real}^2 + Z_{im}^2}$  on the screen. The  $|Z|$  values for different frequencies are also saved into another file.

J. Yang 06/12/01


Documentation of AutoZPlot.




This is the splash screen. Codes associated with splash:

```
frmSplash - 1
Private Sub Form_Load()
  lblVersion.Caption = "Version " & App.Major & "." & App.Minor & "." & App.Revision
  lblProductName.Caption = App.Title
End Sub
```


J. Y. 7/12/02



This is the about screen, displayed after clicking the help manual. Codes are not complete for the help, and therefore, the codes are not documented in this book. J.Y. 06/12/01



J.Y. 06/12/01 J. Y. 2/14/02



This is the main form. clicking the

configuration button takes the user to the

configuration form. (see page 10)

Run button takes the user to the Run form (page 12)

J. Jones  
2/12/02

### Codes For Main Form.

```
frmMain - 1

Private Declare Function OSWinHelp% Lib "user32" Alias "WinHelpA" (ByVal hwnd%, ByVal HelpFiles,
ByVal wCommand%, dwData As Any)

Private Sub cmdEISConfiguration_Click()
    If multiOption.Value = True Then
        frmConfigAC.Show
    Else
        frmConfigAC1.Show
    End If
    Me.Hide
End Sub

Private Sub cmdEISRun_Click()
    If (Not ConfigACDone) Then
        MsgBox "Please Setup the Configuration First"
        Exit Sub
    End If

    Me.Hide
    If multiOption.Value = True Then
        frmRunAC.Show
    Else
        frmRunAC1.Show
    End If

End Sub

Private Sub cmdExit_Click()
    Me.Hide
End Sub

Private Sub Form_Initialize()
    ConfigACDone = False
End Sub

Private Sub Form_Load()
    Me.Left = GetSetting(App.Title, "Settings", "MainLeft", 500)
    Me.Top = GetSetting(App.Title, "Settings", "MainTop", 500)
    Me.Width = GetSetting(App.Title, "Settings", "MainWidth", 4000)
    Me.Height = GetSetting(App.Title, "Settings", "MainHeight", 4000)
End Sub

Private Sub Form_Unload(Cancel As Integer)
    Dim i As Integer
    'close all sub forms
    For i = Forms.Count - 1 To 1 Step -1
        Unload Forms(i)
    Next
    If Me.WindowState <> vbMinimized Then
        SaveSetting App.Title, "Settings", "MainLeft", Me.Left
        SaveSetting App.Title, "Settings", "MainTop", Me.Top
        SaveSetting App.Title, "Settings", "MainWidth", Me.Width
        SaveSetting App.Title, "Settings", "MainHeight", Me.Height
    End If
End Sub

Private Sub mnuHelpAbout_Click()
    frmAbout.Show vbModal, Me
End Sub

Private Sub mnuViewOptions_Click()
    frmOptions.Show vbModal, Me
End Sub

Private Sub mnuViewStatusBar_Click()
    If mnuViewStatusBar.Checked Then
        sbStatusBar.Visible = False
        mnuViewStatusBar.Checked = False
    Else
        sbStatusBar.Visible = True
        mnuViewStatusBar.Checked = True
    End If
End Sub
```

There are other codes that are not related to the data functions.  
They are omitted.

J. Jones 06/12/01

J. Jones 2/14/02

Configuration Form.

Cycle Time is the time interval between each measurement -  
2 ent.

Plotting Parameters Block Size is maximum number of  
measured data shown on  
plot. (see the Run Form)  
page 12

Min & Max frequency gives the frequency  
window to be shown. (page 12)

Z-plot Setup file must be given before the Z-plot  
can be run.

J. Yang 7/15/02

Codes

for Configuration

```
frmConfigAC1 - 1
Dim ScanIntAC1 As Integer ' not to the the ScanInt of the Global, too much headache

Private Sub ACSetup_Click()
    cdlGetACSetupFile.ShowOpen
    txtZplotSetup.Text = cdlGetACSetupFile.filename
    'txtConFile.Text = "C:\Yang\last.cfg"
End Sub

Private Sub cmdBack_Click()
    frmConfigAC1.Hide
    frmMain.Show
    ConfigACDone = True
End Sub

Private Sub cmdDataFile_Click()
    cdlGetDataFile.ShowOpen
    txtDataFile.Text = cdlGetDataFile.filename
End Sub

Function TimeToSec(asciiTIMES)
    '**** convert time-ascii to integer-seconds ****
    'DIM hrs, mint, sec
    TimeToSec = Val(asciiTIMES) * 3600 + Val(Mid$(asciiTIMES, 4, 2)) * 60 + Val(Mid$(asciiTIMES, 7, 2))
End Function

Private Sub Form_Initialize()
    'for ScanInt initialization
    Call txtScanInt_Change
End Sub
    ' these Subs are dealing with the UPDown Properties for Counts

Private Sub Form_Load()
    FileName0$ = txtDataFile.Text
End Sub

Private Sub optHour_Click()
    Call txtScanInt_Change
End Sub

Private Sub optMin_Click()
    Call txtScanInt_Change
End Sub

Private Sub optSec_Click()
    Call txtScanInt_Change
End Sub

Private Sub txtScanInt_Change()
    If Val(txtScanInt.Text) > 60 Or Val(txtScanInt.Text) = 0 Then
        MsgBox "the value is too big!", vbCritical, "Re-type"
        Exit Sub
    End If
    If optSec.Value = True Then
        ScanIntAC1 = txtScanInt.Text
    ElseIf optMin.Value = True Then
        ScanIntAC1 = 60 * txtScanInt.Text
    Else
        ScanIntAC1 = 3600 * txtScanInt.Text
    End If
    ScanIntAC_Global = ScanIntAC1
End Sub

Private Sub txtDataFile_Change()
    FileName0$ = txtDataFile.Text 'Initial filename to store data

    'FileName0$ = "C:\yang\data"
    DataFileTemp = txtDataFile.Text 'Base filename to store temporary data
End Sub
```

J. Yang  
7/15/02



AutoZPlot Run Form.

AC Test

Program Start Time: 6/12/01 3:11:27 PM

Number of Cycles Completed:

Quit

Current Cycle

Data File Name:

Start Time:

Status 6/12/01 3:12 PM

This the form displayed during the program run.

3-D plot:

Z — the impedance data  $\log|Z|$

Y — the frequency, from Min to Max

X — Data measured at different time

The plot will be refreshed when the number of measurement reaches the "Block size"

J. Y. 7/12/02

Codes for Run Form

```
frmRunAC1 - 1

Dim quitflag As Boolean
Dim loopFinished As Boolean
Public ZplotObject As Object ' can also be declared in a module
Dim firstChannel As Boolean
Dim GotoQuitAfterSweep As Boolean
Dim ZplotFinished As Boolean
Dim F(25) As Double 'keep track of frequency
Dim Z(25) As Double 'Keep track of Impedance
Dim ZCycleN As Integer
Dim block As String
Dim DataFileTemp As String 'File to temporary store the Zplot data for one cycle
Dim FileName1 As String ' File Name for data, with no extension name
Dim FileNamePData ' File Name for Processed Data
Dim newFile As Boolean 'track if the data is after the initial loop, need a heading
Dim ScanIntAC1 As Integer ' not to the the ScanInt of the Global, too much headache

Sub LoadZplot()
    Dim IsOK As Boolean
    Dim gs As String
    Dim gb As Boolean

    'If the ISOK variable is "true", the function was succesful
    'If the ISOK variable is "False", the function failed

    'Load ZPlot
    Set ZplotObject = CreateObject("ZPlot.ZPlotApp") 'Windows knows where to start ZPlot
    gs = frmConfigAC1.txtZplotSetup.Text 'load the setup file for Zplot
    IsOK = ZplotObject.OpenSetup(gs)

    'self added:
    'I should get the data points and save my own files
    'I should use ZplotObject.GetPoint(PointNo: Int) VariantArray to get the AC imp. data
    'no need, just do the cycle then pull the data out after the cycle

End Sub

Sub ZplotRun()
    Dim IsOK As Boolean
    Dim gs As String
    Dim gb As Boolean

    'If the ISOK variable is "true", the function was succesful
    'If the ISOK variable is "False", the function failed
    'Open a Setup File
    gs = txtSetupFile.Text ' to be set set such as "mysetup.zpw"

    ' Specify a data file for the experiment
    gs = txtDataFile.Text ' to be set set such as "mydata.z"

    DataFileTemp = FileName1$ + "Temp"
    gs = DataFileTemp 'temporary file name to store the Z plot data for one curve (one channel)
    gb = False 'append condition, True=append, False=overwrite

    ZplotFinished = False

    IsOK = ZplotObject.SetDataFile(gs, gb)

    'Clear the comment text, and add new text
    IsOK = ZplotObject.SetCommentClear
    gs = "comment1" 'to identify channels
    ZplotObject.SetCommentAdd (gs)
    gs = "comment2"
    ZplotObject.SetCommentAdd (gs)

    ' Start the sweep

    IsOK = ZplotObject.StartSweep

    PR = -1
```

frmConfigAC1.

J. Y. 06/12/01

J. Y. 7/12/02

```

frmRunAC1 - 2

While PR <> 0
  DoEvents 'must have a mechanism to prevent timer call the Mainloop
  PR = ZplotObject.GetPointsRemaining
Wend
' MsgBox "After finished PR= " + Str(PR)
ZplotFinished = True

End Sub

' the following should be in a loop

Private Sub cmdCheck_Click()
  Dim PR As Integer
  'Check number of point to be measured
  PR = ZplotObject.GetPointsRemaining
  If PR = -1 Then
    'The instrument are being initialized to sweep
  End If
  If PR = 0 Then
    'The experiment is done
  End If
End Sub

Sub MainLoopInitialize()
  ZCycleN = 0
  txtProgStartTime.Text = Str(Date) + " " + Str(Time)
  quitflag = False

End Sub
Sub MainloopFinish()
  Close #3 'close data file
End Sub
Sub MainLoop()
  '**** MAIN LOOP ****
  ' to be fired by the timer tmrScanCycle
  'Check if this cycle data should be saved to the new sub file.
  'When the block number has been reached, data will be saved into a new file and head
rs for the file will be created
  txtCycleStartTime.Text = Str(Date) + " " + Str(Time)

  '**** MAIN LOOP ****
  ' to be fired by the timer tmrScanCycle
  'Check if this cycle data should be saved to the new sub file.
  'When the block number has been reached, data will be saved into a new file and head
rs for the file will be created

  If quitflag = True Then
    Call MainloopFinish
    Unload Me
    frmMain.Show
  End If

  'pass no Loop Scan No.

  time1 = Time

  Call BlockCheck_for_NewFile(ZCycleN)
  loopFinished = False
  toExitAfterSeep = False
  txtFileName.Text = filename$
  time1 = Now
  GotoQuitAfterSweep = False 'before the start of sweep
  ZCycleN = ZCycleN + 1
  Call ZplotRun
  time2 = Now

  Call Save_N_PlotData(time1, time2)
  txtCycleCompleted.Text = ZCycleN

```

*J. Yang*  
7/15/02

```

frmRunAC1 - 3

If GotoQuitAfterSweep = True Then
  Call cmdQuit_Click
End If
loopFinished = True
'**** end of main loop
End Sub

Sub Save_N_PlotData(time1, time2)

  Open DataFileTemp For Input As #6

  Open filename$ For Append As #3
  tim1 = Mid(Str(time1), 9, 8)
  tim2 = Mid(Str(time2), 9, 8)
  I1 = 1
  jj = 0
  Do Until EOF(6)
    Line Input #6, temp$
    If ZCycleN = 1 Then
      'For begining of file write everything in the Zplot file
      Print #3, " "; " "; " "; " "; " "; temp$
    End If

    'write only the numbers. Note the first 2 lines after "Frequency in the Zplot file will
be ignored

    'test the first three character and see if they are numbers
    If (Val(Mid(temp$, 1, 3)) <> 0) Then
      jj = jj + 1
      If jj >= 2 Then
        'ignore the first 1 lines which are non-zero numbers
        If ZCycleN <> 1 Then
          Print #3, tim1; tim2; temp$
        End If
        'to extract the values for f and Z to plot
        Call ExtractZ(temp$, Fr, Za, Zb)
        jj = jj + 1

        If Fr >= Val(frmConfigAC1.txtMinFreq.Text) Then
          If Fr <= Val(frmConfigAC1.txtMaxFreq.Text) Then
            F(I1) = Fr
            Z(I1) = (Za ^ 2 + Zb ^ 2) ^ 0.5
            I1 = I1 + 1
          End If
        End If
      End If
    Else
      'ignor all descriptive lines
    End If

  Loop
  ' Update Processed Data File
  Print #3, "" ' change the line at the end
  Close #6
  Close #3

  If newFile = True Then 'write the frequency headings
    Open FileNamePData For Append As #9
    Print #9, "Date "; " "; "Time Started "; " ";
    For L = 1 To I1 - 1
      Print #9, F(L); " ";
    Next L
    Print #9, "" 'change line
    Close #9
    newFile = False
  End If

  'Printing the processed data to FileNamePData
  Open FileNamePData For Append As #9
  Print #9, Date$; " "; tim1; " ";

```

*J. Yang*  
7/15/02

```

frmRunAC1 - 4

For L = 1 To I1 - 1
    Print #9, Z(L); ", ";
Next L
Print #9, "" 'change line
Close #9

'Update plotting

With crt3DLine
    'Displays a 3D chart with 8 columns and 8 rows
    'data.

    ZBlock = frmConfigAC1.txtZBlockSize.Text

    'initialize the chart if the ZCycleN=0, first started
    If (ZCycleN = 1) Then 'refresh all data
        .ColumnCount = I1 - 1 'every block, start again
        .RowCount = ZBlock
        For R = 1 To ZBlock
            .Row = R
            For C = 1 To .ColumnCount
                .Column = C
                .Data = 0
            Next C
        Next R
    End If

    R = (ZCycleN - 1) Mod ZBlock + 1
    .Row = R 'make sure scale back after reaching block
    For M = 1 To I1 - 1
        .Column = M 'Plot the Frequency
        .Data = Log(Z(M)) / 2.303 'plot the Impedance
    Next M

    'clear zero for front data
    If R < ZBlock - 1 Then
        .Row = R + 1 'make sure scale back after reaching block
        For M = 1 To I1 - 1
            .Column = M 'Plot the Frequency
            .Data = 0 'plot the Impedance
        Next M
    End If
End With

End Sub

Sub ExtractZ(buffer$, Fr, Za, Zb)
    i = 0
    j = 0
    Flag$ = ""
    Size = Len(buffer$)
    'Scan until we run out of characters, or we hit the delimiter
    Do While (i < Size) And (j <= 6) 'until reaching end
        Do While (Flag$ <> ",") 'until hit ","
            i = i + 1
            Flag$ = Mid(buffer$, i, 1) 'return 1 char from ith position
        Loop
        ' found a ","
        j = j + 1
        Flag$ = 0 'reset flag
        If j = 1 Then
            FreqLen = i - 1
        End If
        If j = 4 Then
            ZaStart = i + 1
        End If
        If j = 5 Then
            ZaLen = i - ZaStart
            ZbStart = i + 1

```

*J. Yarns*  
7/18/02

```

frmRunAC1 - 5

End If
If j = 6 Then
    ZbLen = i - ZbStart
End If

Loop
'Calculate the parameters. Ignore case for VarName$. Ensure there was
' actually data read
If Size > 0 Then
    FValue = Mid(buffer$, 1, FreqLen)
    ZaValue = Mid(buffer$, ZaStart, ZaLen)
    ZbValue = Mid(buffer$, ZbStart, ZbLen)
Else
    'Blank line; ignore it
    ZaValue = ""
    ZbValue = ""
End If
Fr = Val(FValue)
Za = Val(ZaValue)
Zb = Val(ZbValue)
End Sub

'Call the plotting program

' Asking user which frequency range to use for averaging

End Sub

Sub BlockCheck_for_NewFile(ZCycleN):
    blockSize = Val(frmConfigAC1.txtBlockSize.Text)
    If (ZCycleN Mod blockSize = 0) Then
        newFile = True 'tracking whether file is new for writing the frequency headings
        block0 = Int(ZCycleN / blockSize) 'get the block number
        block$ = Str(block0) 'convert to text
        If block0 < 10 Then
            Mid$(block$, 1, 1) = "0" 'change the space to "0"
        Else
            block$ = Right$(block$, Len(block$) - 1) 'leave out the space
        End If

        periodPosition% = InStr(fileName0$, ".") 'position of period in fileName$ to insert "x"
        If periodPosition% <> 0 Then 'If period found
            Position% = periodPosition% - 1
            FileName1$ = Left(fileName0$, Position%) 'chop the extension name
            lengthExtention = Len(fileName0$) - periodPosition%
            FileName2$ = Right(fileName0$, lengthExtention) 'chop the extension name
            filename$ = FileName1$ + block$ + "." + FileName2$

        Else
            FileName1$ = fileName0$
            filename$ = fileName0$ + block$ + ".txt"
        End If
        FileNamePData = FileName1$ + block$ + "PData.txt"
        Close #3

        'Test if the Data File name is exist
        'This section is added May 24, 2001 L.Y.

        On Error Resume Next
        Err.Clear
        Open filename$ For Input As #13

        If Err.Number = 0 Then
            'there is no error, file already exist
            strMsg = "Data File already exist. Do you want to replace it? "
            Response = MsgBox(strMsg, vbYesNo, "Caution")
            If Response = vbNo Then
                End
            End If
        End If

        'Resume error tracking
        On Error GoTo 0

```

*J. Yarns*  
7/18/02

```

frmRunAC1 - 6
    Close #13
    Call DataFileHeading 'setup heading in the data file
Else 'not to start new file
    Close #3 'each time close the file then open it, to be safe with data
    Open filename$ For Append As #3
End If
txtFileName.Text = filename$
End Sub

Sub DataFileHeading()
'**** setup data file header ****
Open filename$ For Output As #3
Print #3, Date$
' Any Comments can be placed here

'Print Out Configuration Infor
Print #3, "Data File:          "; Chr$(9); filename$
Print #3, "Zplot Setup File:   "; Chr$(9); frmConfigAC1.txtZplotSetup.Text
Print #3, "Scan Interval:      "; Chr$(9); ScanIntAC1
Print #3, " " ' Space; line
Print #3, "Start T"; Chr$(9); 'print the same thing to indet the headings with the contents
Print #3, "End T"; Chr$(9);
Print #3, "Z Plot Infor"; Chr$(9) 'the string from Z plot
Close #3

'For the Processed Data
'**** setup data file header ****
Open FileNamePData For Output As #9
Print #9, Date$
' Any Comments can be placed here

'Print Out Configuration Infor
Print #9, "Data File:          "; Chr$(9); FileNamePData
Print #9, "Zplot Setup File:   "; Chr$(9); frmConfigAC1.txtZplotSetup.Text
Print #9, "Scan Interval:      "; Chr$(9); ScanIntAC1
Close #9

End Sub

Private Sub Form_Load()
'initialize the interval
ScanIntAC1 = ScanIntAC_Global

'set channels of instruments to channels in he program
Call Initial_Channels

'set the timer
tmrScanInt.Enabled = True
If ScanIntAC1 <= 60 Then
    tmrScanInt.Interval = ScanIntAC1 * 1000
Else
    tmrScanInt.Interval = 60000
End If

TimerReady = True 'for first measure, set timer at ready state
Call MainLoopInitialize 'start the measure loop initilaization
End Sub

Private Sub tmrScanInt_Timer()
Static ELTime
II = Int(ScanIntAC1 / 60)
If ScanIntAC1 > 3600 Then
    II = ScanIntAC1 / 3600
End If
If ELTime >= II Then
    Beep
    MsgBox "time is up"
    'Time is ready, do the measurement
    If loopFinished = True Then

```

*J. Jans*  
2/18/02

```

frmRunAC1 - 7

' make sure Mainloop not to be called even if the time is up
Call MainLoop
End If
EITime = 0
Else
    ELTime = ELTime + 1
End If
End Sub

Private Sub cmdQuit_Click()
'*****
quitflag = True

If loopFinished = True Then

'Disposing of the object closes ZPlot
Set ZplotObject = Nothing
Call MainloopFinish
! "
    MsgBox "Waiting for the Logging Cycle to Finish and Data to be Saved Please Wait

'cmdStart.Enabled = True
Unload Me
frmMain.Show

'End
Else
    MsgBox "Please Wait for Zplot Sweep to Finish?", vbCritical
    Response = MsgBox("Do you want to quit without waiting to finish all channels?", vbYesNo,
"Channel Loop in Progress")

If Response = vbYes Then
    If ZplotFinished = False Then
        Response = MsgBox("Do you want to Wait for Zplot to finish current sweep?", vbYes
No, "Zplot Sweep (for a Channel) in Progress")

If Response = vbYes Then
        wait for the current sweep to finish
        GotoQuitAfterSweep = True
        loopFinished = True 'Assume all channels has been done, ready to exit at ne
xt call
        Exit Sub ' go back to finish sweep and Save Data then call this sub
    End If
End If
'Disposing of the object closes ZPlot
Set ZplotObject = Nothing

Call MainloopFinish
' MsgBox "Waiting for the Logging Cycle to Finish and Data to be Saved Please Wait!

' frmMain.Show
Unload Me
'cmdStart.Enabled = True
frmMain.Show
End

Else
'Continu to measure
End If
End If

End Sub

Private Sub tmrStart_Timer()
'This timer delay the start of the Mainloop just 0.25 Sec
loopFinished = True 'set this to let tmrScanInt pass
tmrStart.Enabled = False
'start the first cycle without waiting

```

*J. Jans*  
2/18/02

```

frmRunAC1 - 8
  Call LoadZplot
  loopFinished = True 'Allow timer to call the ManiLoop for first cycle
  Call MainLoop 'start the first cycle without waiting
End Sub

```

## Main module for Global variables.

```

mdlMain - 1
Global ConfigACDone As Boolean
Global fMainForm As frmMain
Global LpNum As Integer
Global PBlock As Integer          'number of loops in one cycle
                                   'number of Cycles of data to save
Global buffer$                    'buffer for DMM 2182
Global vvalues(2, 30) ' ?        'values from last scan (readnum, voltage)
Global ScanInt                    ' Scan time interval in Seconds
Global ScanIntAC_Global           ' Scan time interval in Seconds for AC test
Global rparange$(10) ' ?        'channel range
Global filename$                  'filename to store data (name changed for every new block)
Global FileName0$                 'initial given name to store dat
Global startTIME$, ScTime$
Global FiltN                       'total number of scans per loop
Global CycleN                      'total number of measurements since program started
Global chgPROGI                   'status for changing time interval
Global CFGFiles$                  'Config file's name
Global ConfigDon As Boolean
Sub Main()
  frmSplash.Show
  frmSplash.Refresh
  Set fMainForm = New frmMain
  Load fMainForm
  Unload frmSplash
  fMainForm.Show
End Sub

```

## Verification of the program.

Connected the equipment and the cells. The cell  
 → filled with saturated KCl solution (about 1 mm x 1 mm  
 slot, 30 mm length. two Pt electrode at each end).

Program configured to:

Scan interval = 2 sec.

Plotting Block size: 3

Data file Size: 5

Data File Name: test0611a

Frequency Window: 1000 to 100

Run for 10 data points.

test0611a00.txt should have the first 5 measurements  
data

test0611a01.txt should have the last 5 measurements  
data

test0611a00PData.txt should have the processed data  
from the first 5 measurements.

time and  $|z| = \sqrt{z_{\text{real}}^2 + z_{\text{im}}^2}$

at frequencies from 1k to 0.1M

test0611a01PData.txt same as test0611a00PData.txt  
except for the last 5 measurement

test0611aTemp should have the original data for  
the 10th measurement (last one)

J. Young 7/12/02

J. Young  
06/12/01

06-11-2001  
 Data File: c:\yang\test0611a00.txt  
 Zplot Setup File: C:\Yang\setup\_Yang\_Salt1.zpw  
 Scan Interval: 2

test0611a00.txt

Start T End T Z Plot Infor  
 , "ZPlotW Data File: Version 2.2"  
 , "Raw Data"  
 , "Sweep Frequency: Control Voltage"  
 , Date: 06-11-2001 Time: 11:21:53  
 , "Open Circuit Potential (V): 0.0091"  
 , ""  
 , "c:\yang\test0611aTemp"  
 , "Frequency"  
 , 0,3,0,1,100000,1000  
 , 7  
 , " Freq(Hz) Ampl Bias Time(Sec) Z'(a) Z'(b) GD Err Range"  
 , 1.000000E+05, 1.0000E-02, 9.1000E-03, 4.570000E+00, -4.4961E+05, -5.9232E+05, 0.0000E+00, 0, 5  
 , 4.641589E+04, 1.0000E-02, 9.1000E-03, 8.100000E+00, -1.9071E+06, -6.9128E+05, 0.0000E+00, 0, 5  
 , 2.154435E+04, 1.0000E-02, 9.1000E-03, 1.107000E+01, -5.4340E+06, -7.3760E+06, 0.0000E+00, 0, 6  
 , 1.000000E+04, 1.0000E-02, 9.1000E-03, 1.183000E+01, -3.5795E+06, -4.2886E+07, 0.0000E+00, 0, 6  
 , 4.641589E+03, 1.0000E-02, 9.1000E-03, 1.261000E+01, -2.2044E+05, -3.0648E+07, 0.0000E+00, 0, 6  
 , 2.154435E+03, 1.0000E-02, 9.1000E-03, 1.559000E+01, 2.2551E+07, -8.4681E+07, 0.0000E+00, 0, 7  
 , 1.000000E+03, 1.0000E-02, 9.1000E-03, 1.635000E+01, 1.1033E+08, -2.6423E+07, 0.0000E+00, 0, 7

*ist*

*end measurement*

11:22:1111:22:401.000000E+05, 1.0000E-02, -4.9200E-02, 4.580000E+00, -5.5239E+05, -6.1525E+05, 0.0000E+00, 0, 5  
 11:22:1111:22:404.641589E+04, 1.0000E-02, -4.9200E-02, 8.110000E+00, -1.3914E+06, -1.6430E+06, 0.0000E+00, 0, 5  
 11:22:1111:22:402.154435E+04, 1.0000E-02, -4.9200E-02, 1.108000E+01, -3.3036E+06, -7.2941E+06, 0.0000E+00, 0, 6  
 11:22:1111:22:401.000000E+04, 1.0000E-02, -4.9200E-02, 1.185000E+01, -4.8589E+07, 2.5259E+07, 0.0000E+00, 0, 6  
 11:22:1111:22:404.641589E+03, 1.0000E-02, -4.9200E-02, 1.262000E+01, -2.1200E+06, -1.5409E+07, 0.0000E+00, 0, 6  
 11:22:1111:22:402.154435E+03, 1.0000E-02, -4.9200E-02, 1.560000E+01, 4.8764E+07, -1.3461E+08, 0.0000E+00, 0, 7  
 11:22:1111:22:401.000000E+03, 1.0000E-02, -4.9200E-02, 1.637000E+01, -3.1213E+07, 3.3255E+08, 0.0000E+00, 0, 7

11:22:4211:23:111.000000E+05, 1.0000E-02, -6.0600E-02, 4.560000E+00, -5.2639E+05, -5.5051E+05, 0.0000E+00, 0, 5  
 11:22:4211:23:114.641589E+04, 1.0000E-02, -6.0600E-02, 8.090000E+00, -1.9177E+06, -5.2920E+05, 0.0000E+00, 0, 5  
 11:22:4211:23:112.154435E+04, 1.0000E-02, -6.0600E-02, 1.106000E+01, -4.2970E+06, -7.4879E+06, 0.0000E+00, 0, 6  
 11:22:4211:23:111.000000E+04, 1.0000E-02, -6.0600E-02, 1.182000E+01, 3.4666E+06, -1.6885E+07, 0.0000E+00, 0, 6  
 11:22:4211:23:114.641589E+03, 1.0000E-02, -6.0600E-02, 1.260000E+01, -2.0414E+07, 1.4467E+07, 0.0000E+00, 0, 6  
 11:22:4211:23:112.154435E+03, 1.0000E-02, -6.0600E-02, 1.558000E+01, 3.8405E+08, -9.8360E+07, 0.0000E+00, 0, 7  
 11:22:4211:23:111.000000E+03, 1.0000E-02, -6.0600E-02, 1.634000E+01, 3.8143E+07, -1.4340E+08, 0.0000E+00, 0, 7

11:23:1311:23:421.000000E+05, 1.0000E-02, -5.2100E-02, 4.520000E+00, -4.8889E+05, -5.5681E+05, 0.0000E+00, 0, 5  
 11:23:1311:23:424.641589E+04, 1.0000E-02, -5.2100E-02, 8.050000E+00, -1.7578E+06, -1.0861E+06, 0.0000E+00, 0, 5  
 11:23:1311:23:422.154435E+04, 1.0000E-02, -5.2100E-02, 1.102000E+01, -4.4299E+06, -6.4004E+06, 0.0000E+00, 0, 6  
 11:23:1311:23:421.000000E+04, 1.0000E-02, -5.2100E-02, 1.178000E+01, 2.0060E+07, -1.1879E+07, 0.0000E+00, 0, 6  
 11:23:1311:23:424.641589E+03, 1.0000E-02, -5.2100E-02, 1.256000E+01, 4.7353E+07, -1.9454E+07, 0.0000E+00, 0, 6  
 11:23:1311:23:422.154435E+03, 1.0000E-02, -5.2100E-02, 1.554000E+01, -9.3864E+06, -7.7453E+07, 0.0000E+00, 0, 7  
 11:23:1311:23:421.000000E+03, 1.0000E-02, -5.2100E-02, 1.630000E+01, -6.2214E+07, -1.6671E+08, 0.0000E+00, 0, 7

11:23:4411:24:121.000000E+05, 1.0000E-02, -6.3200E-02, 4.560000E+00, -5.9765E+05, -5.6860E+05, 0.0000E+00, 0, 5  
 11:23:4411:24:124.641589E+04, 1.0000E-02, -6.3200E-02, 8.090000E+00, -1.8117E+06, -7.0615E+05, 0.0000E+00, 0, 5  
 11:23:4411:24:122.154435E+04, 1.0000E-02, -6.3200E-02, 1.106000E+01, -7.4747E+06, -8.5262E+06, 0.0000E+00, 0, 6  
 11:23:4411:24:121.000000E+04, 1.0000E-02, -6.3200E-02, 1.182000E+01, -5.5299E+06, -1.4841E+07, 0.0000E+00, 0, 6  
 11:23:4411:24:124.641589E+03, 1.0000E-02, -6.3200E-02, 1.260000E+01, -1.3214E+07, -2.9004E+07, 0.0000E+00, 0, 6  
 11:23:4411:24:122.154435E+03, 1.0000E-02, -6.3200E-02, 1.558000E+01, 1.5992E+07, -9.1563E+07, 0.0000E+00, 0, 7  
 11:23:4411:24:121.000000E+03, 1.0000E-02, -6.3200E-02, 1.634000E+01, 3.4594E+07, -2.7458E+08, 0.0000E+00, 0, 7

*(ohm)*  
*(ohm)*

*Z'* *Z''*

*J. Yang*  
*7/15/02*

06-11-2001  
 Data File: c:\yang\test0611a01.txt  
 Zplot Setup File: C:\Yang\setup\_Yang\_Salt1.zpw  
 Scan Interval: 2

test0611a01.txt

*J. Y. 4/7/06*

*(ohm)* *(ohm)*

Start T End T Z Plot Infor  
 11:24:1511:24:431.000000E+05, 1.0000E-02, -6.0900E-02, 4.560000E+00, -4.8122E+05, -4.2531E+05, 0.0000E+00, 0, 5  
 11:24:1511:24:434.641589E+04, 1.0000E-02, -6.0900E-02, 8.090000E+00, -1.3866E+06, -2.2038E+06, 0.0000E+00, 0, 5  
 11:24:1511:24:432.154435E+04, 1.0000E-02, -6.0900E-02, 1.106000E+01, -3.4866E+06, -5.6345E+06, 0.0000E+00, 0, 6  
 11:24:1511:24:431.000000E+04, 1.0000E-02, -6.0900E-02, 1.183000E+01, -5.8251E+06, -1.4689E+07, 0.0000E+00, 0, 6  
 11:24:1511:24:434.641589E+03, 1.0000E-02, -6.0900E-02, 1.260000E+01, -3.5377E+07, 3.2043E+06, 0.0000E+00, 0, 6  
 11:24:1511:24:432.154435E+03, 1.0000E-02, -6.0900E-02, 1.558000E+01, 4.1575E+07, -6.5245E+07, 0.0000E+00, 0, 7  
 11:24:1511:24:431.000000E+03, 1.0000E-02, -6.0900E-02, 1.646000E+01, 1.9699E+08, -1.1513E+08, 0.0000E+00, 0, 7

11:24:4511:25:141.000000E+05, 1.0000E-02, -6.4900E-02, 4.570000E+00, -4.7017E+05, -5.5957E+05, 0.0000E+00, 0, 5  
 11:24:4511:25:144.641589E+04, 1.0000E-02, -6.4900E-02, 8.100000E+00, -1.1383E+06, -1.3035E+06, 0.0000E+00, 0, 5  
 11:24:4511:25:142.154435E+04, 1.0000E-02, -6.4900E-02, 1.107000E+01, -7.0252E+06, -4.9340E+06, 0.0000E+00, 0, 6  
 11:24:4511:25:141.000000E+04, 1.0000E-02, -6.4900E-02, 1.184000E+01, 1.7486E+07, -4.8168E+07, 0.0000E+00, 0, 6  
 11:24:4511:25:144.641589E+03, 1.0000E-02, -6.4900E-02, 1.261000E+01, 1.9252E+07, -1.3716E+07, 0.0000E+00, 0, 6  
 11:24:4511:25:142.154435E+03, 1.0000E-02, -6.4900E-02, 1.559000E+01, 3.4230E+07, -7.6293E+07, 0.0000E+00, 0, 7  
 11:24:4511:25:141.000000E+03, 1.0000E-02, -6.4900E-02, 1.635000E+01, 6.3508E+07, -7.8955E+07, 0.0000E+00, 0, 7

11:25:1611:25:441.000000E+05, 1.0000E-02, -7.6300E-02, 4.580000E+00, -5.3107E+05, -5.8753E+05, 0.0000E+00, 0, 5  
 11:25:1611:25:444.641589E+04, 1.0000E-02, -7.6300E-02, 8.110000E+00, -1.9930E+06, -1.0403E+06, 0.0000E+00, 0, 5  
 11:25:1611:25:442.154435E+04, 1.0000E-02, -7.6300E-02, 1.108000E+01, -3.3754E+06, -8.1497E+06, 0.0000E+00, 0, 6  
 11:25:1611:25:441.000000E+04, 1.0000E-02, -7.6300E-02, 1.185000E+01, 1.3605E+06, -1.3879E+07, 0.0000E+00, 0, 6  
 11:25:1611:25:444.641589E+03, 1.0000E-02, -7.6300E-02, 1.262000E+01, -7.8890E+06, -2.4201E+07, 0.0000E+00, 0, 6  
 11:25:1611:25:442.154435E+03, 1.0000E-02, -7.6300E-02, 1.560000E+01, -9.7444E+06, -8.5564E+07, 0.0000E+00, 0, 7  
 11:25:1611:25:441.000000E+03, 1.0000E-02, -7.6300E-02, 1.637000E+01, 2.5931E+08, -1.4160E+08, 0.0000E+00, 0, 7

11:25:4711:26:151.000000E+05, 1.0000E-02, -6.6400E-02, 4.580000E+00, -4.8310E+05, -5.2126E+05, 0.0000E+00, 0, 5  
 11:25:4711:26:154.641589E+04, 1.0000E-02, -6.6400E-02, 8.110000E+00, -2.7253E+06, -9.1866E+05, 0.0000E+00, 0, 5  
 11:25:4711:26:152.154435E+04, 1.0000E-02, -6.6400E-02, 1.108000E+01, -2.6191E+06, -6.2904E+06, 0.0000E+00, 0, 6  
 11:25:4711:26:151.000000E+04, 1.0000E-02, -6.6400E-02, 1.185000E+01, -4.4078E+06, -1.2086E+07, 0.0000E+00, 0, 6  
 11:25:4711:26:154.641589E+03, 1.0000E-02, -6.6400E-02, 1.262000E+01, -7.4146E+06, -1.0657E+07, 0.0000E+00, 0, 6  
 11:25:4711:26:152.154435E+03, 1.0000E-02, -6.6400E-02, 1.560000E+01, 2.6131E+07, -1.4086E+08, 0.0000E+00, 0, 7  
 11:25:4711:26:151.000000E+03, 1.0000E-02, -6.6400E-02, 1.636000E+01, 2.3469E+08, -1.3710E+08, 0.0000E+00, 0, 7

*both measurement*

11:26:1711:26:461.000000E+05, 1.0000E-02, -9.7800E-02, 4.570000E+00, -4.5571E+05, -6.2846E+05, 0.0000E+00, 0, 5  
 11:26:1711:26:464.641589E+04, 1.0000E-02, -9.7800E-02, 8.100000E+00, -1.7598E+06, -7.7029E+05, 0.0000E+00, 0, 5  
 11:26:1711:26:462.154435E+04, 1.0000E-02, -9.7800E-02, 1.107000E+01, -6.3606E+06, -1.0374E+07, 0.0000E+00, 0, 6  
 11:26:1711:26:461.000000E+04, 1.0000E-02, -9.7800E-02, 1.184000E+01, 2.6631E+06, -2.1526E+07, 0.0000E+00, 0, 6  
 11:26:1711:26:464.641589E+03, 1.0000E-02, -9.7800E-02, 1.261000E+01, 7.4683E+06, 2.0741E+07, 0.0000E+00, 0, 6  
 11:26:1711:26:462.154435E+03, 1.0000E-02, -9.7800E-02, 1.559000E+01, 3.8171E+07, -1.0844E+08, 0.0000E+00, 0, 7  
 11:26:1711:26:461.000000E+03, 1.0000E-02, -9.7800E-02, 1.635000E+01, 7.3821E+07, -1.4548E+08, 0.0000E+00, 0, 7

*Should add a comma.*

*J. Yang*  
*06/13/01*

06-11-2001  
 Data File: c:\yang\test0611a00PData.txt  
 Zplot Setup File: C:\Yang\setup\_Yang\_Salt1.zpw  
 Scan Interval: 2  
 Date ,Time Started, 100000, 46415.89, 21544.35, 10000, 4641.589, 2154.435, 1000,  
 06-11-2001,11:21:42, 743634.409168914, 2028521.24672136, 9161535.46082751, 43035123.0537337, 30648792.7624172, 87632296.341018, 113449917.712619,  
 06-11-2001,11:22:11, 826841.747010877, 2153007.88665532, 8007350.85843002, 54762286.3109275, 15554153.1752777, 143170457.134145, 334011607.38663,  
 06-11-2001,11:22:42, 761674.269094079, 1989378.27725146, 8633241.30382095, 17237184.821194, 25020501.2939389, 396445572.683061, 148386146.418727,  
 06-11-2001,11:23:13, 740979.627385261, 2066270.56553589, 7783902.24566059, 23313391.8810627, 51193405.0928437, 78019687.9893787, 177940456.040778,  
 06-11-2001,11:23:44, 824919.076334158, 1944454.86255146, 11338748.8961525, 15837773.6759306, 31872273.4049518, 92949056.1167783, 276750648.122096.

test0611a00PData.txt

Z (ohm)

ohm.

Verification.

the last data  $|z| = 276750648$  at 1000 Hz

From the last row of page 21

$F = 1000 \text{ Hz}$ ,  $Z_{real} = Z'(a) = 3.4594 \text{ E}7$

$Z_{im} = Z''(b) = -27458 \text{ E}8$

$$\sqrt{Z_{real}^2 + Z_{im}^2} = 2.76751 \text{ E}8$$

correct!

J.Y.  
 06/13/01  
 2/14/06

06-11-2001  
 Data File: c:\yang\test0611a01PData.txt  
 Zplot Setup File: C:\Yang\setup\_Yang\_Salt1.zpw  
 Scan Interval: 2

test0611a01PData.txt

4/7/06  
 Z values (ohm)

Date ,Time Started, 100000, 46415.89, 21544.35, 10000, 4641.589, 2154.435, 1000,  
 06-11-2001,11:24:15, 642231.488250148, 2603726.94420901, 6626007.0789277, 15801851.5057572, 35521819.5971152, 77365306.5010409, 228166555.393204,  
 06-11-2001,11:24:45, 730875.101368216, 1730560.35433613, 8584741.75732736, 51243696.3928247, 23638277.4330111, 83620061.881106, 101326985.985965,  
 06-11-2001,11:25:16, 791976.34371831, 2248171.05443514, 8821050.68855179, 13945522.6237671, 25454365.4802079, 86117079.7656307, 295452595.351606,  
 06-11-2001,11:25:47, 710702.186291839, 2875968.75601944, 6813869.45648359, 12864684.0940615, 12982601.5174155, 143263284.762705, 271801041.388733,  
 06-11-2001,11:26:17, 776294.773716789, 1921000.4487506, 12168693.7819965, 21690107.8284549, 22044604.46209, 114961988.678867, 163137887.815798,

Verification

The first Data:

$|z| = 642231$  at 100000 Hz.

From the first data row on page 23.

$F = 1 \text{ E}5$ ;  $Z' = -4.8122 \text{ E}5$

$Z'' = -4.2531 \text{ E}5$ .

$$\sqrt{Z'^2 + Z''^2} = 6.42231 \text{ E}5$$

correct!

J.Y.  
 06/13/01  
 2/14/06

"ZPlotW Data File: Version 2.2"  
 "Raw Data"  
 "Sweep Frequency: Control Voltage"  
 Date: 06-11-2001 Time: 11:26:28  
 "Open Circuit Potential (V): -0.0978"  
 ""

test0611aTemp

"c:\yang\test0611aTemp"  
 "Frequency"  
 0,3,0,1,100000,1000

Freq(Hz)	Ampl	Bias	Time(Sec)	Z'(a)	Z''(b)	GD	Err	Range"
1.000000E+05	1.0000E-02	-9.7800E-02	4.570000E+00	4.5571E+05	-6.2846E+05	0.0000E+00	0, 5	
4.641589E+04	1.0000E-02	-9.7800E-02	8.100000E+00	-1.7598E+06	-7.7029E+05	0.0000E+00	0, 5	
2.154435E+04	1.0000E-02	-9.7800E-02	1.107000E+01	-6.3606E+06	-1.0374E+07	0.0000E+00	0, 6	
1.000000E+04	1.0000E-02	-9.7800E-02	1.184000E+01	2.6631E+06	2.1526E+07	0.0000E+00	0, 6	
4.641589E+03	1.0000E-02	-9.7800E-02	1.261000E+01	7.4683E+06	2.0741E+07	0.0000E+00	0, 6	
2.154435E+03	1.0000E-02	-9.7800E-02	1.559000E+01	3.8171E+07	-1.0844E+08	0.0000E+00	0, 7	
1.000000E+03	1.0000E-02	-9.7800E-02	1.635000E+01	7.3821E+07	-1.4548E+08	0.0000E+00	0, 7	

agrees with  
 last paragraph of  
 test0611a01.txt. page 23.

conclusion:

The data files <sup>from AutoZplot</sup> contains the data that are the same as the <sup>original</sup> data from the Zplot alone.

J.Y. 06/13/01

J.Y. 2/14/06

(Ohm)

**Alfa Aesar**  
 A Johnson Matthey Company

30 BOND STREET WARD HILL, MA 01835

SOURCE: RESEARCH INSTITUTE  
 ATOM. J. YANE HIDE 189

5220 CALLEDA RD TX 78236  
 SAN ANTONIO

TEL # 210-522-2275

SHIP TO

EMERGENCY PHONE NUMBER  
 1-800-424-9300

PHONE NUMBER  
 D-U-N-S  
 FEDERAL ID NO.

CUST# 55022509 DAB

DATE SHIPPED	SHIPPING CHARGES
5-17	
SHIPPER	NO. OF PKGS.
DA	1
SHIPPING	GROSS WEIGHT

CUSTOMER ORDER NUMBER	DATE ENTERED	ALFA AESAR ORDER NUMBER	F.O.B. PLANT	PRINT DATE & TIME
1972507	5/17/01	5415571	WARD HILL	5/17/01 15:45:14

TERMS

NET 30 DAYS /COL

STOCK NUMBER	UNIT OF MEASURE	SHIP QTY	I/O QTY	ORDER QTY	SHIPPING DATA AND DESCRIPTION
10285-01	100 CM	1		1	PLATINUM WIRE, 0.5MM (0.02IN)
8A-RE-05	MSDS Y				MSDS Y
610286					MSDS Y

PLATINUM WIRE, 0.5MM (0.02IN)  
 MSDS Y  
 MSDS Y

10286

10760

\*\*\* ORDER IS COMPLETE \*\*\*  
 THANK YOU FOR YOUR ORDER.

Platinum wire to be used as electrode  
 in the conductivity measurement. (similar to counter electrode)

No need to make entries to Lab log book.


J.Y. 2/14/06

J.Y. 06/14/01




Conductivity Cell

File:Teflon\_Cell2



Note: Pencil writings were by machine shop Technician.

2.7 gms  
7/15/02



Cell constant: width : W  
 Thickness :  $\delta$   
 Length : L (related to L')

2.7 gms  
7/15/02

Two cells made.

Cell one (I) :  $W = 2.6 \text{ cm}$  ,  $L = 4.7 \text{ cm}$  ( $L' = 4.5 \text{ cm}$ )  
 Cell Two (II) :  $W = 2.6 \text{ cm}$  ,  $L = 4.5$  ( $L' = 4.4 \text{ cm}$ )

Paper thickness:

① measured 100 sheets <sup>(dry)</sup>, 2.3 cm thick  
 therefore  $\frac{2.3}{100} = 0.023$  cm thick/sheet.

② Soak <sup>dissolved</sup> water then measure water weight  
 using 15 cm diameter circle sheet.  
 Paper average weight: 1.62 g

weight gain when soaked on flat horizontal surface  
 after lifted, remove extra <sup>drops</sup> at bottom.

6.25 g  
 5.94 g      average: 6.07 g  
 6.03 g

weight gain when soaked vertically, and after  
 removing extra drops at bottom.

5.12 g  
 5.68 g      average 5.62 g  
 5.65 g      avg. 0.2/0.1

~~Net weight gain: 6.07 - 1.62 = 4.45 = 5.~~

Average weight gain vertical + horizontal 5.845 g.

Net weight gain  $5.845 - 1.62 = 4.225$  2.2 grams  
0.1/0.2

thickness of water if assume no paper  
 $\delta = \frac{V}{A} = \frac{4.225 \text{ cm}^3}{[(15^2/4) * 3.14]} = 0.0239 \text{ mm}$

Therefore use  $\delta = 0.023$  cm (the dry paper thickness)  
 as the cell thickness for pure water.

cell constant for I  
 $\frac{l}{S} = \frac{4.7}{2.6 * 0.023} = 78.6 \frac{1}{\text{cm}}$

cell constant for II  
 $\frac{l}{S} = \frac{4.5}{2.6 * 0.023} = 75.3 \frac{1}{\text{cm}}$


Jayas  
 08/07/01

conductivity cells

08/01/01

J. Yang

File: Cond\_cell



Quantities: one h=1 mm, w=2 mm  
 one h=0.5 mm, w=2 mm  
 one h=0.5 mm, w=1 mm  
 one h=0.25 mm, w=1 mm

Material: Teflon

Charge Number: 20.1402.561

JUN 1 2001  
 Cal 30-60-3  
 JC# 74923  
 4 pcs

06/01/01

J. Yang

File: Conductivity\_Dimension of cells  
 Conductivity measurement:

	0.2	0.5	1	2	3	5
l=cm (length)	0.2	0.5	1	2	3	5
h=cm (height)	0.1	0.1	0.1	0.1	0.1	0.1
w=cm (width)	0.2	0.2	0.2	0.2	0.2	0.2
S=cm <sup>2</sup>	0.02	0.02	0.02	0.02	0.02	0.02
Cond= 1/cm-ohm	0.2	0.2	0.2	0.2	0.2	0.2
R= (ohm)	1250	750	500	250	250	125
I= A (with 10 mV)	0.008	0.013333	0.02	0.04	0.04	0.08
l=cm	5	3	2	1	1	0.5
h=cm	0.05	0.05	0.05	0.05	0.05	0.05
w=cm	0.2	0.2	0.2	0.2	0.2	0.2
S=cm <sup>2</sup>	0.01	0.01	0.01	0.01	0.01	0.01
Cond= 1/cm-ohm	0.2	0.2	0.2	0.2	0.2	0.2
R= (ohm)	2500	1500	1000	500	500	250
I= A (with 10 mV)	0.004	0.006667	0.01	0.02	0.02	0.04
l=cm (length)	5	3	2	1	1	0.5
h=cm	0.05	0.05	0.05	0.05	0.05	0.05
w=cm	0.1	0.1	0.1	0.1	0.1	0.1
S=cm <sup>2</sup>	0.005	0.005	0.005	0.005	0.005	0.005
Cond= 1/cm-ohm	0.2	0.2	0.2	0.2	0.2	0.2
R= (ohm)	5000	3000	2000	1000	1000	500
I= A (with 10 mV)	0.002	0.003333	0.005	0.01	0.01	0.02
height	5	3	2	1	1	0.5
base width	0.3	0.3	0.3	0.3	0.3	0.3
w=cm	0.2	0.2	0.2	0.2	0.2	0.2
S=cm <sup>2</sup>	0.06	0.06	0.06	0.06	0.06	0.06
Cond= 1/cm-ohm	0.2	0.2	0.2	0.2	0.2	0.2
R= (ohm)	416.6667	250	166.6667	83.33333	83.33333	41.66667
I= A (with 10 mV)	0.024	0.04	0.06	0.12	0.12	0.24

weight calculation

a (width) (cm)	0.2
b (height, cm)	0.3
S (Area, cm <sup>2</sup> )	0.06
L (length, cm)	4
V (volume, cm <sup>3</sup> )	0.24
Weight	water Sat KCl
density	0.24 0.29
volume cm <sup>3</sup>	1 1.2
	0.24 0.241667

J. Yang

7/15/02

06/12/01

conductivity measurement.

Cell:  $h = 1\text{ mm}$   
 $w = 2\text{ mm}$   
 $L = 3.0\text{ cm}$  } see page 32

Platinum Electrode: diameter 0.5 mm  
see page 27.

filtration paper cut to fit the conductivity cell slot. (2 mm x 1 mm (deep) x 30 mm length)

10:09 Paper soaked with saturated KCl solution  
~~lot #~~ dry. 06/12/01 lot # 005573

lot #: 005573 dry paper thickness 0.13 mm. dry. 7/16/01

Paper placed in cell, Pt electrode inserted through the paper to have good contact.


cell placed in GS Blue In Electric Humidity Chamber

Filtration Paper: Whatman Cat No 1004150  
Cut from 150 mm dia, Quantitative Filtration Paper. J. Yarn 7/15/02

GS Blue In Electric Humidity Chamber.

The attached indicator: SN: 91E34523

Calibrated: May 05, 01  
by kWh.  
SWRI Cal-Lab.



J. Yarn  
7/15/02

06/12/01

10:09 VB program (Auto Zplot) started, File: 0612b-SatKCl

10:12 Chamber set to (22<sup>2</sup>/30<sup>2</sup>) (Wet T/Dry T)

	Data #	Twet/Tdry	Impedance (Ω)	setting
10:12	5	22/30	log Z  = 7	22 <sup>2</sup> /30 <sup>2</sup>
10:17	7	29/30	log Z  = 5	
11:06	#57	29/30	log Z  = 3.1	to 19 <sup>2</sup> /30 <sup>2</sup>
12:28	#130	22/30		to 28.5/30 <sup>2</sup>
13:48	#201	27/30	log Z  = 5.5	
15:03	276	27/30	log Z  = 4.3	29.75/30
16:18	350	27/30	log Z  = 3.5	to 30.25/30
16:46		28/29		

program stopped. accidentally.

16:53 program restarted, File: 0612c-Sat-KCl

06/13

8:01	761	28/30		30.25/30
8:08			program halted.	to 29.5/30
8:14			program restarted	0613a-Sat-KCl
9:08	#47	27/30		29.5/30
9:57	#89	27/30		to 29.25/30
11:55	#189	27/30		to 28.75/30
15:22	362	27/30		to 28/30
18:27	368	26/30		

J. Young 7/15/02

06/14/01

	Data in File #	Twet/Tdry	Imp. (Ω)	settings (Twet/Tdry)
11:22				
11:22		27/30	log Z  = 5.2	28/30
11:22		New File:	0614a	to 27/30
11:33		26/30		27/30
13:50		26/30	log Z  = 5.8	to 26/30
16:04	#138	25/30	log Z  = 6.0	
16:18	145	24.5/30	(at 1k - default)	to 24.75/30
18:02	196	24/30	log Z  = 6.5	
18:06				to 23.75/30 <sup>2</sup>
<del>07:44</del>	06/15/01 #598	24/30		to 22/30 <sup>2</sup>
07:44	J.Y. 06/15/01			
7:50		23/30	New File: <del>0615b-Sat-KCl</del>	0615a J.Y. 06/15/01
10:17	#73	23/30		22/30

Calibration the Dial-control

10:18	Set		set to 26/30
10:21	25/30		set to 30/30
10:24	29/30		set to 29/30
10:32	27/30		to 32/30
10:38	29/30		to 31/30
10:43	24.5/30		to 30/30
10:45	28/30		to 29/30
10:49	28/30		to 31/30
10:52	29/30		30/30 33/30 J.Y. 06/15/01

J. Young 7/15/02

J.Y. 06/12/01

10:53 ~~29.5~~ 29.5/30

10:57 29.5/30 to 32.5/30

11:01 29.5/30 to 31.5/30

11:23 29/30 to 30.5/30

from above. the maximum  $T_{wet}$  is  $29.5^{\circ}C/30$   
 at  $100\%$ , so all measured results should offset by  $\underline{-0.5^{\circ}C}$

11:33 28/30 to 29.5/30

11:47 28/30 to 28.5/30

11:50 27/30 to 20/30

14:31 22/30

15:09 Test terminated. Cell rinsed,  
 reloaded with  $NaNO_3$  soaked filter paper.

J. Young 7/15/02


15:24 load  $NaNO_3$  — saturated — solution soaked filter  
 paper. Fisher C.A.C.S.  
 $NaNO_3$  — lot # 961772A.

before load the  $NaNO_3$  soaked paper, put a wet  
 unsoaked paper, same size as used in the real test  
 and put in the chamber. measure when the  
 paper was wet.

15:35 #5  $\log|Z|$  at  $1k = 7.5$   
 program started 15:24; 0615b  
 Proof: paper even with water (Deionized water)  
 does not conduct.

16:23 #29  $\log|Z| = 7.3$  (at  $1k Hz$ )

Go to page 47.




J.Y. 6/19/01

J. Young 2/14/06

0612b\_00\_14a\_01.xls


Chart1

Data For kull Test



Freq: (Hz)

J. Yans 7/15/02




fz (Hz)

This diagram shows that only  $f = 1000$  and  $f = 1585$  respond to humidity changes at lower levels (Section A). ~~See~~ Aug. 06/15/01

See also Page 42.

J. Yans 7/15/02



Should use  $f = 1000$  data because these data respond well to Humidity change at low RH.

J. Jans 7/15/02

J. Jans  
06/15/01


Test Results from June 12 to June 15

Time	Indicated Reading		Wet Bulb Corrected* by +0.5oC	RH** %	Dial Reading		log(Z, ohm)
	Dry Bulb	Wet Bulb			Dry Bulb	Wet Bulb	
6/12/01 16:18	30	28	28.5	90	30	30.25	2.75
6/13/01 8:01	30	27	27.5	84	30	29.5	2.8
6/13/01 9:57	30	27	27.5	84	30	29.25	2.9
6/13/01 11:55	30	27	27.5	84	30	28.75	3
6/13/01 15:27	30	27	27.5	84	30	28	5.3
6/14/01 11:22	30	26	26.5	77	30	27	5.7
6/14/01 13:50	30	25	25.5	71	30	26	6.1
6/14/01 16:18	30	24	24.5	64	30	24.75	6.5
6/14/01 18:06	30	24	24.5	64	30	23.75	6.7

\* see Note Book Page 38

\*\* According to Perry's Chem Eng. Handbook, 6th Ed, page 20-6

Numbers see page 40.



J. Jans  
7/15/02



The plot on page 43 is too coarse. RH error by  $\pm 0.5\%$  on wet T plus  $\pm 0.5\%$  on dry T. we have

About  $\pm 5\%$  — too big.


From data shown on page 36, when the setting Dial temp changes, Z changes. The dial has a better resolution  $0.25^\circ\text{C}$ . Therefore the dial value should be used:

As shown in page 45. There is good correlation between Dial setting temp. and humidity.

Regression:

$$RH(\%) = -0.0368 T_{\text{wet, set}}^2 + 6.2279 T_{\text{wet, set}} - 64.63$$

Results see page 46.



J.Y.  
06/15/01

J. Yarns  
2/14/06


0612b\_DP.xls

Calibration Data, Scientific Note Book P. 36 to 38

	Wet Bulb Corrected*	RH** %	Set Dry T oC	Set Wet T oC
NBook P36	30	22.5	23	19
NBook P36	30	27	27.5	30
NBook P36	30	27	27.5	30
6/12/01 15:03	30	27	27.5	30
6/12/01 16:18	30	28	28.5	30
6/13/01 8:01	30	27	27.5	30
6/13/01 9:57	30	27	27.5	30
6/13/01 11:55	30	27	27.5	30
6/13/01 15:27	30	27	27.5	30
6/14/01 11:22	30	26	26.5	30
6/14/01 13:50	30	25	25.5	30
6/14/01 16:18	30	24	24.5	30
6/14/01 18:06	30	24	24.5	30
6/15/01 10:18	30	26.5	27	30
Notebook P37-38	30	25	25.5	30
Notebook P37-38	30	28	28.5	30
Notebook P37-38	30	27	27.5	30
Notebook P37-38	30	29	29.5	30
Notebook P37-38	30	29	29.5	30
Notebook P37-38	30	28	28.5	30
Notebook P37-38	30	28	28.5	30
Notebook P37-38	30	29	29.5	30
Notebook P37-38	30	29.5	30	31
Notebook P37-38	30	29.5	30	30
Notebook P37-38	30	30	30	31
Notebook P37-38	30	29.5	30	33
Notebook P37-38	30	29.5	30	32.5
Notebook P37-38	30	29	29.5	31.5
Notebook P37-38	30	28	28.5	30.5
Notebook P37-38	30	28	28.5	30.5
Notebook P37-38	30	27	27.5	29.5
6/15/01 11:47	30	24	24.5	28.5

\* see Note Book Page 38

\*\* According to Perry's Chem Eng. Handbook, 6th Ed, page 20-6



J.Y.  
06/15/01

J. Yarns  
2/14/06