


0612b\_DP.xls

Calculation of RH based on the setting dial of the wet temperature controller  
Test Results from June 12 to June 15

Time	Indicated Reading		Wet Bulb Corrected* by +0.5oC		RH** %	Dial Reading		log(Z, ohm)	RH Calc.	Regression expression for calculation RH = -0.0368x2 + 6.2279x - 64.627 Calculated RH
	Dry Bulb	Wet Bulb	Dry Bulb	Wet Bulb		Dry Bulb	Wet Bulb			
6/12/01 16:18	30	28	28.5	28.5	90	30	30.25	2.75	90.1227	-0.0368
6/13/01 8:01	30	27	27.5	27.5	84	30	29.5	2.8	87.1008	6.228
6/13/01 9:57	30	27	27.5	27.5	84	30	29.25	2.9	86.0843	-64.6
6/13/01 11:55	30	27	27.5	27.5	84	30	28.75	3	84.0375	
6/13/01 15:27	30	27	27.5	27.5	84	30	28	5.3	80.9328	
6/14/01 11:22	30	26	26.5	26.5	77	30	27	5.7	76.7288	
6/14/01 13:50	30	25	25.5	25.5	71	30	26	6.1	72.4512	
6/14/01 16:18	30	24	24.5	24.5	64	30	24.75	6.5	67.0007	
6/14/01 18:06	30	24	24.5	24.5	64	30	23.75	6.7	62.5575	

\* see Note Book Page 38

\*\* According to Perry's Chem Eng. Handbook, 6th Ed, page 20-6



agrees well with literature data.

J. Yang. 06/15/01

06/15/01

Cont from page 39.

Lot # 961772A

16:45.

NaNO<sub>3</sub> soaked paper (dry paper originally) placed in cell and cell connected, placed in Humidity chamber

16:

Excess water removed.

17:00 #48 file not changed.

setting (T<sub>w</sub>/T<sub>d</sub>)

Data#, T<sub>w</sub>/T<sub>d</sub> Imped (Ω)

17:02 49

to 27/30

17:15 57

25/30 log(Z) = 3.4

16:17 808

06/16/01

26/30 log Z = 3.3

to 26/30

16:21 810

27/30

06/18/01

7:40 1997

25/30

log Z = 5.5

to 25/30

7:47 1999

24/30

8:46 2029

24/30

log Z = 5.8

to 22/30

9:02 2037

23/30

to 15/30

9:30

23/30

9:35

21/30

10:12

21/30

to 22/30

10:17

23/30

10:31

23/30

10:33

23/30

to 33/30

10:37

27/30

J. Yang. 06/16/01

10:38

28/30

J. Yang 7/15/02

11:04

28/30

Time	Data #	Td/Tw	Z	Set Td/Tw
------	--------	-------	---	-----------

13:08	2160	27/30	log z = 7.2	30/30
-------	------	-------	-------------	-------

13:22				to 27/30
-------	--	--	--	----------

13:33		25/30		
		26/30		

13:46	2184	27/30		
-------	------	-------	--	--


13:24			program quit.	
-------	--	--	---------------	--

Note. From 13:01 to 13:04. open circuit measurement  
see page 50

*[A diagonal line is drawn across the page]*

J.Y. 06/19/01  
J.Y. 2/14/06

Conductivity




0615b00\_05.xls

J.Y. 7/12/02

0615b00\_05.xls

# Maximum Limit of Solartron Equipment, Open Circuit Measurement

Open\_Circuit




4/7/06  
freq.  
(Hz)

- ◆ 100000
- 63096
- ▲ 39811
- × 25119
- ✱ 15849
- 10000
- + 6310
- 3981
- 2512
- ◆ 1585
- 1000

J. Young  
7/15/02

0615b\_00\_05.xls

Conductivity



J. Young  
4/7/06  
freq.  
(Hz)

- ◆ 100000
- 63096
- ▲ 39811
- × 25119
- ✱ 15849
- 10000
- + 6310
- 3981
- 2512
- ◆ 1585
- 1000

J. Young  
7/15/02

RH based on the setting dial of the wet temperature controller  
Test Results from June 16 to June 18, Scientific Notebook Page 39, 47-48

See Page 46

\*\*\* Regression expression for calculation  
RH = -0.0368x2 + 6.2279x - 64.627


Time	Indicated Reading		Corrected* by +0.5oC	RH** %	Dial Reading		log(Z, ohm)	Calc.*** RH, %
	Dry Bulb	Wet Bulb			Dry Bulb	Wet Bulb		
① 6/18/01 13:10	30	28	28.5	90	30	30	3	89.1
② 6/18/01 15:00	30	25.5	26	75	30	27	3.3	76.7
③ 6/18/01 7:00	30	25	25.5	70	30	26	5.6	72.5
④ 6/18/01 8:00	30	24	24.5	64	30	25	6	68.1
⑤ 6/18/01 9:00	30	23	23.5	58	30	22	6.55	54.6
⑥ 6/18/01 10:00	30	21	21.5	43	30	15	6.9	20.5

\* see Note Book Page 38

\*\* According to Perry's Chem Eng. Handbook, 6th Ed, page 20-6

*Handbook value = 73%*

*DP = 75.*



*L. Yang  
7/15/02*



PURCHASE REQUISITION  
SOUTHWEST RESEARCH INSTITUTE™

5 / JUL 03

VENDOR		DATE OF ORDER		SHIP VIA		SUGGESTED SUPPLIER	
ATTN:		F.O.B. POINT		TERMS		COLE PARMER	
PHONE		CONTRACT NO.		PHONE		FAX	
800 323-4340				800 323-4340			
ITEM	QTY.	UNIT	DESCRIPTION	ITEM IDENTIFIER	EST. COST EA.	UNIT PRICE	AMOUNT
A	1	EACH	FAST-RESPONSE THERMOHYGROMETER	#U-03313-66		\$410.00	
B	1	EACH	POWER ADAPTER U-03313-95			\$ 22.00	
C	1	EACH	DATA CABLE U-03313-98			\$71.00	
D	1	EACH	PARALLEL TO SERIAL CONVERTER	#U-03313-99		\$160.00	
<b>RA Requirement: NIST CALIBRATION CERTIFICATE IS REQUIRED</b> <i>Required.</i> <b>NOTE: SWRI Calibration Lab must register and be routed to Cal Lab for checks. See</b>							TOTAL
DATE REQUIRED		REQUISITION DATE		TO BE USED FOR		SPECIAL INSTRUCTIONS	
6/25/01		6/17/01		DELIVER TO NAME/LOCATION		SEND REQ R. - ARD BLDG. - 189 2ND	
ACCOUNT NO (s)		REQUESTOR'S SIGNATURE		EXT. NO.		3. If this requisition is for a repair-is the repair on campus or off campus?	
20-01402-561		L. YANG		L. YANG/BLDG. 57		ON <input type="checkbox"/> OFF <input type="checkbox"/>	
20.0754		DEPARTMENTAL/DIVISIONAL APPROVAL		DATE		If OFF CAMPUS indicate shipping ticket #	
		E. PEARCY		6/18/01		5. Do you want receiving to open and inspect your shipment?	
<input checked="" type="checkbox"/> YES <input type="checkbox"/> NO		<input checked="" type="checkbox"/> YES <input type="checkbox"/> NO		<input checked="" type="checkbox"/> YES <input type="checkbox"/> NO		<input type="checkbox"/> YES <input type="checkbox"/> NO	
4. Government Project? If the item is being charged to a government contract: a. Will it be substantially consumed, destroyed, or exhausted during the performance of the project?		b. If "NO", will it be attached to, built into, or used as an accessory to a piece of equipment either in existence or to be constructed? c. If "NO", is the item uniquely required to instrument the project that is funding the purchase?		d. Is government furnished property being sent to vendor? If "NO", and total estimate exceeds \$3000.00, attach a memo of explanation.		e. QA Approval (if required) Date	
<input checked="" type="checkbox"/> YES <input type="checkbox"/> NO		<input type="checkbox"/> YES <input type="checkbox"/> NO		<input type="checkbox"/> YES <input type="checkbox"/> NO		<input type="checkbox"/> YES <input type="checkbox"/> NO	
CONTRACT ADMINISTRATOR		BUYER SIGNATURE <i>David M. ...</i>		DATE		EXPEDITE DATE	
						REVERSE SIDE	

*L. Yang 7/17/02*



Southwest Research Institute  
 6220 Culebra Road  
 San Antonio, TX 78238  
 (210) 522-5215  
 Department of Quality Assurance  
 Calibration Laboratory



Certificate #  
 0972-01

# Certificate of Calibration

5 July 2001

**Issued to:** DARRELL DUNN DIV20 B57  
**Manufacturer/Model:** COLE-PARMER 03313-66  
**Description:** THERMOHYGROMETER  
**Serial Number:** 21189381  
**Asset Number:** 008788  
**Work Order Number:** 444044223

This certifies the above item was calibrated in compliance with MIL-STD-45662A and ANSI/NCSL Z540-1-1994. The results of this calibration relate only to the individual item as described above. Standards used in this calibration, described in the referenced calibration procedure with associated uncertainties or tolerances, are traceable to the National Institute of Standards and Technology (NIST). Supporting documentation relative to traceability is on file and available for examination upon request. This certificate is not to be reproduced, except in full, without the written approval of the Southwest Research Institute Department of Quality Assurance Calibration Laboratory.

This laboratory is accredited by the American Association for Laboratory Accreditation (A2LA) and the results of this calibration certificate were determined in accordance with the terms of accreditation unless stated otherwise below.

The uncertainty of the calibration was sufficient to determine that the item met the manufacturer's published specifications unless stated otherwise below.

**Ambient Conditions:** Temperature: 75.0 Degrees Fahrenheit Humidity: 50 % RH

**Calibration Date:** 5 Jul 01 **Calibration Procedure:**

**Condition as Received:** SEE ATTACHED DATA

**Condition as Released:** SEE ATTACHED DATA

**Remarks:** SEE ATTACHED DATA SHEET FOR INSTRUMENT ERROR AND UNCERTAINTY.

Approved by:

*Walt Hill*

Walt Hill, Supervisor  
 Institute Calibration Laboratory

Measurements performed by:

*Roger Dykstra*

Roger Dykstra, Technician

*Walt Hill 07/04/01* *J. Jones 2/14/06*

**TRACEABLE®**

**Certificate of Calibration for Digital Hygrometer/Thermometer**

Certificate Number 4085.057974	Model Number New 03313-66	Serial Number 21189381
-----------------------------------	------------------------------	---------------------------

This Electronic Digital Hygrometer/Thermometer was calibrated against National Institute of Standards and Technology Traceable Instrumentation. This calibration complies with the requirements of ISO 9000 Certification.

**Calibration Test Information**

Test Equipment Humidity Standard 118	Serial Number 99324814	Calibration Due Date 2/08/02
TEMPERATURE STANDARD	20267165	2/05/02

NIST Traceable Test Number(s) **NCSL Z540-1  
100082023/1000101790**

**Accuracy**  
 Testing was performed on the unit as shown below. Test results are as follows:

Standard %RH	Reading %RH	Standard °C	Reading °C
23.08	21.63	24.44	24.45
52.53	52.97		
78.76	78.06		

The maximum error of this unit at the time of calibration did not exceed the specified accuracy of: ± 1.5% RH with a combined uncertainty of 0.240% and a confidence level of 95% 0.2°C

Test Conditions:	Temperature °C	Relative Humidity %	Barometric Pressure (inHg)
	27	41	29.71

**Maintaining Accuracy**  
 Once measured and calibrated your Electronic Digital Hygrometer/Thermometer should maintain its accuracy. There is no exact way to determine how long calibration will be maintained. Electronics change little, if any at all, but can be affected by aging, temperature, and shock.

**Calibration Dates**

Factory Calibration Date 05/07/01	Next Calibration Due Date 05/07/03
--------------------------------------	---------------------------------------

We recommend that the unit's accuracy be recertified on an annual basis for those users with critical needs such as accreditation demands, government specifications, or ISO 9000 requirements.

Tester's Initials PUL	Metrology Manager <i>Wallace Berry</i>
--------------------------	---

**Recalibration**  
 For factory calibration and recertification of this Hygrometer/Thermometer contact  
 Control Company • 308 West Edgewood • Friendswood, Texas 77546 • USA  
 Phone 281 482-1714 • Fax 281 482-9448  
 Control Company is an ISO 9001 Quality Certified Company  
 ISO 9001 Certificate No. CERT-01805-2000-AQ-HOU-RAB  
Traceable is a registered trademark of Control Company

*J. Jones*  
*07/04/01*  
*J. Jones*  
*2/14/06*

SWRI Calibration Laboratory Data Sheet

Work Order: 444044223	Manufacturer: Cole-Parmer	Technician: R Dykstra
Asset Number: 008788	Model: 03313-66	Procedure: Pending
Serial Number: 21189381	Type: Thermo-hygro	Calibration Date: 7/3/01

**Remarks:** The instrument error of this unit did not exceed the specified limit of +/- 1.5 % R.H. with a combined uncertainty (K=2) of 0.5 % RH and a confidence level of approximately 95%.  
 The instrument error of this unit did not exceed +/-0.2 degree C with a combined uncertainty (K=2) of 0.2 degree C and a confidence level of approximately 95%.

Humidity Tolerance: 1.5  
 Range: 10 to 95 % RH

Applied Value % RH	As Found Ind. Value (% RH)	Instrument Error	Instrument Tolerance	Measurement Uncertainty
24.99	23.53	1.46	1.5	0.5
50.78	49.71	1.07	1.5	0.5
75.03	74.15	0.88	1.5	0.5


Applied Value % RH	As Left Ind. Value (% RH)	Instrument Error	Instrument Tolerance	Measurement Uncertainty
24.99	23.53	1.46	1.5	0.5
50.78	49.71	1.07	1.5	0.5
75.03	74.15	0.88	1.5	0.5

Temperature Tolerance: 0.2  
 Range: -40 to 60 Degree C

Applied Value Deg C	As Found Ind. Value (Deg C)	Instrument Error	Instrument Tolerance	Measurement Uncertainty
24.97	25.10	-0.13	0.2	0.2

Applied Value Deg C	As Left Ind. Value (Deg C)	Instrument Error	Instrument Tolerance	Measurement Uncertainty
24.97	25.10	-0.13	0.2	0.2

*J. Young*  
 6/29/01



Information potentially subject to copyright protection was redacted from this location. The redacted material was from a catalog for thermohygrometers.

*J. Young*  
 6/29/01

07/16/01

Experiments using Arc Cell. (See Pages 28-29)

Test will be conducted at 50°C because the humidity chamber can not go down to 35% of RH at 30°C

Cell II — KCl saturated (at 24°C) soln  
lot # 005573  
Cell I — KNO3 saturated (at 24°C) soln  
lot # 702860-p0

Put as much solution as possible on to the paper until it flows off.

Electrodes are 0.5 mm Pt (see page 27)

9:02 cells installed in chamber and set  $T_d/T_w = 50.2/51.9$

until 07/18 14:16 having problem with the humidity controller.

07/18/01

14:16 Program started. File name: 0718a.  
14:16 45/50  $\log|Z|_I = 271 \Omega$ , cell connected to I  
14:30  $\log|Z|_{II} = 210 \Omega$ , cell connected to II

Note: Solartom system was manually connected to one of the two cells at a time. J. Yano 7/18/02

Note: Hygrometer was used, see page 57. It was inserted into the chamber.

Time	Data # in File	$T_w/T_d$	$Z, \Omega$	Hygrom. Cell #	Control $T_w/T_d$
14:39	#10	45/50		100%, II	(41/50, 35°C)
				** 35°C — cooling set Temp.	
14:52	#16	43/50		94.6% -100%, II	to (39/50, 35°C)
15:34	#38	42/50		97.8-85.9%, II	to I
16:02	#51	41.5/50		94.24 -83.24%	to II
16:43	73	41/50		94.2-81.3%	to I
17:11					program terminated because the C-22 test need equipment.

07/19/01

07/					
08:19				program started, file: 0719a	to I
08:19		41/50		78.3-89.7%	
8:37					to II
9:01	#20				to I
9:36	#	41/50			to (37/50, 35°C)
10:09		40/50			
10:57	#74				to II

J. Yano 7/18/02

Time	Date#	Tw/Td Td/Tw	h.g. 07/20/01	Z, R	Hygrom.	Cell #	Control	Tw/Td Td/Tw
13:16	#141	40/50	07/20/01			to Cell I	to <del>II</del>	
15:56	#213	40/50				I	to (37/50.2, 35.2)	

\* Tw/Td ; \*\* Temperature of cooling water

16:04	#216	39/50					to (33/50.2, 30.2)	
16:04	#219	38/50				to II		
16:23	#226	37/50						

Program terminated for C-22 testing.

07/20/01

8:38						Program restarted	File: 0720a	
8:38		42/50	(water in wet bulb gone)			water refilled,	to Cell I	
9:08						to Cell II		
10:01		37/50				to Cell I (at 11:14)		
11:39		37/50					to (37/50.2, 35.2)	
12:04		40/50						
12:28						to Cell II		
14:28		40/50				to Cell I		
14:56							to (39/50.2, 35.2)	
15:49		41/50						
18:56	#205					to Cell II		
16:57	#234					to Cell I		
17:16	#243	41/50					to (40/50.2, 35.2)	
17:25		42/50				to Cell II		
17:29		42/50						
17:33	#250					Program terminated for C-22 testing		

J. J. Jones 7/15/02

07/21/01								
14:20		42/50	(after water added)			Program started,	File: 0721a	Cell I
14:32		41/50				to Cell II		
14:41	#4							to (41.2/50.2, 35.2)

07/23/01								
08:15	#1205	43/50	(water in w. bulb gone)			to Cell I		
10:02		42/50	(after water refilled)			to Cell II		
12:17						to Cell I		
12:33		41/50						to (42.5/50.2, 35.2)
12:49		42/50						
12:57		43/50						to (43.1/50.2, 35.2)
13:24		43/50				to Cell I		
13:55		43/50						to (43.0/50.2, 35.2)

16:55								to (43/50.2, 35.2, 3.16)
-------	--	--	--	--	--	--	--	--------------------------

$\frac{3}{16}$  — means  $\frac{3}{16}$  turn of the valve from being closed.

must turn this value to achieve high RH (> 43/50)

17:20		43/50						
07/24/01								
8:40		44/50				Program restarted	File: 0724a	Cell I
9:49		45/50				Program restarted after bug fixing,	File: 0724b	
9:52						to Cell II		

9:56	#15	44/50						to (45.0/50.2, 35.2, 3.16)
------	-----	-------	--	--	--	--	--	----------------------------

J. J. Jones 7/15/02




10:44 #40 45/50 to cell I  
 12:22 #88 45/49.5 to II  
 12:39 #97 45/50 to I to (47/50.2, 35, 3/16)  
 12:56 #105 46/50  
 13:18 #116 46/50  
 13:21  
 13:31 46/50  
 15:06 to II  
 15:19 46/50 to (49.5/50.2, 35.2, 1/8)  
 15:53 47/51 to I  
 16:32 49/51  
 16:48 #223 49/51 to II set (43/50, 35.2, 3/16)  
 program terminated for C-22 testing.

07/26/01  
 8:15 44/50 program started, File: 0725A, with cell I  
 8:47 #15 44/50 to II  
 9:16 44/50 to (49.5/50.2, 35.2, 1/8)  
 12:32 50/51 to I  
 14:30 #144 to (50.2/50.2, 35, 1/2)  
 15:01 #211 51/52 to (50.2/49.5, 35, 1/8)  
 15:10 #216 51/52 to I  
 17:16 #221 to II to (50.2/51.5, 35.2, 1/8)  
 17:18 to  
 17:26 #236 to (46.5/50.2, 35.2, 3/16 turn)

J. Yang 7/26/01

07/26/01  
 07:07 47/50  
 07:34 program started, 0726A, with cell I  
 08:01 #14, 47/50 to cell II  
 13:37 47/50. program restarted (it was stopped at 8:38 by a bug)  
 13:47 #4 to cell I  
 16:35 #84 47/50 to cell II  
 16:40 #92 program was terminated




J. Yang.

07/27/01

Test Results

0718a\_0725a.xls

9/4/01




$f_g$   
(Hz)

- ◆ 100000
- 63095.73
- ▲ 39810.72
- × 25118.86
- ✱ 15848.93
- 10000
- + 6309.573
- 3981.072
- 2511.886
- 1584.893
- 1000

J. Yano  
7/15/02

0718a\_0725a.xls

9/4/01




$f_g$   
(Hz)

- ◆ 100000
- 63095.73
- ▲ 39810.72
- × 25118.86
- ✱ 15848.93
- 10000
- + 6309.573
- 3981.072
- 2511.886
- 1584.893
- 1000

use data of low frequency (1000 - 6310 Hz)  
see page 42.


J. Yano  
7/15/02

0718a\_0725a.xls



*J. Young  
7/18/02*


0718a\_0725a.xls



*J. Young  
7/18/02*

0718a\_0725a.xls


9/4/01



II has been always lower than I  
 Why it is higher than I this time?  
 J. Yans  
 7/18/02

0718a\_0725a.xls

9/4/01



J. Yans  
 7/18/02

0718a\_DP.xls Data

File: 0718a\_DP.xls

Data time	RH (%)	KNO3, Cell I		KCl, Cell II	
		Z, ohm	log Z	Z, ohm	log Z
7/18/01 15:00	80	2.70E+02	2.43E+00	2.20E+02	2.34E+00
7/18/01 17:00	63	2.50E+05	5.40E+00		
7/19/01 9:00	59	2.70E+05	5.43E+00	9.50E+04	4.98E+00
7/19/01 15:00	55	6.00E+05	5.78E+00	2.00E+05	5.30E+00
7/20/01 9:00	48	2.00E+06	6.30E+00	1.00E+06	6.00E+00
7/20/01 12:00	55	8.00E+05	5.90E+00	4.50E+05	5.65E+00
7/20/01 15:00	59	5.00E+05	5.70E+00	2.50E+05	5.40E+00
7/21/01 14:00	61	4.00E+05	5.60E+00		
7/21/01 19:00	61			1.20E+05	5.08E+00
7/23/01 8:00	63	4.00E+05	5.60E+00	2.00E+06	6.30E+00
7/23/01 15:00	66	3.00E+05	5.48E+00	1.50E+06	6.18E+00
7/23/01 17:00	73	1.00E+05	5.00E+00		
7/24/01 12:00	73			3.50E+02	2.54E+00
7/24/01 17:00	90	2.80E+02	2.45E+00	1.00E+03	3.00E+00
7/25/01 9:00	71				

why so high.  
see comments on page 69

From Graph, Page 71:

DH KNO3=77%, Literature\*=84.8% Difference= +7.8%  
 DH KCl=71%, Literature\*= 81.2% Difference= +10.2

Should be corrected by adding  $(+7.8+10.2)/2=$  9 %

Data time	Corrected RH (%)	MgCl2-NaCl, Cell I		MgCl2-SH2O, Cell II	
		Z, ohm	log Z	Z, ohm	log Z
7/18/01 15:00	89	2.70E+02	2.43E+00	2.20E+02	2.34E+00
7/18/01 17:00	72	2.50E+05	5.40E+00		
7/19/01 9:00	68	2.70E+05	5.43E+00	9.50E+04	4.98E+00
7/19/01 15:00	64	6.00E+05	5.78E+00	2.00E+05	5.30E+00
7/20/01 9:00	57	2.00E+06	6.30E+00	1.00E+06	6.00E+00
7/20/01 12:00	64	8.00E+05	5.90E+00	4.50E+05	5.65E+00
7/20/01 15:00	68	5.00E+05	5.70E+00	2.50E+05	5.40E+00
7/21/01 14:00	70	4.00E+05	5.60E+00		
7/21/01 19:00	70			1.20E+05	5.08E+00
7/23/01 8:00	72	4.00E+05	5.60E+00	2.00E+06	6.30E+00
7/23/01 15:00	75	3.00E+05	5.48E+00	1.50E+06	6.18E+00
7/23/01 17:00	82	1.00E+05	5.00E+00		
7/24/01 12:00	82			3.50E+02	2.54E+00
7/24/01 17:00	99	2.80E+02	2.45E+00	1.00E+03	3.00E+00
7/25/01 9:00	80				

J.Y.  
8/24/02


\* L. Greenspan, J. Research of N. B. S., Vol 81A, No1, 1977, Page 89-96

File: 0718a-DP.xls was accidentally deleted. All data in that file are on this page.

J. Young 10/22/01

0718a\_DP.xls Uncorrected\_Chart


Data see Page 70



DP for KNO3= 84.78 % at 50cC  
 DP for KCl = 81.20% at 50cC

J. Young  
7/15/02


0718a\_DP.xls Corrected\_Chart



Data for KCl are not good.  
reason not known!  
See comment on page 68.

J. Yang  
7/15/02

0718a\_DP.xls Corrected\_Chart\_KNO3



J. Yang  
7/15/02

07/26/01

NaCl - KNO<sub>3</sub> sat. solution (30 ml water)

}	NaCl	12.8 mol/kg	→ x 30 ml	0.38 mol	→ x 58.44	22.44 g
	KNO <sub>3</sub>	6.1 mol/kg	→ x 30 ml	0.18 mol	→ x 101.10	18.14 g

solubility of calculated see: Page 87.

NaCl: lot # 010166 - Fisher

KNO<sub>3</sub>: lot # 7028 KDPO by Mallinckrodt.

water → Deionized. 30 ml.

Heated and stirred. then cooled to 22°.

Cell I, 2 mm wider than base of PTFE, KNO<sub>3</sub>-NaCl soln.

Arc cell

Cell II, same as base of PTFE, NaNO<sub>3</sub> soln.

Saturated NaNO<sub>3</sub> solution made from lot # 961722A

17:13 Cells installed in chamber.

Program started. File Name: 0726C  
set to (36/50, 30°, 1/2)

#13, #14 with Cell II

after #16 with Cell I

J. Yama  
7/15/02

Z, r

Cell #

control Tw/Td

17:43 # 40/50 (56%) to (36/50, 30°, 1/2)

17:56 40/50 to (36/50, 15°, 1/2)

18:10 35/44 <sup>J. Yama.</sup> ~~to II~~ 07/27/01

07/27/01

8:04 #435 42/50 (Wet bulb) <sup>J. Yama.</sup> ~~to II~~ 07/27/01 to Cell II  
dried out  
water added.

8:43 42/50 / Z<sub>II</sub> = 10<sup>6</sup> Ω

10:05 #489 34/50 to Cell I.

11:02 #515 34/50 to (36/50, 25°, 1/2)

11:19 37/50 (45%)

11:34 37/50 to Cell II

13:06 37/50 to Cell I

13:31 #586 37/50 to (37/50, 25°, 1/2)

13:53 37/50 to (38.5/50, 30°, 1/2)

14:12 #606 39.5/50 to (38.5/50, 25°, 1 turn)

14:18 38/50 too low to (38.5/50, 27.5°, 1/2)

14:21 38.5/50

14:24 38.5/50

14:38 39/50 (52%)

14:41 to Cell II

14:46 #623 to (40/50, 27.5°, 1/2)

J. Yama  
7/15/02

power of medium

07/27/01

14:48 to cell I

14:51 39/50 to (40/50, 30<sup>o</sup>, 1/4)

14:58 40/50

15:14 40/50 to cell II

16:11 #665 40/50 to cell I

16:18 to (41.5/50, 31.8<sup>o</sup>, 1/4)

16: #685 40.5/50 to (36/50.5, 15<sup>o</sup>, 1/4)

17:11 38/50 program terminated and restarted: File: 07276.

07/30/01

12:04 #1815 43/50, (wet. Bulb dried) to cell II  
water added.

12:13 34/50

12:15 to (42/50.5, 30<sup>o</sup>, 1/4)

12:26 40.5/50

12:30 to cell I

12:56 40/50

13:05 to cell II

13:14 to cell I

13:24 #1855 40/50 to (43/50.5, 34, 1/4)

13:33 #1860 40.5/50

14:00 41/50 to cell II

14:25 41/50 to (43/50.5, 34, 1/4)

14:32 43/50

14:38 43/50 to cell I

14:58 43/50 to cell II

J. Jones  
7/15/02

15:13 43/50 to cell I to (45/50, 34, 1/8 turn)

15:26 #1417 46/50 to cell II

15:40 46/50 to cell I

15:44 46/50 Z<sub>I</sub> = 143 $\Omega$  to (49/49.5, 34.2, 1/8)

15:52 #1430 50/51 to cell II

15:54 Z<sub>I, 1k</sub> = 312 $\Omega$  to (45/50.5, 34, 1/8)

16:08 53/53 to cell I

16:22 #1447 45/50 Z<sub>I</sub> = 215 $\Omega$

16:33 41.5/50 to cell II

16:38 to (47/50.5, 34.2, 1/4)

16:43 43/50 to cell I

16:51 43/50

16:59 43/50 to cell II

17:09 43/50 to cell I

17:22 43/50

17:29 43/50 ← to cell II

17:41 41/50

17:47 to cell I

7/31/01

8:31 #2452 40.5/50.5 Z<sub>I, 1k</sub> = 240 $\Omega$  to cell II  
Z<sub>I, 1k</sub> = 2470 $\Omega$

8:45 #2459 to (38.5/50.5, 25<sup>o</sup>, 1/4)

8:53 37/50

9:45 38/50

10:47 38/50 to cell I

11:56 37/50 to cell II.

13:33 37/50 to cell I

15:01 36/50 to cell II

J. Jones  
7/15/02



15:25 32/50 to cell I to (36/50, 1+2, 1)

16:04 32/50

16:07 to cell II

20:43 #2802 32/50 to cell I

08/01/01

8:20 #3117 32/50-44 to cell II


10:25 34/50-44 Chamber opened.

7/18/02  
20/11/02  
amf.t

66.84 Pa  
69.04 Pa  
Char  
VW03 - Nucl - I  
NaA03 - II


0726c00\_27b07.xls

8/2/01




20/11/02  
amf.t

Enlarged Figure for 7/27/01 sea page 83




20/11/02  
Enlarged




20/11/02  
Enlarged

0726c00\_27b07.xls Chart1



2. yans  
7/18/02

0726c00\_27b07.xls Chart1



2. yans  
7/18/02

0726c\_DP.xls Data

July 28/29/01

0726c\_DP

Point #	Data time	RH (%)	KNO3, Cell I		NaNO3, Cell II	
			Z, ohm	log Z  KNO3-NaCl	Z, ohm	log Z  NaNO3
6	7/30/01 15:30	80	2.50E+02	2.40E+00	3.50E+02	2.54E+00
7	7/30/01 16:30	67	2.20E+02	2.34E+00	3.50E+02	2.54E+00
8	7/31/01 8:20	59	2.10E+02	2.32E+00	2.50E+04	4.40E+00
9	7/31/01 15:20	48	3.50E+04	4.54E+00	2.00E+05	5.30E+00
10	7/31/01 20:00	38	3.50E+05	5.54E+00	1.80E+06	6.26E+00
5	7/30/01 12:00	36	1.00E+06	6.00E+00	5.00E+06	6.70E+00
	7/30/01 13:00	55			1.00E+05	5.00E+00
1	7/27/01 10:00	36	8.00E+05	5.90E+00	1.80E+06	6.26E+00
2	7/27/01 13:00	45	2.50E+05	5.40E+00	8.00E+05	5.90E+00
3	7/27/01 14:50	52	2.50E+04	4.40E+00	1.50E+05	5.18E+00
4*	7/27/01 17:10	56	2.50E+02	2.40E+00	8.00E+04	4.90E+00

\* 4 estimated

From Graph, Page 85:  
DH NaNO3=63%, Literature\*=69.04% Difference= +6%

Should be corrected by adding

July 28/29/01

6%

Data time	Corrected RH (%)	KNO3, Cell I		NaNO3, Cell II	
		Z, ohm	log Z  KNO3-NaCl	Z, ohm	log Z  NaNO3
7/30/01 15:30	86	2.50E+02	2.40E+00	3.50E+02	2.54E+00
7/30/01 16:30	73	2.20E+02	2.34E+00	3.50E+02	2.54E+00
7/31/01 8:20	65	2.10E+02	2.32E+00	2.50E+04	4.40E+00
7/31/01 15:20	54	3.50E+04	4.54E+00	2.00E+05	5.30E+00
7/31/01 20:00	44	3.50E+05	5.54E+00	1.80E+06	6.26E+00
7/30/01 12:00	42	1.00E+06	6.00E+00	5.00E+06	6.70E+00
7/30/01 13:00	61			1.00E+05	5.00E+00
7/27/01 10:00	42	8.00E+05	5.90E+00	1.80E+06	6.26E+00
7/27/01 13:00	51	2.50E+05	5.40E+00	8.00E+05	5.90E+00
7/27/01 14:50	58	2.50E+04	4.40E+00	1.50E+05	5.18E+00
7/27/01 17:10	62	2.50E+02	2.40E+00	8.00E+04	4.90E+00


\* L. Greenspan, J. Research of N. B. S., Vol 81A, No1, 1977, Page 89-96

Correction

This correction -> close to 9% in Page 70


July 28/29/01

0726c\_DP.xls Uncorrected\_Chart



July 28/29/01

0726c\_DP.xls Corrected\_Chart



Dp of KNO<sub>3</sub>-NaCl is 65%

J. J. Jones 7/15/02

Calculation of solubility and activity of water (RH) of salt mixture, using OLI 6.5 - esp.

Calculation method same as that described in Page 69 of Scientific Notebook # 430

500 g, 7/31/01

	H2O	KCl	KNO3	NANO3	NaCl
Input (mol/hr)	55.55		10		10
Stream	In	In	out	out	out
Phase	Aqueous	Solid	Aqueous	Solid	Vapor
Temperature	25	25	50	50	50
Pressure, ε	1	1	0.081288	0.081288	0.081288
pH	6.94875		6.58556		
Total mol/h	74.41866	10.5463	77.8098	8.84085	7.56E-04
Flow Units	mol/hr	mol/hr	mol/hr	mol/hr	mol/hr
H2O	55.55		55.5492		7.56E-04
HCL	1.20E-13		1.21E-12		1.02E-13
HNO3	5.03E-10		3.01E-09		3.99E-14
KCL	0.01487		0.03059		
NANO3	0.023849		0.027113		
OHION	2.03E-08		4.84E-08		
HION	1.98E-08		4.54E-08		
KION	3.06917		4.53978		
NAION	6.3458		6.56167		
NO3ION	3.05019		3.54266		
CLION	6.35478		6.55819		
NaCl		3.630347		3.41122	
KNO3		6.915953		5.42963	
KOH.1H2C					
KOH.2H2C					
KOH					
NaOH.1H2					
NaOH					
Total g/hr	1684.82	911.394	1847.89	748.315	0.013611
Volume, L/l	1.26718	0.429607	1.37928	0.352433	0.246216
Enthalpy, c	-4.75E+06	-1.17E+06	-4.92E+06	-9.72E+05	-43.5155
Density, g/l	1329.59	2121.46	1339.75	2123.28	0.055279
Vapor fract					1
Solid fract			1		1
Organic fra					
Osmotic Pr	529.757		607.38		
Redox Pot,					
E-Con, 1/ol	0.343559		0.520565		
E-Con, cm:	21.7675		27.3113		
Abs Visc, c	2.05076		1.38261		
Rel Visc	2.30236		2.52663		
Ionic Stren:	9.40797		11.0933		
<b>RH</b>					<b>0.665953</b>

solubility mol/kg water

Calculated RH = 66.6%

higher than 65%, measured.

J. J. Jones 7/15/02

Calculation for Pure NaNO<sub>3</sub>, Next page

Continued from page 87.

	H2O	KCl	KNO3	NaNO3	NaCl
	55.55			15	
Stream	In	In	out	out	out
Phase	Aqueous	Solid	Aqueous	Solid	Vapor
Temperature	25	25	50	50	50
Pressure, $\epsilon$	1	1	0.079427	0.079427	0.079427
pH	7.139		6.735		
Total mol/h	77.23457	4.08411	82.99731	1.18782	7.06E-04
Flow Units	mol/hr	mol/hr	mol/hr	mol/hr	mol/hr
H2O	55.55		55.5493		7.06E-04
HCl					
HNO3	1.50E-09		6.82E-09		1.34E-13
KCl					
NaNO3	0.147172	4.08411	0.176412	1.18782	
OHION	2.27E-08		4.77E-08		
HION	2.12E-08		4.09E-08		
KION					
NaION	10.7687		3.6358		
CLION					
NaCl					
KNO3					
KOH.1H2C					
KOH.2H2C					
KOH					
NaOH.1H2					
NaOH					
Total g/hr	1928.55	347.128	2174.7	100.959	0.01271
Volume, L/l	1.36768	0.153807	1.53019	0.044733	0.235314
Enthalpy, c	-4.96E+06	-4.57E+05	-5.25E+06	-1.32E+05	-40.6356
Density, g/l	1410.09	2256.9	1421.2	2256.9	0.054012
Vapor fract					1
Solid fractic		1		1	
Organic fra					
Osmotic Pr	459.309		617.253		
Redox Pot,					
E-Con, 1/ol	0.193679		0.275214		
E-Con, cm:	17.6594		18.9457		
Abs Visc, c	2.63048		2.06255		
Rel Visc	2.95321		3.76916		
Ionic Stren:	10.7607		13.6258		
			0.650707		

RU

DP of  $\text{NaNO}_3 = 65.1\%$  which is lower than published data by 4% (see page 84).  
 J. Young  
 7/18/02

8/01/01

Test with  $\text{NaNO}_3 - \text{NaCl}$  (I),  $\text{NaNO}_3 - \text{NaCl} - \text{HNO}_3$  (II)

Solution preparation.

	sol*	double sol	in 30 ml	FW g	g	lot #
I: $\text{NaNO}_3$	6.8	13.6	0.408 mol	84.44	34.68 g	96.1772A
$\text{NaCl}$	3.8	7.6	0.278	58.4	13.3 g	006924
II: $\text{NaNO}_3$	7.0	14	0.42	84.44	31.7 g	96.1772A
$\text{NaCl}$	3.8	8	0.24	58.44	14.0 g	006924
$\text{HNO}_3$	2.1	5	0.12	101.1	11.16 g	7028 WPD Mallinckrodt

Fisher

\* — solubility at 25°C, calculated, see page 90.

Solution was well mixed, heated, then cooled to 25°C.

J. Young

7/18/02

File: knachno3\_b.xls

See page 37.

mod by winter

Chemistry Model: KNACINO3

In		out		out		out	
Aqueous		Solid		Aqueous		Solid	
25		25		50		50	
1		1		1		1	
Temperature, C	7.05619			0.071613	0.071613	0.071613	0.071613
Pressure, atm	80.97702	17.2391		88.04159	13.6949	8.56E-04	
pH	7.05736			6.69067			
Total mol/hr	8.57E-14	55.55		55.5491		8.56E-04	
Flow Units	7.28E-10	7.56E-13		7.56E-13		8.43E-14	
H2O	0.004038	0.00702		0.00702		1.39E-13	
HCL	0.090808	3.014599		0.110546			
HNO3	1.32E-08	2.81E-08		2.81E-08			
KCL	1.25E-08	3.15586		3.15586			
NANO3	1.99957	13.0316		13.0316			
OHION	10.6665	6.8578		6.8578			
HION	3.76789	7.996413		7.996413			
KION	6.228088	6.8371		6.8371			
NAION	2017.49	1428.67		2354.1	1092.05	0.015412	
VOLUME, L/hr	1.40793	0.665097		1.62524	0.513001	0.316508	
ENTHALPY, cal/hr	-5.13E+06	-1.89E+06		-5.50E+06	-1.48E+06	-49.2752	
DENSITY, g/L	1432.95	2148.07		1448.47	2128.74	0.048694	
VAPOR FRACTION							
SOLID FRACTION							
ORGANIC FRACTION							
OSMOTIC PRES, atr	631.531	765.088		765.088			
REDUX POT, volts							
E-CON, 1/ohm-cm	0.259518	0.360716		0.360716			
E-CON, cm2/ohm-1	12.1795	15.2369		15.2369			
ABS VISC, cP	2.96375	2.29093		2.29093			
REL VISC	3.32737	4.18652		4.18652			
IONIC STRENGTH	12.6567	16.1757		16.1757			
ACTIVITY OF WATER		0.586695		0.586695			

solubility

RH

solubility

RH

knachno3

In		out		out		out	
Aqueous		Solid		Aqueous		Solid	
25		25		50		50	
1		1		1		1	
Temperature, C	7.05736			0.075628	0.075628	0.075628	0.075628
Pressure, ε	76.5806	9.43681		81.38682	7.01728	7.56E-04	
pH	7.05736			6.67905			
Total mol/hr	8.85E-14	55.55		55.5492		7.56E-04	
Flow Units	7.74E-10	8.32E-13		8.32E-13		7.14E-14	
H2O	0.004038	0.00702		0.00702		1.25E-13	
HCL	0.090808	3.014599		0.110546			
HNO3	1.32E-08	2.81E-08		2.81E-08			
KCL	1.25E-08	3.15586		3.15586			
NANO3	1.99957	13.0316		13.0316			
OHION	10.6665	6.8578		6.8578			
HION	3.76789	7.996413		7.996413			
KION	6.228088	6.8371		6.8371			
NAION	2017.49	1428.67		2354.1	1092.05	0.015412	
VOLUME, L/hr	1.40793	0.665097		1.62524	0.513001	0.316508	
ENTHALPY, cal/hr	-5.13E+06	-1.89E+06		-5.50E+06	-1.48E+06	-49.2752	
DENSITY, g/L	1432.95	2148.07		1448.47	2128.74	0.048694	
VAPOR FRACTION							
SOLID FRACTION							
ORGANIC FRACTION							
OSMOTIC PRES, atr	631.531	765.088		765.088			
REDUX POT, volts							
E-CON, 1/ohm-cm	0.259518	0.360716		0.360716			
E-CON, cm2/ohm-1	12.1795	15.2369		15.2369			
ABS VISC, cP	2.96375	2.29093		2.29093			
REL VISC	3.32737	4.18652		4.18652			
IONIC STRENGTH	12.6567	16.1757		16.1757			
ACTIVITY OF WATER		0.586695		0.586695			

solubility

RH

08/02/01 program started, file: 0802a cell II

8:29  $13\frac{1}{2}/2$  (Room)  $Z_{II} = 10^5 \Omega$  in arr

8:32  $13\frac{1}{2}/2$  (Room)  $Z_I = 10^4 \Omega$  changed to cell I in arr

9:03 installed in chamber to  $(50/50, 312, \frac{1}{8})$

9:09  $3/39$  to cell II

9:22 #11  $50/51$   $Z_{II} = 357 \Omega$  to cell I

9:27  $Z_I = 329 \Omega$  to  $(44/50, 252, \frac{1}{8})$

9:29  $48/51$  to cell II

9:36  $41/50$  to  $(44/50, 343, \frac{1}{8})$

9:42  $43/50$  to cell I

9:53 to cell II

10:19 #41  $44/50$  to cell I

12:21 #104  $44/50$  to cell II to  $(42/50, 325, \frac{1}{8})$

13:29  $41/50$  to cell I

15:12 to cell II

15:16  $41/50$  (52%) to  $(39/50, 272, \frac{1}{8})$

15:55  $39/50$  (52%) to cell I

17:20  $39/50$  to cell II

8/03/01

8:17 #728  $39/50$  (52%)  $Z_{II} = 214 \Omega$  to cell I

8:20  $37/49$   $Z_I = 30651 \Omega$

9:22  $39/50$  to  $(36.5/50, 272, \frac{1}{8})$

8:27  $37/50$  to cell II

8:40  $36/50$  to  $(36.5/50, 202, \frac{1}{8})$

9:24 to cell II

10:10  $36/50$  (42%) J. Young

11:04 #811  $36/50$  to cell I 7/15/02

12:20 to cell II

8/3/01

14:07 #899 to cen I

14:21 36/50 to (37.5/50.5, 23<sup>2</sup>, 1/4)

14:24 37/50

14:37 37/50

16:03 #954 37/50

16:58 37/50 to II

17:06 37/50 to (42/50.5, 32.5, 1/4)

17:30 #997 41/50 to I

17:45 41/50 to (32.5/50.5, 17<sup>2</sup>, 1/4)

17:46 to II

17:48 #1006 37/50

08/06/01

8:06 #2752 48/50 (dried)  $Z_{II} = 10^4 \Omega$  to I

8:10  $Z_I = 10^6 \Omega$  add water, to II

8:20 33/50

10:49 34/50 to (42/50.5, 32.5, 1/4)

11:24 34/50 to I

16:40 41/50

16:37 41/50 to II

16:39 41/50 to (30 41/50.5, 30<sup>2</sup>, 1/4)

08/07/01

9:02 40/50 to I

16:15 40/50 to II to (39/50, 25<sup>2</sup>, 1/4)

16:42 37/50

16:55 37/50

17:17 37/50

J. Young  
7/15/02

08/08/01

8:27 38/50 (48%)  $Z_{II} = 10^4 \Omega$  to I

10:49 38/50  $Z_I = 10^5 \Omega$

11:01 inserted Cole parmer probe to II.  
just calibrated. see page 94.

11:07 38/50 - 37/50

11:08 37.7 (47%)  $51.36\% - 48.72\% *$ , 47.68<sup>2</sup>  
Ave: 50%

\* — Hygrometer - Colparmer.

11:20 Experiment terminated. open circuit.

Data processing go to page 103

J. Young  
7/15/02



*It was noticed that the Cole-Parmer Humidity Probe was not good!*

# SOUTHWEST RESEARCH INSTITUTE

6220 CULEBRA ROAD • POST OFFICE DRAWER 28510 • SAN ANTONIO, TEXAS, 78228-0510 • TEL (210) 522-5215 • FAX (210) 522-3692

**To:** Lietai Yang, Div 20, Bld 189

**From:** Walt Hill, Institute Calibration Laboratory Supervisor

**CC:** Rodney Weber, Institute Quality Assurance Manager

**Date:** Aug. 6, 01

**Subject:** Out-of-tolerance Notice

The purpose of this notice is to alert you of a condition, which could have caused erroneous measurements. The as-found readings are provided for your evaluation to determine if the instrument had any impact on your operations and if further action is required. If we can be of assistance, please contact the Calibration Laboratory at 522-5215.

**Manufacturer:** Cole Parmer      **Model:** 03313-66

**Description:** Thermohygrometer      **Serial Number:** 21189381

**Asset Number:** 008788      **User ID Number:**

**Last Calibration:** 7/03/01

**Date Received for Service:** Aug. 3, 01      **Work Order Number:** 444044659

**Service Requested:** Calibrate before use

**Remarks:** TI probe was contaminated. Cleaned and recalibrated TI. E-mail was sent to Mr. Yang on 8/3/01.

### AS-FOUND DATA

PARAMETER OR FUNCTION	APPLIED OR NOMINAL VALUE	INSTRUMENT READING	INSTRUMENT ERROR	INSTRUMENT TOLERANCE
Humidity	39.95 % RH	46.9 % RH	6.9 % RH	0.5 % RH
	55.05 % RH	65.8 % RH	10.8 % RH	0.5 % RH
	70.07 % RH	85.1 % Rh	15.1 % RH	0.5 % RH
	77.39 % RH	94.7 % RH	17.3 % RH	0.5 % RH


Page 1 of 1 **OUT OF TOLERANCE** Reviewed by *Walt Hill*

*J. Yang*  
*7/15/02*

Cole\_Parmer\_Calibration.xls Data

Data from Calibration Lab Standard reading	Standard	50oC	20oC	22.5oC
		39.95	46.88	
		55.05	65.83	
		70.07	85.08	
		77.39	94.67	
	50		63.24	
	33.85			41.93
	74.8			91.25

Cole\_Parmer\_Calibration.xls Chart1



*J. Yang*      *7/15/02*

12/13/02

Work Order 444044659	Mfr. Cole-Parmer	Technician R Dykstra
Asset No. 008788	Model 03313-66	Procedure CL-13
Serial No. 21189381	Type Thermohygrometer	Cal Date 8/6/01

Remarks: The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor of  $k=2$ , which provides a level of confidence of approximately 95%. The test limits are specified in meter instructions. As found humidity out of tolerance. Probe was contaminated. Cleaned probe and recalibrated.

Parameter	Test Point	As Found	Error	Test Limits	Uncertainty
Humidity	39.95 % RH	46.88 % RH	6.93 % RH	1.5 % RH	0.5 % RH
	55.05 % RH	65.83 % RH	10.78 % RH	1.5 % RH	0.5 % RH
	70.07 % RH	85.08 % RH	14.69 % RH	1.5 % RH	0.5 % RH
	77.39 % RH	94.67 % RH	17.28 % RH	1.5 % RH	0.5 % RH

Parameter	Test Point	As Left	Error	Test Limits	Uncertainty
Humidity	24.99 % RH	24.43 % RH	0.56 % RH	1.5 % RH	0.5 % RH
	50.1 % RH	50.86 % RH	0.76 % RH	1.5 % RH	0.5 % RH
	75.22 % RH	75.92 % RH	0.70 % RH	1.5 % RH	0.5 % RH

Parameter	Test Point	As Found	Error	Test Limits	Uncertainty
Temperature	50.0 °C	50.01 °C	.01 °C	0.2 °C	0.2 °C

Parameter	Test Point	As Left	Error	Test Limits	Uncertainty
Temperature	25.03 °C	25.09 °C	.06 °C	0.2 °C	0.2 °C



Southwest Research Institute  
6220 Culebra Road  
San Antonio, TX 78238  
(210) 522-5215  
Department of Quality Assurance  
Calibration Laboratory



Certificate #  
0972-01

# Certificate of Calibration

7 August 2001

Issued to: DARRELL DUNN DIV20 B57  
 Manufacturer/Model: COLE-PARMER 03313-66  
 Description: THERMOHYGROMETER  
 Serial Number: 21189381  
 Asset Number: 008788  
 Work Order Number: 444044659

This certifies the above item was calibrated in compliance with MIL-STD-45662A and ANSI/NCSL Z540-1-1994. The results of this calibration relate only to the individual item as described above. Standards used in this calibration, described in the referenced calibration procedure with associated uncertainties or tolerances, are traceable to the National Institute of Standards and Technology (NIST). Supporting documentation relative to traceability is on file and available for examination upon request. This certificate is not to be reproduced, except in full, without the written approval of the Southwest Research Institute Department of Quality Assurance Calibration Laboratory.

This laboratory is accredited by the American Association for Laboratory Accreditation (A2LA) and the results of this calibration certificate were determined in accordance with the terms of accreditation unless stated otherwise below.

The uncertainty of the calibration was sufficient to determine that the item met the manufacturer's published specifications unless stated otherwise below.

Ambient Conditions: Temperature: 76.0 Degrees Fahrenheit Humidity: 49 % RH

Calibration Date: 6 Aug 01 Calibration Procedure:

Condition as Received: OUT OF TOLERANCE

Condition as Released: IN TOLERANCE

Remarks:

Approved by: Walt Hill

Walt Hill, Supervisor  
Institute Calibration Laboratory

Measurements performed by: Roger Dykstra

Roger Dykstra, Technician

probe was contaminated and was cleaned, recalibrated.

2. J. Jones, 7/13/02

J. Jones, 7/13/02



Southwest Research Institute  
 6220 Culebra Road  
 San Antonio, TX 78238  
 (210) 522-5215  
 Department of Quality Assurance  
 Calibration Laboratory



Certificate #  
 0972-01

# Certificate of Calibration

7 August 2001

Issued to: STEVE MACLEOD DIV03 B151  
 Manufacturer/Model: ROTRONIC HT205R  
 Description: HYGROMETER  
 Serial Number: R37096  
 Asset Number: 006989  
 Work Order Number: 444044678

*Borrowed from  
 Division 03*

This certifies the above item was calibrated in compliance with MIL-STD-45662A and ANSI/NC SL Z540-1-1994. The results of this calibration relate only to the individual item as described above. Standards used in this calibration, described in the referenced calibration procedure with associated uncertainties or tolerances, are traceable to the National Institute of Standards and Technology (NIST). Supporting documentation relative to traceability is on file and available for examination upon request. This certificate is not to be reproduced, except in full, without the written approval of the Southwest Research Institute Department of Quality Assurance Calibration Laboratory.

This laboratory is accredited by the American Association for Laboratory Accreditation (A2LA) and the results of this calibration certificate were determined in accordance with the terms of accreditation unless stated otherwise below.

The uncertainty of the calibration was sufficient to determine that the item met the manufacturer's published specifications unless stated otherwise below.

Ambient Conditions: Temperature: 77.0 Degrees Fahrenheit Humidity: 46 % RH

Calibration Date: 7 Aug 01 Calibration Procedure: CL-19, 5/99

Condition as Received: IN TOLERANCE

Condition as Released: IN TOLERANCE

Remarks:

Approved by: Walt Hill  
 Walt Hill, Supervisor  
 Institute Calibration Laboratory

Measurements performed by: Roger Dykstra  
 Roger Dykstra, Technician

*Figures 7/18/02*

Date Calibrated 8-7-01  
 Technician R Dykstra Calibration Procedure HT200-03/30/95  
 Unit Under Test Humidity/Temperature Transmitter  
 Manufacturer Rotronic Model HT205R SN R37096 ASN 6989  
 Work Order 444044678

STEP	FUNCTION OR RANGE	APPLIED	TOLERANCE		MEASURED VALUES		P/F
			MIN	MAX	AS FOUND	RELEASED	
NOTE: Manufacturer state tolerance of 1.0 to 1.5 % is based on less than a 4:1 Test Uncertainty Ratio. Limit all calibrations of this unit to $\geq \pm 2\%RH$ .							
Standards: Thunder Scientific 2500 Humidity 10 to 95%RH $\pm 0.5\%$ Azonic A1011/wA12001 Temperature -183 to 480 °C $\pm 0.02$ °C Tektronix DM50X DMM (2ea) 0-20V, $\pm 0.01\%$ + 2 counts Power Supply (10 to 35 VDC), Uncertainty N/A Set power supply to $\approx 18$ V							
See page 14 of Mfg. manual for conversion formula for voltage to %RH and Temperature.							
	Humidity	35 % RH	33%RH	37%RH	34.5	Same v	P
	Temperature	20 °C	19.7°C	20.3°C	19.9	as found	P
	Humidity	50%	48%RH	52%RH	50.6		P
	Temperature	25 °C	24.7°C	25.3°C	24.8		P
	Humidity	80%RH	78%RH	82%RH	78.6		P
	Temperature	20 °C	19.7°C	20.3°C	20.1		P
	Humidity	20 %RH	18%RH	22%RH	19.4		P
	Temperature	20 °C	19.7°C	20.3°C	20.0		P

*CL 171  
 2-80ms 7/18/02*

Aug. 7, 2001

Attached is what you will need to operate the TI you borrowed from Div 03. Both instruments the Cole-Parmer and the Rotronic are ready to be picked up on 8/8/01.

Roger Dykstra Calibration Lab.

Equipment needed to operate Rotronic Hygromer series 1200.  
DC Power Supply- 18 vdc  
DVM 0-5 VDC -should have a good accuracy on 20 VDC scale or less

The TI needs to be hooked up to a DC Power supply set to approximately 18 VDC.  
Connect positive to the red wire terminal and negative to the black wire terminal on the 4-wire connector.

Humidity output is monitored at the 2-wire terminal. 0-5 VDC is equal to 0 to 100 % RH.  
Positive is clear wire, negative is black (2 wire terminal.)  
(Voltage reading / 5)\*100 = % RH.

Temperature output is monitored at the 4-wire terminal. 0-5 VDC is equal to 0 to 200 Degree F.  
Positive is the white wire, negative is the green wire (4-wire terminal.)  
(Voltage reading / 5)\*200 = Degree F.

*Dykstra*  
*8/8/01*

*ordered*      *K/03*

08/10/01 04:02 Fisher Scien p 2/2

AUG 10, 2001		F I S H E R   S C I E N T I F I C		O R D E R   A C K N O W L E D G E M E N T		PAGE: 1	
THANK YOU FOR YOUR ORDER... ANY QUESTIONS REGARDING THIS ORDER, CALL 1-(800) 766-7000 TO SPEAK TO A CUSTOMER SERVICE REPRESENTATIVE							
ORDER NBR: H12220574		ACCT NBR: 784214-001		ORDER DATE: AUG 10, 2001		REL NBR: 50578	
CUST PURCH NBR: 1.20.01402.561		CALLER: LEITAL YANG		PHONE NBR: (210)522-2483			
SHIP TO: SOUTHWEST RES INSTITUTE BLDG- 189 5704 BUSINESS PARK SEND C OF A W/CHEMICAL SAN ANTONIO TX 78218 ATTN: LEITAL YANG BL. 189				SOLD TO: SOUTHWEST RES INSTITUTE P O DRAWER 28510 SAN ANTONIO TX 78228			
LN	QTY	UN	CATALOG NBR	DESCRIPTION	UNIT PRC	TOTAL AMT	
				SEND C OF A WITH CHEMICALS			
001	3	EA	P263500	POT NITRATE CERTIF ACS 500G 3 SHIPPING FROM CHS	29.31	87.93	
				MERCHANDISE TOTAL:		87.93	
THE INVOICE AMOUNT MAY INCLUDE TRANSPORTATION, TAX, AND OTHER MISC ORDER CHARGES AS DETERMINED BY YOUR CONTRACT.							

*Certificate see page 102*

*Dykstra 8/20/01*

*Entry made to Chemical Inventory and  
Certs.*

FROM: Fisher Scientific TO: Extension PAGES: 3

**Fisher Scientific Company**  
Chemical Manufacturing Division

**Certificate of Analysis**

Fisher Scientific's Quality System is  
Certified to ISO9002 (1994) standard by DNV  
Cert. # 96-HOU-AQ-8052

1 Reagent Lane  
Fairlawn, NJ 07410  
Phone: (201) 796-7100 Fax: (201) 796-1329

Catalog Number	P263	Report Date	5/1/2001	Mfg. Date	12/7/2000
Lot Number	010186	Sample ID	P263.010186.B1.		
Description	POTASSIUM NITRATE, A.C.S.				

This is to certify that units of the above mentioned lot number were tested and found to comply with the specifications of the grade listed. The following are the actual analytical results obtained:

Result Name	Units	Specifications	Test Value
APPEARANCE	REPORT	colorless to white crystals not wet	FINE WHITE CRYSTALS
ASSAY	%	99.0 Minimum	99.2000
CALCIUM IN %	%	0.005 Maximum	0.0005
CHLORIDE	%	0.002 Maximum	0.0010
HEAVY METALS (AS Pb)	PPM	3 Maximum	0.30
INSOLUBLE MATTER	%	0.005 Maximum	0.0020
IODATE (IO3) IN PPM	PPM	5 Maximum	5.0000
IRON (Fe)	PPM	3 Maximum	0.300
MAGNESIUM IN %	%	0.002 Maximum	0.0002
NITRITE	%	0.001 Maximum	0.00050
PH 5% SOLN @ 25DEG C		4.5 to 8.5	5.70
PHOSPHATE (PO4)	PPM	5 Maximum	0.500
SODIUM (Na) IN %	%	0.005 Maximum	0.00100
SULFATE (SO4)	%	0.003 Maximum	0.0003

**CERTIFIED BY**

*Ant Doral*  
Lab Manager Fair Lawn

*Edgar E. Hesse*  
Lab Manager BPF


Note: The data listed is valid for all package sizes of this lot of product, expressed as a extension of the catalog number listed above. If there are any questions with this certificate, please call Chemical Services at (800) 227-6701

*J. Yancy 8/15/02*

Continued from page 93.


10  
0802b00\_08.xls Chart1

*J. Yancy 8/15/02*




*J. Yancy 7/15/02*

0802b00\_03.xls Chart1



20/5/12  
J. Yama


0802b00\_03.xls Chart1



20/5/12  
J. Yama


0802b00\_10.xls Chart1

1075  
two points  
each freq.




J. Young  
7/18/02

0802b00\_10.xls Chart1



J. Young  
7/18/02

0802b00\_10.xls Chart1



J. Yano  
7/15/02

0802b\_DP\_b.xls Data

un corrected.

J. Yano 4/7/06

08/02/01  
J. Yano

#	Data time	RH (%)	KNO <sub>3</sub> , Cell I		NaNO <sub>3</sub> , Cell II	
			Z, ohm	log Z	Z, ohm	log Z
1	8/2/01 12:00	71	3.20E+02	2.51E+00	2.50E+02	2.40E+00
2	8/2/01 15:00	58	3.20E+02	2.51E+00	2.50E+02	2.40E+00
3	8/3/01 8:00	52	3.50E+04	4.54E+00	3.00E+02	2.48E+00
5	8/3/01 17:00	47	3.50E+05	5.54E+00	7.50E+04	4.88E+00
4	8/3/01 14:00	42	7.30E+05	5.86E+00	2.00E+05	5.30E+00
6	8/6/01 8:00	35	1.00E+07	7.00E+00	9.00E+05	5.95E+00
7	8/6/01 17:00	58			2.00E+02	2.30E+00
8	8/7/01 15:00	55	2.00E+04	4.30E+00	2.00E+02	2.30E+00
9	8/8/01 11:00	48	1.50E+05	5.18E+00		

corrected RH.

Should be corrected by average of previous test  
page 70: 7.8%, 10.2%, Page 84: 6%

J. Yano 4/7/06

J. Yano 08/02/01

#	Data time	Corrected RH (%)	KNO <sub>3</sub> , Cell I		NaNO <sub>3</sub> , Cell II	
			Z, ohm	log Z	Z, ohm	log Z
1	8/2/01 12:00	79	3.20E+02	2.51E+00	2.50E+02	2.40E+00
2	8/2/01 15:00	66	3.20E+02	2.51E+00	2.50E+02	2.40E+00
3	8/3/01 8:00	60	3.50E+04	4.54E+00	3.00E+02	2.48E+00
5	8/3/01 17:00	55	3.50E+05	5.54E+00	7.50E+04	4.88E+00
4	8/3/01 14:00	50	7.30E+05	5.86E+00	2.00E+05	5.30E+00
6	8/6/01 8:00	43	1.00E+07	7.00E+00	9.00E+05	5.95E+00
7	8/6/01 17:00	66			2.00E+02	2.30E+00
8	8/7/01 15:00	63	2.00E+04	4.30E+00	2.00E+02	2.30E+00
9	8/8/01 11:00	56	1.50E+05	5.18E+00		


\* L. Greenspan, J. Research of N. B. S., Vol 81A, No1, 1977, Page 89-96

J. Yano

7/15/02




0802b\_DP\_b.xls Uncorrected\_Chart



*J. Yans*  
 7/12/02


0802b\_DP\_b.xls Corrected\_Chart




*corrected RH %*

*Modeled Result: (page 90)*  
 KNO<sub>3</sub>-NaNO<sub>3</sub> NaCl: 58.7%  
 NaNO<sub>3</sub>-NaCl : 62%  
 Close to the above measured.


*J. Yans*  
 7/12/02



J. Yams  
7/18/02



J. Yams  
7/18/02



J. Gunn  
7/18/02

Measurement for

- I — MgCl<sub>2</sub> + NaCl
- II — MgCl<sub>2</sub> · 6H<sub>2</sub>O

- I — Added excess of MgCl<sub>2</sub> · 6H<sub>2</sub>O, lot # 00036, heated and cooled, added NaCl, lot # 006924 cooled at room Temp. overnight.
- II — Saturated MgCl<sub>2</sub> · 6H<sub>2</sub>O lot # 00036 cooled at room Temp. for many days.

- I — Paper 1.5 mm narrower than the PTFE base
- II — Paper 1 mm narrower than the PTFE base.

11:45 Cells I and II installed in chamber

11:55 program started. File: 0808a with cell II.  
to (31.5/50, 17<sup>o</sup>, 1/1).  
to (32.1/50, 17<sup>o</sup>, 1/1)

12:26	#18	34/50	Z <sub>II</sub> = 10 <sup>2.6</sup> = 5010 <sup>2</sup>	to Cell #I
12:38		34/50	Z <sub>I</sub> = 10 <sup>2.4</sup> = 4530 <sup>2</sup>	
13:48		34/50		to II to (31.5/50, 15 <sup>o</sup> , 1 turn)
14:15		33/50		to
16:22		33/50		to I

J. Gunn  
7/18/02

08/09/01  
 8:35 #621 48/50 dried out  $Z_I = 10^{2.7}$  to II  
 8:38  $Z_{II} = 2584$   
 8:43 to I  
 8:46 water added to wet bulb  $Z_I = 533$   
 8:47 33/50 to (30/50.5, 13.5°, 1/4)  
 10:25 #677 32/50  $Z_I = 560$  to II  
 10:53  $Z_{II} = 628$  to (29/50.5, 11.5°, 1 turn)  
 12:14 32/50 to (28.0/50.9, 11, 1/4)  
 12:41 31/50  
 17:04 31.3/49.8 (28%)  $Z_{II} = 652$  to I to (28/51.0, 0°, 2 turn)  
 17:26 31/50 to high power.

8/10/01  
 9:08 #1362 49/51 (dried out)  $Z_I = 10^{2.75}$  to II  
 9:15 add water to wet bulb.  $Z_{II} = 725 \Omega$  to I  
 $Z_I = 671 \Omega$

9:19 #1367 31/51  
 13:24 32/51 to (28/51.0, 0°, Full)  
 \* — Fully open the valve of cooling

17:00 32/51  
 08/13/01  
 8:58 #3513 49/51 (dried out)  $Z_I = 10^5$  to II  
 9:13  $Z_{II} = 1523 \Omega$  to I  
 $Z_I = 146241 \Omega$   
 9:15 water added.  
 J. Yano 7/18/02

9:40 30/51  
 11:11 30.5/51  
 15:26 31/51 to (0°/51.2, 0°, Full)  
 15:36 28/51 to II to (15.2/51.2, 0°, Full)  
 17:21 27/51  $Z_{II} = 10^{4.7} \Omega$  to I  
 17:23  $Z_I = 408416 \Omega$  to (22.2/51.2, 0°, Full)

8/14/01  
 8:34 #4205 49/51 (dried out)  $Z_I = 1.09 \times 10^6$  (1k) to II  
 8:44 49/51  $Z_{II} = 730492 \Omega$  water added.  
 8:46 23/51 (at 1k)  
 8:55 38/44 system was Flooded.  $Z_{II} = 629 \Omega$  by raising the reservoir water.

8:57 37/49.2 (41%)  
 13:06 29/51  $Z_{II} = 10^{3.0} \Omega$  to I  
 17:07 29/51

8/12/01  
 8:18 49/51 (dried out)  $Z_I = 10^{4.5}$  to II  
 8:28  $Z_{II} = 10^{4.9} \Omega$   
 11:48 49/51  $Z_{II} = 10^{5.2}$  to I  
 $Z_{I, 1k} = 517146 \Omega$


11:59 to (26/51, 0°, Full)  
 12:09 39/51  $Z_{I, 1k} = 506200$   
 \* all resistance values given more than two digits are obtained at 1k Hz frequency.  
 J. Yano 7/18/02

68/15/01  
 13:14 30/51  $Z_{I,1k} =$  to Cell II  
 16:45 30/51 to I  
 8/16/01  
 8:25  $Z_I = 10^{5.4}$  to II  
 8:32 water added to wet bulb  
 8:37 30/51 to (27.5, 02, Full)  
 11:49 31/51  $Z_{II} = 10^{4.3}$  (dropping) to I  
 14:14 to II to (29/51, 02, Full)  
 15:15 32/51  
 16:47 program stopped !  
 16:49 program restarted 0816a (filename)

08/17/01  
 8:02 49/51 (dried out)  $Z_0 = 10^{4.7}$  to I  
 8:25 water added  $Z_I = 10^5$   
 11:20 32/51  $Z_I = 10^5$  to II to (30.5/51, 02, Full)  
 11:55 33/51  $Z_{II}$  dropping quickly  
 12:37 33/51 to I  
 15:38 33/51  $Z_I = 10^{2.6}$  to II  
 17:00 33/51 to I  
 17:14 33/51 to (9/51, 02, Full)  
 17:29 28/51  
 17:37 28/51  
 08/20/01 (Monday)  $Z_I = 10^6$  to II  
 8:15 water added to wet bulb  $Z_{II} = 2.2 \times 10^6$   
 8:35 27/51  
 9:06 27/51 test terminated

*J. Young*  
 7/15/02

File: 0808a00\_0816a06.xls Chart1




*J. Young*  
 7/15/02

Chart1

8/29/01

0808a00\_



fa  
(1/2)


- ◆ 100000
- 63095.73
- ▲ 39810.72
- × 25118.86
- ✱ 15848.93
- 10000
- + 6309.573
- 3981.072
- 2511.886
- 1584.893
- 1000

20/11/02  
smg/af

Chart1

8/29/01


0808a00\_




power  
channel  
fa  
(1/2)

- ◆ 100000
- 63095.73
- ▲ 39810.72
- × 25118.86
- ✱ 15848.93
- 10000
- + 6309.573
- 3981.072
- 2511.886
- 1584.893
- 1000

20/11/02  
smg/af




20/8/02 2/18/02




20/8/02 2/15/02

File: 0808a00\_1



29/51  
7/18/02


File: 0808a00\_1



29/51  
7/18/02




File: 0808a00\_1




2. years  
7/15/02

File: 0808a00\_1



2. years  
7/15/02

File: 0808a00\_1



*J. Yans*  
2/15/02

0808a\_DP\_b.xls Data

	Data time	RH (%)	MgCl2-NaCl, Cell I		MgCl2.6H2O, Cell II	
			Z, ohm	log Z	Z, ohm	log Z
				MgCl2-NaCl		MgCl2.6H2O
2	8/8/01 15:00	32.5	5.40E+02	2.73E+00	5.80E+02	2.76E+00
3	8/9/01 9:00	32.5	5.30E+02	2.72E+00	5.80E+02	2.76E+00
4	8/9/01 11:00	29	5.70E+02	2.76E+00	6.10E+02	2.79E+00
5	8/9/01 17:00	27	6.10E+02	2.79E+00	6.50E+02	2.81E+00
6	8/10/01 9:00	27	6.50E+02	2.81E+00	7.00E+02	2.85E+00
7	8/14/01 9:00	20	1.30E+06	6.11E+00	8.00E+05	5.90E+00
8	8/15/01 15:00	23	3.50E+05	5.54E+00	2.00E+05	5.30E+00
9	8/16/01 9:00	23	4.00E+05	5.60E+00	2.50E+05	5.40E+00
10	8/16/01 15:00	25	2.00E+05	5.30E+00	1.60E+05	5.20E+00
11	8/17/01 8:00	26	1.00E+05	5.00E+00	1.30E+05	5.11E+00
12	8/17/01 17:00	30	6.00E+02	2.78E+00	5.70E+02	2.76E+00
13	8/20/01 8:00	16	4.00E+06	6.60E+00	3.00E+06	6.48E+00


From Fig in Page 130, DP of the MgCl2.6H2O is 26.5%  
Literature value is 30.54\*  
Correction by 30.54-26.5 =

	Data time	Corrected RH (%)	MgCl2-NaCl, Cell I		MgCl2.6H2O, Cell II	
			Z, ohm	log Z	Z, ohm	log Z
				MgCl2-NaCl		MgCl2.6H2O
2	8/8/01 15:00	36.54	5.40E+02	2.73E+00	5.80E+02	2.76E+00
3	8/9/01 9:00	36.54	5.30E+02	2.72E+00	5.80E+02	2.76E+00
4	8/9/01 11:00	33.04	5.70E+02	2.76E+00	6.10E+02	2.79E+00
5	8/9/01 17:00	31.04	6.10E+02	2.79E+00	6.50E+02	2.81E+00
6	8/10/01 9:00	31.04	6.50E+02	2.81E+00	7.00E+02	2.85E+00
7	8/14/01 9:00	24.04	1.30E+06	6.11E+00	8.00E+05	5.90E+00
8	8/15/01 15:00	27.04	3.50E+05	5.54E+00	2.00E+05	5.30E+00
9	8/16/01 9:00	27.04	4.00E+05	5.60E+00	2.50E+05	5.40E+00
10	8/16/01 15:00	29.04	2.00E+05	5.30E+00	1.60E+05	5.20E+00
11	8/17/01 8:00	30.04	1.00E+05	5.00E+00	1.30E+05	5.11E+00
12	8/17/01 17:00	34.04	6.00E+02	2.78E+00	5.70E+02	2.76E+00
13	8/20/01 8:00	20.04	4.00E+06	6.60E+00	3.00E+06	6.48E+00

\* L. Greenspan, J. Research of N. B. S., Vol 81A, No1, 1977, Page 89-96


*J. Yans*  
2/14/06

0808a\_DP\_b.xls Uncorrected\_Chart



II MgCl<sub>2</sub> · 6H<sub>2</sub>O  
 I MgCl<sub>2</sub> · 6H<sub>2</sub>O - NaCl have the same DP.  
 J - Yang 7/15/02

0808a\_DP\_b.xls Corrected\_Chart




L. Yang 7/29/01

Pages 1 - 132 copied and given  
to QA for record.


J. Yarns

10/09/01



J. Yarns

7/15/02



J. Young  
7/15/02


### Deliquescence Point Measurement

Objective: To determine the deliquescence points of salts experimentally.

The experimental set-up is shown below:

Several bottles are placed in the oven. The bottles with pure salts are used as controls for the method and water.

File: DP\_Hyfrometer



J. Young 2/14/06  
02/08/01 J. Young