

**From:** Willie Lee  
**To:** jkeck@sibley.org  
**Date:** Tue, Dec 12, 2006 11:37 AM  
**Subject:** Sibley Memorial Hospital, License Amendment, Control 139587

REQUEST FOR ADDITIONAL INFORMATION

LICENSEE: Sibley Memorial Hospital  
LICENSE NUMBER: 08-07398-03  
DOCKET NUMBER: 030-14754  
CONTROL NUMBER: 139587  
DATE: December 12, 2006

ATTN: Jordie Keck, Medical Physicist and Radiation Safety Officer

The following information is needed for completion of the license amendment request received on 10/16/06 regarding the addition of Cesium-131 brachytherapy sources to your NRC license:

1. Please describe how you will assure that members of the public will receive less than 500 milli-Rem exposure from a released patient containing Cesium-131 brachytherapy sources.

Please note that you may not reply to this request by return e-mail. Your prompt reply must be in writing by letter, signed by a member of Senior Management, which can be sent by facsimile to (610) 337-5269. Contact me if there are any questions regarding this request.

Willie J. Lee, Health Physicist  
US Nuclear Regulatory Commission  
Region 1  
475 Allendale Road  
King of Prussia, PA 19406  
(610) 337-5090 (Phone)  
(610) 337-5269 (FAX)

**Mail Envelope Properties** (457EDABE.DE7 : 7 : 26753)

**Subject:** Sibley Memorial Hospital, License Amendment, Control 139587  
**Creation Date** Tue, Dec 12, 2006 11:37 AM  
**From:** Willie Lee

**Created By:** WJL1@nrc.gov

**Recipients**

sibley.org  
jkeck (jkeck@sibley.org)

**Post Office**

**Route**  
sibley.org

<b>Files</b>	<b>Size</b>	<b>Date &amp; Time</b>
MESSAGE	1880	Tuesday, December 12, 2006 11:37 AM

**Options**

**Expiration Date:** None  
**Priority:** Standard  
**ReplyRequested:** No  
**Return Notification:** None

**Concealed Subject:** No  
**Security:** Standard