

PORTABLE INSTRUMENT ACCOUNTABILITY FORM

RP Supervisor Review W. B. [Signature] (1)

COPY

PORTABLE/GAMMA P.R.I.S. & SOURCE CHECK FORM

E 600 Meter Type SPA-3 Detector Type DETS 51847 Detector Number 277 Source ID 201077.6 Net Acceptance Criteria - 20% 301616.4 Net Acceptance Criteria + 20%

PRE USE CHECKS								POST USE CHECKS							
Date	Time	Audible Check	Alarm Check	BKG Counts	SRC Counts	Net Counts	Int	Date	Time	Audible Check	Alarm Check	BKG Counts	SRC Counts	Net Counts	Int
7-7-04	0540	SAT	N/A	5380	252000	246620	WS								①
7-10-06	0545	SAT	N/A	5060	250000	244940	WS								①
7-11-6	0645	SAT	NA	4950	253000	248550	CF	7-11-6	12:45	SAT	NA	5290	252000	246710	CF
7-12-04	0545	SAT	N/A	5280	251000	245720	WS	7-12-6	16:45	SAT	ND	7140	253000	245860	CF
7-13-06	0545	SAT	N/A	5200	252000	246800	WS	7-13-6	1530	SAT	N/A	5300	250000	244700	WS
7-14-06	0540	SAT	N/A	4780	251000	246220	WS	7-14-6	1600	SAT	NA	5730	252000	246270	CF
7-17-06	0545	SAT	N/A	5550	249000	249450	WS	7-17-6	1405	SAT	NA	4920	252000	247050	CF
7-18-6	0610	SAT	NA	5070	251000	245910	CF	7-18-6	1600	SAT	NA	4700	251000	246600	CF
7-19-06	0550	SAT	N/A	4460	250000	245540	WS	7-19-6	1635	SAT	NA	4760	251000	246020	CF
7-20-06	0545	SAT	N/A	4600	252000	247400	WS	7-20-06	0550	SAT	N/A	5880	251000	245920	WS
7-21-06	0550	SAT	N/A	5080	251000	245920	WS								①
7-24-06	0610	SAT	N/A	4860	251000	246140	WS								①
7-25-06	0545	SAT	N/A	4700	251000	246300	WS	7-25-6	1530	SAT	NA	5460	251000	245540	CF
7-26-06	0600	SAT	N/A	5450	250000	244850	WS	7-26-6	1620	SAT	NA	4940	250000	245260	CF
7-27-04	0600	SAT	N/A	5490	249000	243510	WS	7-27-6	1525	SAT	NA	5480	251000	245520	CF

RP Supervisor Review: [Signature] (1) NOT USED

(1) If any post-use source check failures occur, ensure that the condition is documented by a Condition Report.

COPY

PORTABLE INSTRUMENT ACCOUNTABILITY FORM

RP Supervisor Review

(1)

- DPF-8504.1
rev. 17

COPY

PORTABLE INSTRUMENT ACCOUNTABILITY FORM

RP Supervisor Review W. Bulge (1)

COPY

PORTABLE/GAMMA RAIS & SOURCE CHECK FORM

Meter Type C-600 Detector Type SPA-3 Detector Number 725329 Source ID 277 Net Acceptance Criteria - 20% 188/20 Net Acceptance Criteria + 20% 282/80

PRE USE CHECKS								POST USE CHECKS							
Date	Time	Audible Check	Alarm Check	BKG Counts	SRC Counts	Net Counts	Int	Date	Time	Audible Check	Alarm Check	BKG Counts	SRC Counts	Net Counts	Int
8-16-06	0700	SAT	N/A	5080	222000	216920	VS	8-16-06	071030	SAT	N/A	4280	219000	214720	VS
8-17-06	0550	SAT	N/A	4870	218000	213430	VS	8-17-06	1640	SAT	N/A	4930	213000	208010	VS
8-18-06	0545	SAT	N/A	5110	220000	214890	VS								①
8-19-06	0610	SAT	N/A	5500	224000	215500	VS								①
8-21-06	0610	SAT	N/A	4710	219000	214280	VS	8-21-06	1725	SAT	N/A	5550	221000	215450	VS
8-22-06	0550	SAT	N/A	5080	221000	215920	VS	8-22-06	0505	SAT	N/A	5700	222000	216400	VS
8-22-06	0905	SAT	N/A	5000	222000	216100	VS	8-22-06	1630	SAT	N/A	4370	218000	213030	VS
8-23-06	0730	SAT	N/A	5030	221000	215970	VS	8-23-06	1705	SAT	N/A	4740	221000	216260	VS
8-24-06	0600	SAT	N/A	5400	219000	213510	VS	8-24-06	1140	SAT	N/A	4740	222000	217220	VS
8-25-06	0600	SAT	N/A	4720	219000	214280	VS	8-25-06	1550	SAT	N/A	4860	233000	228040	VS
8-26-06	0440	SAT	N/A	5310	232000	226690	VS	8-26-06	1550	SAT	N/A	4690	235000	230010	VS
8-28-06	0550	SAT	N/A	5280	231000	225720	VS	8-28-06	1640	SAT	N/A	5310	232000	226690	VS
8-29-06	0550	SAT	N/A	4440	233000	228260	VS	8-29-06	1545	SAT	N/A	5040	232000	226690	VS
8-29-06	0600	SAT	N/A	4520	233000	228710	VS								①
8-29-06	0550	SAT	N/A	4490	232000	227510	VS	8-31-06	1620	SAT	N/A	3800	231000	227000	VS

RP Supervisor Review: Wayne B. B. (1) NOT USED

(1) If any post-use source check failures occur, ensure that the condition is documented by a Condition Report.

COPY

PORTABLE/GAMMA RATE & SOURCE CHECK FORM

6600 Meter Type
30A-3 Detector Type
2056 Detector Number
277 Source ID
187748 Net Acceptance Criteria - 20%
272422 Net Acceptance Criteria + 20%

PRE USE CHECKS						POST USE CHECKS					
Date	Time	Audible Check	Alarm Check	BKG Counts	SRC Counts	Net Counts	Date	Time	Audible Check	Alarm Check	BKG Counts
8-24-06	0600	SMT	N/A	5490	221000	217000	8-24-06	1655	SMT	N/A	6020
8-25-06	0610	SMT	N/A	5000	217000	229000	8-25-06	1600	SMT	N/A	4940
8-26-06	0645	SMT	N/A	4770	229000	224230	8-26-06	1545	SMT	N/A	5800
8-28-06	0550	SMT	N/A	5350	229000	223050	8-28-06	1545	SMT	N/A	5800
8-28-06	0550	SMT	N/A	5880	229000	223920	8-28-06	1545	SMT	N/A	5800
8-30-06	0620	SMT	N/A	5210	229000	223970	8-30-06	1705	SMT	N/A	4890
8-31-06	0550	SMT	N/A	5280	227000	221720	8-31-06	1705	SMT	N/A	4890
9-1-06	0545	SMT	N/A	4180	229000	224820	9-1-06	1645	SMT	N/A	4250
9-5-06	0610	SMT	N/A	8010	227000	221990	9-5-06	1645	SMT	N/A	4250
9-6-06	0610	SMT	N/A	4540	227000	222400	9-6-06	1645	SMT	N/A	4250
9-7-06	0600	SMT	N/A	4860	225000	220140	9-7-06	1145	SMT	N/A	4310

RP Supervisor Review: W. J. Smith

If any post-use source check failures occur, ensure that the condition is documented by a Condition Report.

DPF-8504.5
Rev. 17

COPY

PORTABLE INSTRUMENT ACCOUNTABILITY FORM

RP Supervisor Review *W. B. J.* (1)

COPY

PORTABLE/GAMMA FIELD SOURCE CHECK FORM

6600 Meter Type
 20A-3 Detector Type
 2056 Detector Number
 277 Source ID
 187148 Net Acceptance Criteria - 20%
 272622 Net Acceptance Criteria + 20%

PRE USE CHECKS								POST USE CHECKS							
Date	Time	Audible Check	Alarm Check	BKG Counts	SRC Counts	Net Counts	Int	Date	Time	Audible Check	Alarm Check	BKG Counts	SRC Counts	Net Counts	Int
8-24-06	0600	SAT	N/A	5490	221000	215510	MS	8-24-06	1655	SAT	N/A	6070	217000	210530	MS
8-25-06	0610	SAT	N/A	5200	217000	211400	MS	8-25-06	1600	SAT	N/A	4940	229000	224060	MS
8-26-06	0645	SAT	N/A	4770	229000	224230	MS	8-26-06	1545	SAT	N/A	5820	231000	225200	MS
8-28-06	0530	SAT	N/A	5350	229000	223650	MS								MS
8-29-06	0550	SAT	N/A	5880	229000	223420	MS								MS
8-30-06	0600	SAT	N/A	5210	229000	223390	MS								MS
8-31-06	0550	SAT	N/A	5280	227000	221720	MS	8-31-06	1705	SAT	N/A	4890	230000	225110	MS
9-1-06	0545	SAT	N/A	4180	229000	224820	MS								MS
9-5-06	0610	SAT	N/A	5210	227000	221990	MS	9-5-06	1045	SAT	N/A	4250	225000	220750	MS
9-6-06	0610	SAT	N/A	4540	227000	222460	MS	9-6-06	1645	SAT	N/A	4620	229000	224980	MS
9-7-06	0600	SAT	N/A	4860	225000	220140	MS	9-7-06	1145	SAT	N/A	4360	232000	227690	MS
						N									
						A									

RP Supervisor Review: W. J. [Signature] (1) NOT USED

(1) If any post-use source check failures occur, ensure that the condition is documented by a Condition Report.

COPY

PORTABLE/GAMMA RATE & SOURCE CHECK FORM

C-600
Meter
Type

SPT-3
Detector
Type

725329
Detector
Number

217
Source
ID

188120
Net
Acceptance
Criteria
- 20%

282160
Net
Acceptance
Criteria
+ 20%

PRE USE CHECKS								POST USE CHECKS							
Date	Time	Audible Check	Alarm Check	BKG Counts	SRC Counts	Net Counts	Int	Date	Time	Audible Check	Alarm Check	BKG Counts	SRC Counts	Net Counts	Int
9-1-06	0545	SAT	N/A	340	23000	22960	US	9-1-06	1110	SAT	N/A	4220	23160	226780	US
9-5-06	0610	SAT	N/A	5300	23000	225700	US								①
9-10-06	0610	SAT	N/A	370	23000	236770	US	9-10-06	1456	SAT	N/A	4410	23800	230190	US
9-1-06	0600	SAT	N/A	4570	224000	219430	US	9-1-06	1200	SAT	BA	4400	230000	225100	US
9-12-06	0730	SAT	N/A	520	231000	225780	US								①
9-12-06	0815	SAT	N/A	510	231000	225890	US	9-12-06	1615	SAT	N/A	4270	215000	210730	US
9-13-06	0910	SAT	N/A	4530	217000	212480	US	9-13-06	1525	SAT	BA	3660	218000	212320	US
9-14-06	0610	SAT	N/A	7680	217000	212320	US								①
9-15-06	0605	SAT	N/A	4850	217000	212050	US								①
9-18-06	0600	SAT	N/A	4700	216000	211240	US	9-18-06	1840	SAT	N/A	4540	218000	213460	US
9-18-06	0650	SAT	N/A	4050	217000	212950	US	9-19-06	1725	SAT	N/A	6040	219000	212960	US
9-20-06	0640	SAT	N/A	4446	216000	211560	US								①
9-26-06	1155	SAT	N/A	4010	216000	211390	US								①
10-2-06	0640	SAT	N/A	4350	216000	211650	US								①
10-3-06	0610	SAT	N/A	4790	216000	211210	US	10-3-06	0805	SAT	BA	6160	216000	210820	US

RP Supervisor Review:

[Signature]

(1) If any post-use source check failures occur, ensure that the condition is documented by a Condition Report.

- (1) NOT USED
(2) Setup w/12 cable