

MATERIALS LICENSE

Pursuant to the Atomic Energy Act of 1954, as amended, the Energy Reorganization Act of 1974 (Public Law 93-438), and Title 10, Code of Federal Regulations, Chapter I, Parts 30, 31, 32, 33, 34, 35, 36, 39, 40, and 70, and in reliance on statements and representations heretofore made by the licensee, a license is hereby issued authorizing the licensee to receive, acquire, possess, and transfer byproduct, source, and special nuclear material designated below; to use such material for the purpose(s) and at the place(s) designated below; to deliver or transfer such material to persons authorized to receive it in accordance with the regulations of the applicable Part(s). This license shall be deemed to contain the conditions specified in Section 183 of the Atomic Energy Act of 1954, as amended, and is subject to all applicable rules, regulations, and orders of the Nuclear Regulatory Commission now or hereafter in effect and to any conditions specified below.

<p style="text-align: center;">Licensee</p> <p>1. Department of the Army U.S. Army Communications- Electronics Command AMSEL-SF-RER</p> <p>2. Fort Monmouth, New Jersey 07703-5024</p>	<p>In accordance with the letter dated February 15, 2001,</p> <p>3. License number 29-01022-14 is amended in its entirety to read as follows:</p> <hr/> <p>4. Expiration date October 31, 2003</p> <hr/> <p>5. Docket No. 030-29741 Reference No.</p>
--	---

6. Byproduct, source, and/or special nuclear material	7. Chemical and/or physical form	8. Maximum amount that licensee may possess at any one time under this license
A. Cobalt 60	A. []	A. []
B. Cobalt 60	B. []	B. []
C. Cobalt 60	C. []	C. []
D. Krypton 85	D. Sealed sources (USAEA Dwg. No. B124-12-8)	D. Not to exceed 6 millicuries per source and 120 curies total
E. Strontium 90	E. []	E. []

DD/8

Information in this record was deleted in accordance with the Freedom of Information Act, exemptions 2
2006-0238

**MATERIALS LICENSE
SUPPLEMENTARY SHEET**

License Number
29-01022-14

Docket or Reference Number
030-29741

Amendment No. 21

6. Byproduct, source, and/or special nuclear material

7. Chemical and/or physical form

8. Maximum amount that licensee may possess at any one time under this license

F. Strontium 90

F. Sealed sources (ECOM Dwg. No. SM-B-509048)

F. Not to exceed 150 microcuries per source and 45 millicuries total

G. Strontium 90

G.

G.

H. Strontium 90

H. Sealed sources (3M Dwg. No. 12-1921-0474-8)

H. Not to exceed 36 microcuries per source and 18 millicuries total

I. Cesium 137

I.

I.

J. Cesium 137

J.

J.

K. Plutonium 239

K. Electroplated sources (Eberline Instrument Corp., Model 594-1)

K. Not to exceed 23 micrograms (1.4 microcuries) per set and 0.0115 grams total

L. Americium 241

L. Sealed sources (Amersham Radiochemical Center, Amersham Code 2084)

L. Not to exceed 10 millicuries per source and 50 millicuries total

M. Americium 241

M. Sealed sources (Amersham Model AMR 8122)

M. Not to exceed 1 microcurie per source and 100 microcuries total

N. Americium 241

N. Sealed sources (Amersham Model AMRB 8152)

N. Not to exceed 10 microcuries per source and 50 microcuries total

O. Americium 241

O. Sealed sources (Amersham Model AMRB 1659)


O. Not to exceed 20 microcuries per source and 100 microcuries total

**MATERIALS LICENSE
SUPPLEMENTARY SHEET**

License Number
29-01022-14

Docket or Reference Number
030-29741

Amendment No. 21

6. Byproduct, source, and/or special nuclear material	7. Chemical and/or physical form	8. Maximum amount that licensee may possess at any one time under this license
P. Thorium 230	P. Electroplated source (Eberline Instrument Corp., Model No. CS-12)	P. Not to exceed 0.98 micrograms (20 nanocuries) per source and 1 milligram total
Q. Thorium 232	Q. Metal foils	Q. Not to exceed 2.7 grams (300 nanocuries) per source and 4.05 kilograms total
R. Plutonium 239	R. Electroplated sources (Eberline Instrument Corp., Model No. CS-1)	R. Not to exceed 163 nanograms (10 nanocuries) per source and 1 gram total
S. Thorium 232	S. Solid (Thorium Fluoride coating on optical systems)	S. Not to exceed 3 grams (0.330 microcuries) per optical system and 40 kilograms total
T. Cesium 137		
U. Cesium 137		
V. Hydrogen 3	V. Tritiated paint in Lensatic Compasses (NSN 6605-00-846-7618)	V. 120 millicuries per compass and 480 curies total
W. Hydrogen 3	W. Sealed light sources in Lensatic Compasses (NSN-6605-00-151-5337)	W. 190 millicuries per compass and 5700 curies total
X. Depleted Uranium	X. Metal	X. 1870 kilograms

9. Authorized use:

- A. through R. Calibration and operational checking of radiation detection instrumentation.
- S. Optical coating on thermal imaging devices.

**MATERIALS LICENSE
SUPPLEMENTARY SHEET**License Number
29-01022-14Docket or Reference Number
030-29741

Amendment No. 21

- T. Use in FEMA Model CDV-794 calibrators for instrument calibrations.
- U. Use in FEMA Model CDV-790 calibrators for instrument calibrations.
- V. and W. Possession, storage, and distribution to any U.S. Department of Defense elements and reserve components including the U.S. Army, U.S. Navy, U.S. Marine Corps, U.S. Air Force, Defense Supply Agency, the National Guard and the Air National Guard.
- X. Shielding for CDV-794 instrument calibrator.

CONDITIONS

10. Licensed material may be used only at the licensee's facilities located at Fort Monmouth, New Jersey, and at Department of Defense installations anywhere in the United States.
11. A. Licensed material shall only be used by, or under the supervision and in the physical presence of, individuals who have completed the training described in application dated July 20, 1992 and letter dated May 1, 1998, with enclosures.
- B. The Radiation Safety Officer for this license is Richard J. Lovell.
12. A. Sealed sources shall be tested for leakage and/or contamination at intervals not to exceed the intervals specified in the certificate of registration issued by the U.S. Nuclear Regulatory Commission under 10 CFR 32.210 or under equivalent regulations of an Agreement State.
- B. Notwithstanding Paragraph A of this Condition, sealed sources designed to primarily emit alpha particles shall be tested for leakage and/or contamination at intervals not to exceed 3 months.
- C. In the absence of a certificate from a transferor indicating that a leak test has been made within the intervals specified in the certificate of registration issued by the U.S. Nuclear Regulatory Commission under 10 CFR 32.210 or under equivalent regulations of an Agreement State, prior to the transfer, a sealed source received from another person shall not be put into use until tested and the test results received.
- D. Sealed sources need not be tested if they contain only hydrogen-3; or they contain only a radioactive gas; or the half-life of the isotope is 30 days or less; or they contain not more than 100 microcuries of beta- and/or gamma-emitting material or not more than 10 microcuries of alpha-emitting material.

**MATERIALS LICENSE
SUPPLEMENTARY SHEET**License Number
29-01022-14Docket or Reference Number
030-29741

Amendment No. 21

- E. Sealed sources need not be tested if they are in storage and are not being used; however, when they are removed from storage for use or transferred to another person and have not been tested within the required leak test interval, they shall be tested before use or transfer. No sealed source shall be stored for a period of more than 10 years without being tested for leakage and/or contamination.
- F. The leak test shall be capable of detecting the presence of 0.005 microcurie (185 becquerels) of radioactive material on the test sample. If the test reveals the presence of 0.005 microcurie (185 becquerels) or more of removable contamination, a report shall be filed with the U.S. Nuclear Regulatory Commission in accordance with 10 CFR 30.50(c)(2), and the source shall be removed immediately from service and decontaminated, repaired, or disposed of in accordance with Commission regulations.
- G. Tests for leakage and/or contamination, including leak test sample collection and analysis, shall be performed by the licensee or by other persons specifically licensed by the U.S. Nuclear Regulatory Commission or an Agreement State to perform such services.
- H. Records of leak test results shall be kept in units of microcuries.
13. Sealed sources or detector cells containing licensed material shall not be opened or sources removed from source holders by the licensee.
14. The licensee shall conduct a physical inventory every six months to account for all sealed sources and devices containing licensed material received and possessed under the license.
15. The licensee shall not acquire licensed material in a sealed source or device unless the source or device has been registered with the U.S. Nuclear Regulatory Commission pursuant to 10 CFR 32.210 or equivalent regulations of an Agreement State.
16. The licensee is authorized to transport licensed material in accordance with the provisions of 10 CFR Part 71, "Packaging and Transportation of Radioactive Material."

**MATERIALS LICENSE
SUPPLEMENTARY SHEET**

License Number

29-01022-14

Docket or Reference Number

030-29741

Amendment No. 21

17. Except as specifically provided otherwise in this license, the licensee shall conduct its program in accordance with the statements, representations, and procedures contained in the documents, including any enclosures, listed below. The Nuclear Regulatory Commission's regulations shall govern unless the statements, representations, and procedures in the licensee's application and correspondence are more restrictive than the regulations.
- A. Application dated July 20, 1992
 - B. Letter dated June 15, 1993
 - C. Letter dated April 6, 1994
 - D. Letter dated February 28, 1997
 - E. Letter dated July 30, 1997
 - F. Letter dated August 27, 1997, with attachment
 - G. Letter dated September 10, 1997
 - H. Letter dated May 1, 1998
 - I. Letter dated July 2, 1998
 - J. Letter dated May 13, 1998, with attached survey report
 - K. Letter dated July 14, 1999 with attached survey report
 - L. Letter dated September 1, 1999
 - M. Letter dated September 10, 1999
 - N. Letter dated April 19, 2000, with attached survey report
 - O. Letter dated July 6, 2000
 - P. Letter dated August 18, 2000
 - Q. Letter dated September 14, 2000, with enclosure

For the U.S. Nuclear Regulatory Commission

Date March 12, 2001

By

Original signed by Sattar Lodhi, Ph.D.

Sattar Lodhi, Ph.D.
Nuclear Materials Safety Branch 2
Division of Nuclear Materials Safety
Region I
King of Prussia, Pennsylvania 19406