

11/6/70  
FF/12

### Scope of Program

1. Number of individuals occupationally exposed 13.
2. Type of Irradiator (eg. Pool, Pit, etc.) Lic 07 In Air-cask storage  
Pool irradiator.
3. Number of Curies 97-500-1800 of Co-60.
4. Frequency of use: 671-2  
14-1 times per Week.

### Exposure Evaluation

1. Personnel
  - a. Film Badge ✓
  - b. Dosimeter ✓
  - c. Other chip per.
2. Facility
  - a. Independent area radiation monitor Yes
  - b. Survey meter when enter HRA Yes - for in div.

### Surveys

1. Radiation levels in unrestricted areas < 0.1 mR/hr.
2. Contamination smears in restricted area B.E. in pool area.
3. Leak Tests
  - a. Frequency 6 mo.
  - b. Method adequate Yes.
4. Interlocks into HRA
  - a. Frequency of Testing Quarterly - will be each time  
use
  - b. Functional at time of inspection OK at end of inspection.  
Not functional before.

Ex 2

c. Are they intentionally bypassed or deleted. Yes No

*They were deleted*

(1) Procedure if yes

*None*

d. In accordance with license?

*No*

e. Adequate?

*No*

5. Routine maintenance of Hot cell equipment adequate. Yes No

Instrumentation

1. Adequate type and number: Yes No

2. Calibration as required: Yes No

Evaluation of Effluent

1. Liquid *OK*

2. Airborne *N/A*

Training

1. Std. Procedures *Yes*

2. Emergency Procedures *Yes*

3. NRC Regulations *Yes.*

Signs/Posting

1. CRM

2. CHRA

3. 19.11

*Yes.*

Evaluation of Incoming Packages (20.205)

None received

Disposal

None

Evaluation of Outgoing Shipments - (DOT)

N/A

Unusual Occurrences or Events

None

Independent Measurements (Van, Inspector)

Pool sample  
wipes in pool irradiator room.

For irradiations not completely self-shielded containing:

379 Ci cobalt-60

1042 Ci iridium-192

1515 Ci cesium-137 or more, the following must be determined:

I. Control Devices

1. What control device will prevent entry of individuals into the irradiator when the source is exposed?

*Microswitch on door retracts source when door is opened.*

2. What control device will retract the source if an individual attempts entry?

*Same.*

3. What control device prevents operation of the source if an individual is present in the irradiator?

*Alarm rings - switch inside is hit to retract source.*

4. Do any of the above control devices prevent egress from the irradiator?

*No - door can always be opened.*

I. If the Entry Control Devices Fails:

1. What control device will retract the source?

*None -*

*one will be added.*

2. Are visible and audible alarm signals generated to warn individuals entering of the hazard, and to alert another knowledgeable individual? *There is an alarm for 10 seconds before the plug comes up. Flashing light in cell but none at entrance.*

If there is credible probability, the physical radiation barriers can fail: *No - only the door.*

1. What control device will cause the source to retract?

2. Are visible and audible alarms signals generated to warn individuals entering of the hazard, and to alert another knowledgeable individual?

If the Source Is Stored In A Liquid Shield:

1. Is loss of liquid level adequately signaled for immediate action?

*Not on 10/20/78 - Was on 11/1/78*

Exposing the Source

1. What device will automatically generate visible and audible alarm signals to alert individuals before exposing the source?

*See N. 2 above*

2. What clearly identified device can be activated from inside the irradiator which will prevent the source from becoming exposed?

*See M. 3*

3. Is there a procedure to assure that the area is clear of individuals prior to exposing the source?

*Operator must make visual check before exposing the source.*

Physical Radiation Measurements

1. Is a physical radiation measurement made upon entry to the irradiator after source operation?

*Yes - & source to check for functioning*

Tests of Entry Control Devices

*instrument was installed by 11/*

1. Are tests of the entry control devices conducted each day prior to initial operation of the source? (Note: These tests are not required if operations are uninterruptedly continued from the previous day.)

*No but will be*

2. Are records of these tests maintained?

*No but will be*

Control of Portals Into Irradiator

*N/A*

1. What safety devices and administrative procedures are used to prevent entry by individuals through portals that convey materials in and out?

2. Are exit portals equipped to detect and signal presence of loose radiation sources and to automatically prevent them from being carried out?

I. Independent Measurements

1. Take water sample and split with licensee.

- a. Licensee results

*$< 2.5 \times 10^{-6}$   $\mu$  Ci/ml*

- b. IE:I Results

*$< 1 \times 10^{-8}$   $\mu$  Ci/ml*

2. Planchet or bottle source standard.

a. Value

b. Licensee's results

3. Results of interlock checks

Source was not retracted on 10/20

" was retracted on 11/1

4. Is water continuously circulating through demineralizer?

Yes

5. Results of surveys around demineralizer.

$< 0.3$  mR/hr.

6. Demineralized conductivity measurement

Yes.

7. Results of PH check with litmus paper

N/I

8. Restricted area survey results with meter

$< 0.3$  mrem/hr.

9. Restricted area survey results with wipes

$< 100$  dpm



## 10. Unrestricted area survey results:

$< 0.1 \text{ mR/hr}$

## 11. Results of check of liquid level indicator

on 10/20, alarm did not operate  
on 11/1 it did.

INSPECTION FINDINGS AND LICENSEE ACKNOWLEDGMENT

E-III

1. LICENSEE  Dept of the Army U. S. Army Electronics Command Fort Monmouth, N.J. 0703	2. REGIONAL OFFICE  U. S. ATOMIC ENERGY COMMISSION Division of Compliance, Region I 970 Broad Street Newark, N.J. 07102
3. LICENSE NUMBER(S)  29-1022-7	4. DATE OF INSPECTION  January 19, 1970 R/I

5. INSPECTION FINDINGS

- ☒ A. No item of noncompliance was found.
- ☐ B. Rooms or areas were not properly posted to indicate the presence of a RADIATION AREA.  
10 CFR 20.203(b) or 34.42
- ☐ C. Rooms or areas were not properly posted to indicate the presence of a HIGH RADIATION AREA.  
10 CFR 20.203(c) (1) or 34.42
- ☐ D. Rooms or areas were not properly posted to indicate the presence of an AIRBORNE RADIOACTIVITY AREA.  
10 CFR 20.203(d)
- ☐ E. Rooms or areas were not properly posted to indicate the presence of RADIOACTIVE MATERIAL.  
10 CFR 20.203(e)
- ☐ F. Containers were not properly labeled to indicate the presence of RADIOACTIVE MATERIAL.  
10 CFR 20.203(f) (1) or (f) (2)
- ☐ G. A current copy of 10 CFR 20, a copy of the license, or a copy of the operating procedures was not properly posted or made available. 10 CFR 20.206(b)
- ☐ H. Form AEC-3 was not properly posted. 10 CFR 20.206(c)
- ☐ I. Records of the radiation exposure of individuals were not properly maintained. 10 CFR 20.401(a) or 34.33(b)
- ☐ J. Records of surveys or disposals were not properly maintained. 10 CFR 20.401(b) or 34.43(d)
- ☐ K. Records of receipt, transfer, disposal, export or inventory of licensed material were not properly maintained.  
10 CFR 30.51, 40.61 or 70.51
- ☐ L. Records of leak tests were not maintained as prescribed in your license, or 10 CFR 34.25(c)
- ☐ M. Records of inventories were not maintained. 10 CFR 34.26
- ☐ N. Utilization logs were not maintained. 10 CFR 34.27

  
CHARLES E. CONNER  
(AEC Compliance Inspector)

6. LICENSEE'S ACKNOWLEDGMENT


The AEC Compliance Inspector has explained and I understand the items of noncompliance listed above. The items of noncompliance will be corrected within the next 30 days.

(Date)

(Licensee Representative — Title or Position)

ORIGINAL: LICENSEE. COPIES: ☐ CO REGION ☐ CO HEADQUARTERS ☐ CO ENFORCEMENT

INSPECTION FINDINGS AND LICENSEE ACKNOWLEDGMENT

<b>1. LICENSEE</b>  Dept of the Army U. S. Army Electronics Command Fort Monmouth, N.J. 07703	<b>2. REGIONAL OFFICE</b>  U. S. ATOMIC ENERGY COMMISSION Division of Compliance, Region I 970 Broad Street Newark, N.J. 07102
<b>3. LICENSE NUMBER(S)</b>  29-1022-6	<b>4. DATE OF INSPECTION</b>  January 19, 1970
<b>5. INSPECTION FINDINGS</b>  <input checked="" type="checkbox"/> A. No item of noncompliance was found.  <input type="checkbox"/> B. Rooms or areas were not properly posted to indicate the presence of a RADIATION AREA. 10 CFR 20.203(b) or 34.42  <input type="checkbox"/> C. Rooms or areas were not properly posted to indicate the presence of a HIGH RADIATION AREA. 10 CFR 20.203(c) (1) or 34.42  <input type="checkbox"/> D. Rooms or areas were not properly posted to indicate the presence of an AIRBORNE RADIOACTIVITY AREA. 10 CFR 20.203(d)  <input type="checkbox"/> E. Rooms or areas were not properly posted to indicate the presence of RADIOACTIVE MATERIAL. 10 CFR 20.203(e)  <input type="checkbox"/> F. Containers were not properly labeled to indicate the presence of RADIOACTIVE MATERIAL. 10 CFR 20.203(f) (1) or (f) (2)  <input type="checkbox"/> G. A current copy of 10 CFR 20, a copy of the license, or a copy of the operating procedures was not properly posted or made available. 10 CFR 20.206(b)  <input type="checkbox"/> H. Form AEC-3 was not properly posted. 10 CFR 20.206(c)  <input type="checkbox"/> I. Records of the radiation exposure of individuals were not properly maintained. 10 CFR 20.401(a) or 34.33(b)  <input type="checkbox"/> J. Records of surveys or disposals were not properly maintained. 10 CFR 20.401(b) or 34.43(d)  <input type="checkbox"/> K. Records of receipt, transfer, disposal, export or inventory of licensed material were not properly maintained. 10 CFR 30.51, 40.61 or 70.51  <input type="checkbox"/> L. Records of leak tests were not maintained as prescribed in your license, or 10 CFR 34.25(c)  <input type="checkbox"/> M. Records of inventories were not maintained. 10 CFR 34.26  <input type="checkbox"/> N. Utilization logs were not maintained. 10 CFR 34.27  <div style="text-align: right;"> CHARLES E. LOWER (AEC Compliance Inspector)</div>	
<b>6. LICENSEE'S ACKNOWLEDGMENT</b>  The AEC Compliance Inspector has explained and I understand the items of noncompliance listed above. The items of noncompliance will be corrected within the next 30 days.  <div style="display: flex; justify-content: space-between;"><div data-bbox="247 1837 702 1881">_____ (Date)</div><div data-bbox="759 1837 1577 1881">_____ (Licensee Representative — Title or Position)</div></div>	

ORIGINAL: LICENSEE. COPIES: ☐ CO REGION ☐ CO HEADQUARTERS ☐ CO ENFORCEMENT

*BM*  
To: PRN From: CEE

Subject: Dept. of the Army

U.S. Army Electronics Command

Fort Monmouth, N.J. 07703

Lic Nos 29-1022-6 and -7

License No -6 is properly classified as on E(1-A)II, and Lic No -7

as on EIII. I saw no cause for our concern for health and safety in

this well managed program. I recommend reinspection of Lic No -6

at the normal interval (1yr) plus 6mo. (7/71). License No -7 should be

reinspected at the normal interval (2yrs) plus 12mo. (1/73).

REGION I, DIVISION OF COMPLIANCE  
NEWARK, NEW JERSEY

~~SPECIAL LIMITED INSPECTION~~

1. Name and address of licensee:

Dept. of the Army  
U.S. Army Electronics Command  
Fort Monmouth N.J. 07703

2. Date of Inspection: 1/19/70

3. Type of Inspection: Announced reinspect

4. License number(s), docket number(s), number and date of last amendment for each license. Category and Priority of each license:

29-1022-C Amend #12 9/9/69 E(1-A)B

-7 Amend #2 10/28/68 ETE

5. Date of previous inspection: None - C 10/10/68 - T 10/13/66

6. Is "Company Confidential", or proprietary, or classified information contained in report?

Yes \_\_\_\_\_ No ✓

(Specify paragraphs)

7. Scope of inspection: Complete

Charles E. Conn

2/11/70

Paul B. Nelson

4/28/70

licensee: \_\_\_\_\_

Summary

*Moderate sized program - well managed*

Noncompliance and Safety Items

*None*

Unusual Occurrences

*None*

Status of Previously Reported Noncompliance or Safety Items

*Both licenses Clear 5/91*

Management Interview

*Congratulated Mr. Admittis on an excellent program*

licensee: \_\_\_\_\_

## DETAILS

### Participants

C. Pullen, Supervisor Radiation Facility, Sec. of Comm.  
Dr. W. Ramm, RSO  
B. Markow, Chairman Isotope Comm.  
Mr. J. Admitis Asst. to Deputy Electronics Comm.

### Scope of Licensee Program

Very little unsealed material used. Large gamma sealed sources used in hot cell and in collimated calibrators for development and calibration of Dept of Army radior instruments.

### Organization

According to Pullen Lt. Lorenz (who was on leave) does all the log work on health physics. He reports to Dr. Ramm who reports to Markow who reports to Dr. R.S. Wiseman Jr. Deputy for Labs. Lorenz also reports to Savatko the past safety officer. A list of the present radiorisotope committee is on the reverse side of this page. Meetings are held quarterly and the minutes are recorded. I inspected the records.

### Administrative Control

Pullen stated that all procurements are authorized by either him or Markow and all deliveries are made to Pullen's office.

Licensee: \_\_\_\_\_

E. Use of Material

Inventory: (taken from 12/69 monthly report to the Radiological Comm.)

Am 241 - 1.1 mCi

C 14 - 15.66 mCi (storage)

Cs 137 - 102.2 mCi plus one 120 Ci sealed source used in UDM-1A calibrator and 8 each 220 Ci sealed sources stored in UDM-1A calibrators

Co 57 - 52 mCi

H 3 -

Co 60 (all sealed sources) - 3540 Ci (license - 7), 7.12 Ci UDM calibrator, 2 each 1 mCi

Kr 85 - 50 mCi (sealed, never opened or used)

Ni 63 - 2 mCi (gas chrono)

Po 210 (sealed) 3.5 mCi (6 sources)

Pu-239 - 3.62 mCi (held under Gen. Authorization to DOE)

Pm 147 (sealed) 4 sources of 300 mCi each

Receipts for 1968-69: 12/18/68 - 1 mCi Am 241, 7/26/68 - 100 mCi Am 241, 11/24/69 10 mCi Ra 2 (sealed) 12/8/69 - 57.2 Ci H-3 (target, not used) 12/9/69 100 mCi Cs 137 (not used, storage)

~~XXXXXXXXXX~~

Pullen stated that the only material in use was as follows:

Am 241 1 mCi Bldg 45

Am 241 100 mCi Rm 3956

Cs 137 2.2 mCi Rm 4008

Co 57 50 mCi OA 402

Co 57 2 mCi OA 504

C 14 0.16 mCi Rm 4008

Co 60 ( ) Bldg 401

Cs 137 ( ) Bldg 401

Co 60 ( ) Underground vault (Lic-7)

Ni 63 2 mCi Bldg 45

Po 210 1.4 mCi Rm 7 Bldg 11

Po 210 500 mCi Labo house

Pm 147 1.82 mCi Rm 4008

H 3 6.4 Ci Fort Hancock

H 3 9.5 Ci Bldg 401

Ex 2



Licensee: \_\_\_\_\_

G. Equipment

I noted that Pullen had several cabinets full of a variety of radiation detection equipment. Specifically I noted the following items:

400 Rod-Tec Chippers

1500 0-200 mR pocket dosimeters

300 0-5R pocket dosimeters

200 AN/PDR-275 GM survey meters (0-500 mR/hr)

200 AN/PDR-39 ionization chamber rate meters 0-50 R/hr

Victoreen Model 740 Cutie pie

I noted that all survey meters were labeled to show last date of calibration and that the dates were within the past three months.

H. Radiological Safety Procedures

Pullen stated that all users are required to affirm that they have read EUCOM Regulations 385-9 (11/15/61) which are the licensee's safety regulations. Pullen had copies of the licensee and at 10 CFR 20 and 30.

censee: \_\_\_\_\_

Personnel Monitoring and Exposure to External Radiation

Pullen stated that all persons who are occupationally exposed to ionizing radiation are required to wear film badges. I inspected the DD Form 1141 and noted that no person had been exposed to more than 500 mR in any quarter since 1966.

Exposure of Employees to Concentrations of Radioactive Materials

None - sealed sources only

Facilities

I noted that the facilities had not changed from the description given by Higgenball in 1968 and my report of 1966.

~~Records of Leak Test of Sealed Sources~~ Records of Leak Test of Sealed Sources

I inspected the records and found that alpha sources were leak tested at 3 mo. intervals, beta gamma at 6 mo. intervals and no source showed removable contamination in excess of 0.005  $\mu$ Ci.

licensee: \_\_\_\_\_

. Disposals

Pullen stated none since 1966

. Miscellaneous Surveys, Evaluations and Records

I inspected records of surveys made by Lt Lorenz on monthly basis. They covered use and storage areas and showed negligible exposure rates and no detectable contamination

. Special License Conditions

I reviewed each license condition with Pullen and found no items of noncompliance

Licensee: \_\_\_\_\_

1. Posting and Labeling

*I noted that all areas and containers were properly posted and labeled.*

2. Independent Measurements

*None*

3. Operations Observed

*None*

Licensee: \_\_\_\_\_

R. Incidents, Overexposures, Theft or Loss, Equipment Malfunction

Pollen stated none

S. Other information or continuation from previous paragraphs.

None