


**Customer:** 92093L  
**Location ID** 00000DPT  
**Report Number:** 1740


Sealed Radioactive Source Leak Test Certificate  
 Save this report for your records

Benefits Healthcare  
 Kari L. Cann  
 Radiation Oncology  
 Slatten Cancer Institute  
 1117 29th Street South  
 Great Falls, MT 59405

Isotope	Activity/ Date	Source Manufacturer:	Collected By:	Analyzed By:	Result (uCi)
Sr-90	100 mCi 4/13/1977	3M Model: 6DIA Serial: 6115	Kari L. Cann, MS, DABR Collected On: 7/11/2006 Using GDS Q1 ki	GDS Counting Lab Analyzed On: 07/21/2006	0.0000

The Leak Tests were analyzed in accordance with Global Dosimetry Solutions, Inc. procedures and Radioactive Material License #4635-30. Calibration and Quality Control records are maintained for the Leak Test Counting Equipment at Global Dosimetry Solutions, Inc.

  
 \_\_\_\_\_  
 Radiation Safety Officer Review

  
 \_\_\_\_\_  
 Date

