



SOUTHWEST RESEARCH INSTITUTE®

6220 Culebra Road, P.O. Drawer 28510
Institute Quality Systems
Institute Calibration Laboratory
Phone: 210-522-5215 Fax 210-522-4834



Certificate of Calibration

0972-01

Submitted By: DIV20

Address: B57

Contact: DON BANNON

Manufacturer Model: FISHER SCIENTIFIC 15-166A

Description: THERMOMETER, GLASS

Serial No: C96-816

Asset No: 005433

Procedure: THERMOMETERS, GLASS - 11 SEP 2006

Work Order: 303071829

Date Issued: Nov 28, 2006

Calibration Date: Nov 28, 2006

*Calibration Due: May 28, 2007

Calibration Location: Bldg. 64

Environment: Temp. 73.0°F Hum. 40 %RH

**Data Type: FOUND-LEFT

DivID/Location:

This certificate documents traceability to the National Institute of Standards and Technology (NIST) and the International System of Units (SI). The Laboratory quality system conforms to ISO/IEC 17025, 1999, ANSI/NC SL Z540-1-1994 and relevant requirements of the ISO 9000-2000 standard. This certificate shall not be reproduced, except in full, without the written approval of the Southwest Research Institute Calibration Laboratory. This certificate shall not be used to claim product endorsement by Southwest Research Institute, American Association for Laboratory Accreditation (A2LA) or any agency of the U. S. Government. Results of this calibration relate only to the instrument described above at the time of calibration and does not imply any long term stability of the instrument.

*Determined by the customer, does not imply the instrument will remain within tolerance as any number of factors may cause an out-of-tolerance condition before this date. **Found/Left = adjustment and/or repair was not required, As Left = adjusted and/or repaired was required. The client has sole responsibility for determination of in-/out-of-tolerance or compliance/noncompliance. See Remarks or attached Measurement Report with the same Work Order number for data.

Reported uncertainty calculated in accordance with the ISO "Guide to the Expression of Uncertainty in Measurement" (GUM) and represents an expanded uncertainty with a coverage factor of k=2 to approximate a 95% confidence level.

Remarks: None

Standards Used

Asset No.	Serial No.	Manufacturer	Model	Description	Cal Due
009917	604420	HART SCIENTIFIC	5612	RTD	Apr 06, 07
012305	A51826	HART SCIENTIFIC	1502A	TEMPERATURE READOUT	Apr 06, 07

Reviewed by: bit () jrg () pwc () wgh ()

Metrology Technician

m:\a2la1.rpt Rev date August 15, 2005

Measurements by: Bob Trollinger

Metrology Technician

Southwest Research Institute
 Calibration Laboratory
 Measurement Report

Work Order:	303071829	Mfr:	Fisher Scientific	Technician:	blt
Asset No:	005433	Model:	15-166A	Cal Date:	28-Nov-06
Serial No:	C96-816	Type:	THERMOMETER		
Remarks:	Accuracy Reference ANSI Z236.1-1983				

Function/Range	Test Point	TI Reading	Difference	+/-Limit	+/-Uncertainty	Found/Left
Temperature	°C	°C	°C	°C	°C	Result
	-19.83	-20.0	-0.17	0.5	0.29	Pass
	0.09	0.0	-0.1	0.5	0.29	Pass
	50.03	50.0	0.0	0.5	0.29	Pass
	99.96	100.0	0.0	0.5	0.29	Pass
	150.07	150.0	-0.1	0.5	0.29	Pass

END OF REPORT