

**SOUTHWEST RESEARCH INSTITUTE  
CALIBRATION LABORATORY  
MEMORANDUM**

**November 28, 2006**

**To:** DON BANNON DIV20 B57

**From:** Institute Calibration Laboratory

**Subject:** Status of Calibration Supplier

**Manufacturer/Model:** OHIO SEMITRONICS PC5-103D

**Description:** AC WATT TRANSDUCER

**Serial Number:** 03080279

**Asset Number:** 010467

**Work Order Number:** 303071532

**Date Calibrated:** November 2, 2006

**Supplier:** CMI, CENTERVILLE GA - A2LA - 800 525-0408

**Remarks:**

☒ Supplier is on the Approved Suppliers List (ASL).

☐ Supplier is not on the Approved Suppliers List.

☒ Calibration is ISO 17025 accredited.

☐ Calibration is not ISO 17025 accredited.

☐ There is no known supplier to meet ISO 17025 accreditation at this time.

Please contact the Institute Calibration Laboratory, extension 5215, if you have any questions about the condition of this equipment or calibration documentation.

Attachment(s) 2

**Customer Information**

Southwest Research Institute  
6220 Culebra Rd  
San Antonio, TX 78238

PO #: 787060J  
Reference #: 07002812  
Account #: SO1043  
SO #: 60463

**Instrument Identification**

Instrument Id: **03080279**

Noun: AC Watt Transducer

Serial #: 03080279

Mfr: Ohio Semitronics

Model: PC5-103D

Accuracy:  $\pm 0.5\%$  of Full Scale

Expanded Measurement Uncertainty at K=2:  $\pm 0.1\%$  of Full Scale

**Certification Information**

Reason For Service: Calibration

Type Of Calibration: Normal

As Found Condition: In Tolerance

As Left Condition: Left As Found

Procedure: Mfr Manual :

Technician: Scott McLeod

Cal Date: 02 NOV 06

Cal Due: 02 NOV 07

Temperature: 24.0 °C

Humidity: 42.0 %

Quality Remarks: Calibration complies with ISO/IEC 17025 & ANSI/NCSL Z540 requirements.


*This instrument has been calibrated using standards with accuracies traceable to the National Institute of Standards and Technology, derived from natural physical constants, derived from ratio measurements, or compared to consensus standards.*

*Reported uncertainties and/or "test uncertainty ratios" (TUR's) are expressed as expanded uncertainty values at approximately the 95% confidence level using a coverage factor of K=2. A TUR of 4:1 is routinely observed unless otherwise noted on the Certificate. Statements of compliance, where applicable, are based on test results falling within specified limits with no reduction by the uncertainty of the measurement.*


*Certified Measurements, Inc.'s Quality System complies with the applicable requirements of ANSI/NCSL Z540-1 and ISO/IEC 17025.*

*The results contained herein relate only to the item calibrated. Calibration due dates appearing on the Certificate of Calibration and label are determined by the client for administrative purposes and do not imply continued conformance to specification.*

*This certificate shall not be reproduced except in full, without the written permission of Certified Measurements, Inc.*



Deborah J. Nazimiec, Technical Manager



Robert S. Lamb, President

✓ In Tolerance    ✗ Out of Tolerance

## Calibration Data

Range	Nominal	As Found		As Left		Min	Max
Watt transducer, output in volts dc							
20 W	2.00	2.013	✓	Left as Found	■	1.95	2.05
40 W	4.00	4.028	✓	Left as Found	■	3.95	4.05
60 W	6.00	6.032	✓	Left as Found	■	5.95	6.05
80 W	8.00	8.032	✓	Left as Found	■	7.95	8.05
100 W	10.00	10.025	✓	Left as Found	■	9.95	10.05

End of Datasheet

## Calibration Standards

<u>NIST Traceable #</u>	<u>Instrument ID#</u>	<u>Description</u>	<u>Model</u>	<u>Calibration Date</u>	<u>Date Due</u>
1000206177	22956	Calibrator	811A	15 JUN 2006	30 JUN 2008
1000212690	27504	Multimeter	1271	10 OCT 2006	30 APR 2007

Your calibration records are available on-line at <http://www.cmi-labs.com>

Printed: 15 Nov 2006 03:11 PM

**CERTIFIED MEASUREMENTS, INC.**  
*Raising the Calibration Standard...*

510 North Houston Lake Blvd.  
Centerville, Georgia 31028  
Tel. 478.953.5171  
Fax: 478.953.2688