

**U.S. NUCLEAR REGULATORY COMMISSION**

Date: 12/29/06

**TELEPHONE CONVERSATION RECORD**

Time: 9:50 AM

Mail Control 139572 License No(s) STC-133 Docket No(s) 04000341  
 or Report No(s).

Name of Licensee: Defense Logistics Agency  
 Name of Participant(s): Claude Wiblin, CHP, Contractor  
 Chesapeake Nuclear Services, Inc.  
 Telephone No. 410-923-4653  
 Subject: RAI, Reference Readings

(NOTE: This will be used as the Documents Title in ADAMS)

Summary: The reference areas and material was not included on 13CFL03, 13CWL01, and 13CWL02. The material background was 4.6 which is higher than any background location for instrument serial number 235489. Also Duplicate measurement in reference 10AFL01 and 14DFL01 using detector SHP 380 was not documented.

Action Required: Will review the information and send correction or additional data.

Document Availability:  Publicly Available  Non-Publicly Available  
 Non-Sensitive  Non-Sensitive Copyright  Sensitive  Sensitive Copyright  
 Immediate Release  Normal Release  Delay Release Date

Prepared & SUNSI Review Completed By: Dennis Lawyer / RA / Date: 12/29/2006