

NRC FORM 314 (6-2004) 10 CFR 30.36(j)(1); 40.42(j)(1); 70.38(j)(1); and 72.54(j)(1)	U.S. NUCLEAR REGULATORY COMMISSION APPROVED BY OMB: NO. 3150-0028 EXPIRES: 06/30/2007 Estimated burden per response to comply with this mandatory collection request: 30 minutes. This submittal is used by NRC as part of the basis for its determination that the facility is released for unrestricted use. Send comments regarding burden estimate to the Records and FOIA/Privacy Services Branch (T-5 F52), U.S. Nuclear Regulatory Commission, Washington, DC 20555-0001, or by internet e-mail to infocollects@nrc.gov, and to the Desk Officer, Office of Information and Regulatory Affairs, NEOB-10202, (3150-0028), Office of Management and Budget, Washington, DC 20503. If a means used to impose an information collection does not display a currently valid OMB control number, the NRC may not conduct or sponsor, and a person is not required to respond to, the information collection.
---	---

CERTIFICATE OF DISPOSITION OF MATERIALS

LICENSEE NAME AND ADDRESS American Video Glass Company, L.L.C. 777 Technology Dr Mt. Pleasant, PA 15266	LICENSE NUMBER 37-30348-01	DOCKET NUMBER 030-34274
LICENSE EXPIRATION DATE March 31, 2007		

A. LICENSE STATUS (Check the appropriate box)

This license has expired. This license has not yet expired; please terminate it.

B. DISPOSAL OF RADIOACTIVE MATERIAL
 (Check the appropriate boxes and complete as necessary. If additional space is needed, provide attachments)

The licensee, or any individual executing this certificate on behalf of the licensee, certifies that:

1. No radioactive materials have ever been procured or possessed by the licensee under this license.

2. All activities authorized by this license have ceased, and all radioactive materials procured and/or possessed by the licensee under this license number cited above have been disposed of in the following manner:

a. Transfer of radioactive materials to the licensee listed below:

Qal-Tek Associates
 3998 Lemmerce Cr.
 Idaho Falls, Idaho 83401
 Phone (208) 523-5557

Transfer Date: 10/16/2006
 Contact: Jared Eggleston
 NRC/Agreement State License 11-27610-01

b. Disposal of radioactive materials:

1. Directly by the licensee:

2. By licensed disposal site:

3. By waste contractor:

c. All radioactive materials have been removed such that any remaining residual radioactivity is within the limits of 10 CFR Part 20, Subpart E, and is ALARA.

C. SURVEYS PERFORMED AND REPORTED

1. A radiation survey was conducted by the licensee. The survey confirms:

a. the absence of licensed radioactive materials

b. that any remaining residual radioactivity is within the limits of 10 CFR 20, Subpart E, and is ALARA.

2. A copy of the radiation survey results:

a. is attached; or b. is not attached (Provide explanation); or c. was forwarded to NRC on: _____ Date _____

3. A radiation survey is not required as only sealed sources were ever possessed under this license, and

a. The results of the latest leak test are attached; and/or b. No leaking sources have ever been identified.

RECEIVED REGION 1
 2006 DEC 26 AM 8:57

The person to be contacted regarding the information provided on this form:

NAME Doug Orbin	TITLE Facilities Engineering Manager	TELEPHONE (Include Area Code) 724-696-7171	E-MAIL ADDRESS Doug.Orbin@am.sony.com
--------------------	---	---	--

Mail all future correspondence regarding this license to:
 Doug Orbin, 1001 Technology Dr, Mt. Pleasant, PA 15266

C. CERTIFYING OFFICIAL

I CERTIFY UNDER PENALTY OF PERJURY THAT THE FOREGOING IS TRUE AND CORRECT

PRINTED NAME AND TITLE William Hanna, V.P. STP	SIGNATURE <i>William Hanna</i>	DATE 12/19/06
---	-----------------------------------	------------------

WARNING: FALSE STATEMENTS IN THIS CERTIFICATE MAY BE SUBJECT TO CIVIL AND/OR CRIMINAL PENALTIES. NRC REGULATIONS REQUIRE THAT SUBMISSIONS TO THE NRC BE COMPLETE AND ACCURATE IN ALL MATERIAL RESPECT. 18 U.S.C. SECTION 1001 MAKES IT A CRIMINAL OFFENSE TO MAKE A WILLFULLY FALSE STATEMENT OR REPRESENTATION TO ANY DEPARTMENT OR AGENCY OF THE UNITED STATES AS TO ANY MATTER WITHIN ITS JURISDICTION.



**Qal-Tek
Associates**

www.qaltek.com

Instrumentation and Professional Services

Rich Walker
American Video Glass Company
777 Technology Drive
Mt. Pleasant, PA 15666

Date: 10/16/2006

This is to advise that the Radioactive Material as detailed below has been received by Qal-Tek Associates as of 10/16/2006 and we have taken custody of them:

<u>Manufacturer</u>	<u>Model</u>	<u>Serial Number</u>	<u>Isotope</u>	<u>Activity (mCi)</u>
Nordion Inc.	C-163	AO21	Co-60	2400
Nordion Inc.	C-163	AO22	Co-60	2400

Please retain this record for your files. Should you require further assistance, please contact us at Qal-Tek Associates.

Regards,

Jared Eggleston
Radioactive Source Disposal Specialist

Applied Health Physics, Inc.

2986 Industrial Blvd., Bethel Park, Pa 15102 Phone: (412) 835-9555 Fax: (412) 835-9559
E-Mail address: ahp.inc@adelphia.net - Web Page address: WWW.APPLIEDHEALTHPHYSICS.COM

Mark V Leak Test Certificate

This certificate shall not be reproduced except in full, without the written approval of Applied Health Physics, Inc.

Device Information

Device Serial Number:	321	Source Serial Number:	AO-21
Device Manufacturer:	Level Link	Source Manufacturer:	Nordian Inc.
Device Model:	770-A5	Source Model:	C-163
Device Type:	Gauge	Make:	Location: AVG-1
Radioisotope:	Co-60	Activity:	2400 mCi

Leak Test Information

Leak Test Performed by:	Anthony Hull <i>Anthony J. Hull</i>	Leak Test Performed on:	5/3/2006
Leak Test Specimen Number:	AVG-050306-1	Leak Test Analysis Date:	5/3/2006
Leak Test Analysis Performed by:	A. Hull		
Leak Test Results Indicate an Activity of:	< 0.005 uCi		

Notification

The analysis of this leak test sample was provided by Applied Health Physics, Inc. U.S. Nuclear Regulatory Commission License Number 37-09135 and State License Number PA-0228. In the event this specimen indicates leakage and or contamination in excess of 0.005 microcuries, Applied Health Physics, Inc. will notify you promptly by telephone; at such time you will need to take the appropriate actions required by the governing regulatory agency. Applied Health Physics, Inc. is required by our license to report leakage and/or contamination in excess of 0.005 microcuries.

Client Information

American Video Glass Company	Contact:	Ron Gallagher
777 Technology Drive	Phone:	(724) 696-6129
	Contact:	Fax Number
email: rgallagher@amervideo.com	Phone:	(724) 696-6332
Mount Pleasant	PA	15666

Individual notified _____ Time _____ Date _____ - Client

Individual notified _____ Time _____ Date _____ - Agency

This certificate is an essential record and should be maintained for inspection by the regulatory agency.

Next LT Due 11/3/2006

Applied Health Physics, Inc.

2986 Industrial Blvd., Bethel Park, Pa 15102 Phone: (412) 835-9555 Fax: (412) 835-9559
E-Mail address: ahp.inc@adelphia.net - Web Page address: WWW.APPLIEDHEALTHPHYSICS.COM

Mark V Leak Test Certificate

This certificate shall not be reproduced except in full, without the written approval of Applied Health Physics, Inc.

Device Information

Device Serial Number:	322	Source Serial Number:	A022		
Device Manufacturer:	Level Link Incorporated	Source Manufacturer:	Nordion Incorporated		
Device Model:	770-A5	Source Model:	C-163		
Device Type:	Gauge	Make:		Location:	AVG-2
Radioisotope:	Co-60	Activity:	2400 mCi		

Leak Test Information

Leak Test Performed by:	Anthony Hull <i>Anthony J. Hull</i>	Leak Test Performed on:	5/3/2006
Leak Test Specimen Number:	AVG-050306-2	Leak Test Analysis Date:	5/3/2006
Leak Test Analysis Performed by:	A. Hull		
Leak Test Results Indicate an Activity of:	< 0.005 uCi		

Notification

The analysis of this leak test sample was provided by Applied Health Physics, Inc. U.S. Nuclear Regulatory Commission License Number 37-09135 and State License Number PA-0228. In the event this specimen indicates leakage and or contamination in excess of 0.005 microcuries, Applied Health Physics, Inc. will notify you promptly by telephone; at such time you will need to take the appropriate actions required by the governing regulatory agency. Applied Health Physics, Inc. is required by our license to report leakage and/or contamination in excess of 0.005 microcuries.

Client Information

American Video Glass Company	Contact:	Ron Gallagher
777 Technology Drive	Phone:	(724) 696-6129
	Contact:	Fax Number
email: rgallagher@amervideo.com	Phone:	(724) 696-6332
Mount Pleasant	PA	15666

Individual notified _____ Time _____ Date _____ - Client

Individual notified _____ Time _____ Date _____ - Agency

This certificate is an essential record and should be maintained for inspection by the regulatory agency.

Next LT Due 11/3/2006

139882

Express Envelopes

UPS Next Day Air[®]
UPS Worldwide ExpressSM
Shipping Document

SHIPMENT FROM

UPS ACCOUNT NO.

REFERENCE NUMBER

Aug Orbin TELEPHONE *721-496-1111*
my Electronics Inc.
Technology Dr.
CA 95666

DELIVERY TO

TELEPHONE

Engineering Asst. Section TEL. *337-5000*
Heavy Materials Safety Branch USN
75 Allendale Rd.
Prussia PA 19400-1411

WEIGHT	WEIGHT	DIMENSIONAL WEIGHT

IME Standard
 The shipper authorizes UPS to act as forwarding agent for export control and customs purposes.
 The shipper certifies that these commodities, technology or software were exported from the United States in accordance with the Export Administration Regulations. Diversion contrary to U.S. law is prohibited.

- EXPRESS (INT'L)
- DOCUMENTS ONLY

SATURDAY DELIVERY

J150 929 993 1



J150 929 993 1

EXPORT

UPS Next Day Air[®]
 EXTREMELY URGENT

J150 929 993 1



J150 929 993 1

DELIVERY

DATE OF SHIPMENT

475 ALLENDALE RD

KING OF PRUSSIA PA 19406-1415

P: RED S: 09 I: 4F

0211-RDC

J1509299931

1200

VTTZCL Power224 Dec 22 00:33:52 2006
TB 1940 HIP 5.10.3 LP2044

This is to acknowledge the receipt of your letter/application dated

12/19/2006, and to inform you that the initial processing which includes an administrative review has been performed.

- TEAM. 37-30348-d There were no administrative omissions. Your application was assigned to a technical reviewer. Please note that the technical review may identify additional omissions or require additional information.
- Please provide to this office within 30 days of your receipt of this card

A copy of your action has been forwarded to our License Fee & Accounts Receivable Branch, who will contact you separately if there is a fee issue involved.

Your action has been assigned **Mail Control Number** 139882.
When calling to inquire about this action, please refer to this control number.
You may call us on (610) 337-5398, or 337-5260.