



UNITED STATES
NUCLEAR REGULATORY COMMISSION
REGION I
475 ALLENDALE ROAD
KING OF PRUSSIA, PENNSYLVANIA 19406-1415

January 3, 2007

Docket No. 03009149
Control No. 139757

License No. 29-15459-01

David P. Gibbons
Assistant Vice President Operation
Kennedy Memorial Hospitals
Hurffville-Cross Keys Road
Turnersville,, NJ 08012

SUBJECT: KENNEDY MEMORIAL HOSPITALS, REQUEST FOR ADDITIONAL
INFORMATION CONCERNING APPLICATION FOR AMENDMENT TO
LICENSE, CONTROL NO. 139757

Dear Mr. Gibbons:

This is in reference to your letter dated November 14, 2006 requesting to amend Nuclear Regulatory Commission License No. 29-15459-01. In order to continue our review, we need the following additional information:

1. In support of your request to authorize Dr. Carolyn Horowitz for HDR use, please submit a copy of her vendor training and a preceptor statement for HDR use.
2. Please describe the principal use of the area above the HDR treatment room. Additionally, describe any voids in the shielding and indicate how these voids are shielded.
3. Please confirm that extremity monitors will be provided to individuals who may be called upon to respond to an emergency involving an unretracted or stuck HDR source.
4. Please confirm that your radiation safety training will be provided in accordance with 10 CFR 19.12 to all personnel involved with use of the HDR.

Current NRC regulations and guidance are included on the NRC's website at www.nrc.gov; select **Nuclear Materials; Medical, Academic, and Industrial Uses of Nuclear Material**; then **Toolkit Index Page**. Or you may obtain these documents by contacting the Government Printing Office (GPO) toll-free at 1-888-293-6498. The GPO is open from 7:00 a.m. to 8:00 p.m. EST, Monday through Friday (except Federal holidays).

We will continue our review upon receipt of this information. Please reply to my attention at the Region I Office and refer to Mail Control No. 139757. If you have any technical questions regarding this deficiency letter, please call Shirley Xu at (610) 337-5006.

D. Gibbons
Kennedy Memorial Hospitals

2

If we do not receive a reply from you within 30 calendar days from the date of this letter, we will assume that you do not wish to pursue your application.

Sincerely,

Original signed by Randolph C. Ragland, Jr.

Randolph C. Ragland, Jr.
Senior Health Physicist
Medical Branch
Division of Nuclear Materials Safety

cc:
Lester M. Tripp, Radiation Safety Officer

DOCUMENT NAME: C:\FileNet\ML070040381.wpd

SUNSI Review Complete: S Xu

After declaring this document "An Official Agency Record" it will be released to the Public.

To receive a copy of this document, indicate in the box: "C" = Copy w/o attach/encl "E" = Copy w/ attach/encl "N" = No copy

OFFICE	DNMS/RI	N	DNMS/RI	N	DNMS/RI			
NAME	S Xu / SSX /		RRagland / RCR1 /					
DATE	1/3/07		1/3/07					

OFFICIAL RECORD COPY