

From: Nima Ashkeboussi
To: John Miller
Date: 1/3/2007 10:18:19 AM
Subject: Questions for amendment to NR-1235-S-102-S

Hi John,

In order for us to complete your amendment request for NR-1235-S-102-S, we will need the following additional information:

2. On the last page of Enclosure 1 to your letter, you provided the calibration dates for the instruments used to measure the radiation profile for Model BM06S-57. The date that the exposure rates were recorded are past the calibration due date of the instruments. Please provide justification on how the rates recorded are accurate.

Please contact me if you have any additional questions.

Sincerely,

Nima Ashkeboussi, Mechanical Engineer
Source Safety and Security Branch
US Nuclear Regulatory Commission
301-415-7637
naa@nrc.gov

Mail Envelope Properties (459BC93B.7BA : 10 : 35736)

Subject: Questions for amendment to NR-1235-S-102-S
Creation Date 1/3/2007 10:18:19 AM
From: Nima Ashkeboussi

Created By: NAA@nrc.gov

Recipients
intisoid.com
jjmiller (John Miller)

Post Office

Route
intisoid.com

Files	Size	Date & Time
MESSAGE	1314	1/3/2007 10:18:19 AM

Options
Expiration Date: None
Priority: Standard
ReplyRequested: No
Return Notification: None

Concealed Subject: No
Security: Standard