

07-07

Application and Review Checklist for (Acceptance, 1st, or 2nd) Review for SSD 00-000

SUMMARY DATA	
Name and Complete Mailing Address of the Applicant: <i>International Isotopes</i>	Name, Title, and Telephone Number of the Individual to Be Contacted If Additional Information or Clarification Is Needed by the NRC: <i>John Miller</i>
The Applicant is (check one):	If the Applicant Is Not the Manufacturer, Provide the Name and Complete Mailing Address of the Manufacturer:
<input type="checkbox"/> Custom User	
<input type="checkbox"/> Manufacturer	
<input type="checkbox"/> Distributor	
<input checked="" type="checkbox"/> Manufacturer and Distributor	
If the Applicant is a Custom User, Provide the Name and Complete Mailing Address of the Distributor:	Provide the Name, Complete Mailing Address, and Function of Other Companies Involved:
Model Number: <i>NR-1235-S-1023 RMOGS Series</i>	Principal Use Code (see Appendix E): <input checked="" type="checkbox"/>
Name Used by the Industry to Identify the Product (e.g., Radiography Exposure Device, Teletherapy Source, Calibration Source, etc.):	For Use by:
	<input checked="" type="checkbox"/> Specific Licensees Only
	<input type="checkbox"/> General Licensees Only
	<input type="checkbox"/> Both Specific and General Licensees
	<input type="checkbox"/> Persons Exempt from Licensing
Leak-Test Frequency:	Principal Section of the 10 CFR that Applies to the User (e.g., General Licensees under 10 CFR 31.5):
<input type="checkbox"/> Periodic Leak-Testing is Not Required	
<input checked="" type="checkbox"/> 6 Months	Radionuclides and Maximum Activities (including loading tolerance):
<input type="checkbox"/> Attached is justification for a leak test frequency of greater than 6 months	<i>Various</i>
CERTIFICATION:	
THE APPLICANT UNDERSTANDS THAT ALL STATEMENTS AND REPRESENTATIONS MADE IN THIS APPLICATION ARE BINDING UPON THE APPLICANT.	
THE APPLICANT AND ANY OFFICIAL EXECUTING THIS CERTIFICATION ON BEHALF OF THE APPLICANT, NAMED IN ITEM 2, CERTIFY THAT THIS APPLICATION IS PREPARED IN CONFORMITY WITH TITLE 10, CODE OF FEDERAL REGULATIONS, PARTS 30 AND 32 AND THAT ALL INFORMATION CONTAINED HEREIN IS TRUE AND CORRECT TO THE BEST OF THEIR KNOWLEDGE AND BELIEF.	
WARNING: 18 U.S.C. SECTION 1001 ACT OF JUNE 25, 1948 62 STAT. 749 MAKES IT A CRIMINAL OFFENSE TO MAKE A WILLFULLY FALSE STATEMENT OR REPRESENTATION TO ANY DEPARTMENT OR AGENCY OF THE UNITED STATES AS TO ANY MATTER WITHIN ITS JURISDICTION.	
Certifying Officer — Typed Name and Title	
Signature: <i>LRAI</i>	Date: <i>1/3/07</i>

CHECKLIST

Registration Certificate Holder:

Model:

DESCRIPTION	OK/DEF		COMMENTS
	1 st Reviewer	2 nd Reviewer	
DESCRIPTION/CONSTRUCTION			
If registration certificate holder is requesting to register more than one source/device on a certificate, are designs similar enough to do so?	ok.	✓	Amendment to add sources for NR-1235-S-102-S
Device/source design with complete engineering drawings (dimensions, tolerances, list of materials)	ok.	✓	
Assembly methods (screw, welds, etc.); verify integrity	ok.	✓	
Source mounting (size and integrity) and security	Ok.	✓	
Is source ANSI classification sufficient (from ANSI N43.6 and ISO 2919): Radiography - Unprotected 43515 Radiography - In Device 43313 Medical - Radiography 32312 Medical - γ Teletherapy 53524 Medical - Brachytherapy 53211 Medical - Source Applicators 43312 γ Gauges - Unprotected 43333 γ Gauges - In Device 43232 β Gauges, Low Energy γ Gauges, or X-ray fluorescence 33222 Oil Well Logging 56522 Portable Moist/Density 43333 Neutron Applications 43323 Calibration source activity > 30μCi (1 MBq) 22212 γ Irradiators (I) 43323 γ Irradiators (II, III) 43424 γ Irradiators (II, III, IV) 53424 Chromatography 32211 Static Eliminators 22222 Smoke Detectors 32222	ok	✓	22312 i Calibration Calibration Source Source
Definition of shutter operation (locked in Off position, not locked in On position), Fail safe, spacing and tolerances	N/A	—	
On-Off indicators (description, qty., location)	N/A	—	
Safety interlocks, guards, etc. to prevent access to beam or high radiation levels	N/A	—	
Corrosion between unlike materials (e.g., aluminum & steel, depleted uranium & steel, etc.)	ok.	✓	
Shielding efficiency and integrity	N/A	—	
For medical devices: Was a 510(k) provided? (provide written notification to FDA)	N/A	—	
Well logging sources must be nondispersible and nonsoluble. (see Appendix B for a list of approved well logging sources as of November 1991)	N/A	—	
See "ANSI and Other Standards" list for references for particular source/device designs (e.g. radiography, Brachytherapy, etc.)	N/A	—	

CHECKLIST

Registration Certificate Holder:

Model:

DESCRIPTION	OK/DEF		COMMENTS
	1 st Reviewer	2 nd Reviewer	
LABELING			
Copy of label	ok.	✓	
Materials, dimensions, colors (note on registration certificate if labeling is exempt from the color requirements of 10 CFR Part 20)	ok.	✓	
Permanent attachment and location(s) - visible to users?	ok.	✓	
Contents: Model#, Serial#, Isotope, Activity, Manufacturer, Date of Assay, Trefoil, "CAUTION - RADIOACTIVE MATERIAL" (Depleted Uranium information must be included)	ok.	✓	
CONDITIONS OF USE			
Expected working life of the source/device (years, operations)	Ok.	✓	Same as current certificate
Actions to be taken when product reaches end of its working life.	ok.	✓	
Maximum allowable temperature, vibration, shock, corrosion, etc. (during use, handling, storage, and transport)	ok.	✓	
How the device will be used	ok.	✓	
Meets dose limits of Part 32 for distribution general licensees or persons exempt from licensing	N/A	—	
PROTOTYPE TESTING/HISTORICAL USE			
Tests methods and conditions (for source and device)	ok	✓	
Tests results	ok	✓	
Years of use (incidents, failures, etc.)	ok	✓	
Similarities to other sources/devices if they are used as basis.	ok	✓	
RADIATION PROFILES			
Survey instrument used (type, window thickness, sensitivity, etc.)	Def. Ok.	def	Calibration data Profile calculated with software
Conditions: including environments, scatter (product in beam), and use of guards and shields	N/A.	—	Def: Need method of levels for BM065-57
Distance from source/surface (per ANSI 538-1979)	ok.	✓	
Shutter Open and Closed/Source Shielded	N/A	—	
Verify radiation surveys for γ radiation meet inv^2 law.	—	—	
Verify radiation surveys for non- γ radiation have not been calculated using inv^2 law.	—	—	

CHECKLIST

Registration Certificate Holder:

Model:

DESCRIPTION	OK/DEF		COMMENTS
	1 st Reviewer	2 nd Reviewer	
QUALITY ASSURANCE	Ok		
Materials, subassemblies, services	ok	✓	
Assembly methods (screws, welding, etc.)	ok	✓	
Dimensions and tolerances	ok	✓	
Activity, radiation levels, leak tests	ok	✓	
QA Manual and comparison of manual to Regulatory Guide 6.9	-		
INSTALLATION	N/A		
Fixed, portable, movable, fixed installation but portable source housing			
Inherent shielding, inaccessibility			
Beam access: size of air gap/opening to beam and use of interlocks, locks, additional shielding or barriers			
Mounting integrity			
SAFETY INSTRUCTIONS			
Operation, maintenance, calibration, damage/failure, specific warnings, leak test, and radiation surveys	-		
ACCOMPANYING DOCUMENTATION	N/A		
Leak tests results and radiation surveys			
Transportation documents			
Operation, maintenance, calibration, damage/failure, specific warnings, leak test, and radiation survey instructions if applicable			
For Distribution to General Licensees: Verify NRC Regions and Agreement State listing is up-to-date and copies of all pertinent regulations		✓	

CHECKLIST

Registration Certificate Holder:

Model:

DESCRIPTION				OK/DEF		COMMENTS
				1 st Reviewer	2 nd Reviewer	
SERVICING				N/A		
The following activities may be performed by the persons indicated:				↓		
Activity	by a General Licensee	Only by a Specific Licensee	Will be Offered by the Applicant			
Installation						
Relocation						
Maintenance						
Repair						
Source Exchange						
Calibration						
Leak Testing						
Radiation Survey						
Training						
FOREIGN VENDORS				N/A		
Drop ship					—	
Who and where is source installed					—	
Leak test and radiation surveys					—	
QA in the U.S.					—	

1st Reviewer Signature:

Nirima Ashkelouss

Date: 1/3/07

2nd Reviewer Signature:

John P. [Signature]

Date: 1/3/07