

Licensing Assistant Section
Nuclear Materials Safety Branch
U.S. Nuclear Regulatory Commission, Region 1
475 Allendale Road
King of Prussia, PA 19406-1415

Attn: Mr. Ragland

RE: NRC LICENSE AMENDMENT REQUEST of January 10, 2007

AROS Transportation, Inc.
1 Lethbridge Plaza, Suite 20
Route 17 North
Mahwah, New Jersey 07430
RE: NRC License # 29-30593-01

02 03037088

RADIATION SAFETY SURVEY CLOSEOUT

Survey Instrument: Ludlum Model 14C, SN# 197847
Bkg: 0.03 mR/hr Check Source: 1.5 mR/hr
Cal. Expires 2-6-07

AROS TRANSPORTATION CHEVY VAN

(Previously used to transport Varian 200T transportable HDR unit
Sealed Source)

Van is located in Parking Lot of Nyack Hospital, Nyack, NY.

Survey Performed on 1-15-07 by Mark M. Belanich, M.S., D.A.B.R.

Background away from vehicle:	0.03 mR/hr
Van Exterior:	0.03 mR/hr
HDR Lockdown Area:	0.03 mR/hr
Van Floor:	0.03 mR/hr
Lift Gate:	0.03 mR/hr
Cabinet areas:	0.03 mR/hr
Front cab area:	0.03 mR/hr

139953

NMSS/RGN MATERIALS-002-1

REC'D IN LAT JAN 19 2007

All areas in and around the area were found to be at the background levels of radioactivity.

Please contact Mark M. Belanich with any questions at (845) 729-0075.

A handwritten signature in cursive script that reads "Mark M. Belanich". The signature is written in black ink and is positioned above a horizontal line.

Mark M. Belanich, M.S., D.A.B.R.
V.P. of Medical Physics Services, mpowered, inc.