

CONVERSATION RECORD

TIME | DATE
1-16-2007

☐ VISIT ☐ CONFERENCE ☒ TELEPHONE

☐ INCOMING
☒ OUTGOING

NAME OF PERSON(S) CONTACTED OR IN CONTACT
David L. Weber, RSO

ORGANIZATION (OFFICE, DEPT. ETC.)
Midwest Testing, Inc.
3377 Hollenberg Drive
Bridgeton, MO 63044

TELEPHONE NO.
314-574-0437
FAX -739-5429

SUBJECT

Control No. 315637 -- Re - Request for license renewal (24-24619-02) by application dated July 21, 2006.

SUMMARY

In a telephone conversation with Mr. Weber on this date, the following was discussed:

In accord with current NRC policy in support of security issues, we discussed the need for the licensee to specify a total possession limit, to be added to License Item 8 for each radioisotope (Cesium-137 and Americium-241) listed in License Item 6 of the subject license. I related that the possession limit did not have to be limited to their current total possession but could allow for temporary replacement gauge(s) and/or for planned expansions in gauge usage. After discussion, Mr. Weber requested a possession limit of 275 millicuries for Cesium-137 and 1250 millicuries for Americium-241.

ACTION REQUIRED

Licensee reviewer to incorporate above information into license, as renewed.


In accordance with 10 CFR 2.390 of the NRC's "Rules of Practice," a copy of this conversation record will be available electronically for public inspection in the NRC Public Document Room or from the Publicly Available Records (PARS) component of NRC's document system (ADAMS). The NRC's document system is accessible from the NRC Web site at <http://www.nrc.gov/reading-rm/adams.html> (the Public Electronic Reading Room).

NAME OF PERSON DOCUMENTING CONVERSATION

SIGNATURE

DATE

Loren J. Hueter



1-16-2007

ACTION TAKEN

SIGNATURE

TITLE

DATE