Awarenia.

Lt. Col. William E. Froemming, MC Department of the Army Office of the Surgeon General Washington, D. C. 20315

Dear Col. Froemming:

This is in regard to our telephone conversation regarding requests by the Third U. S. Army Medical Laboratory for additional hydrogen 3 foils for use in gas chromatography devices.

The application specifies Radiation Research Corporation foils in a Packard Instrument's Model 807 Detector Cell. The information we have concerning Packard Instruments devices specifies U. S. Radium Corporation foils. Also, the Model 807 Detector Cell is for use with nickel 63 foils and not for tritium foils. In order to amend the license, we need to know the quantity of activity which will be in each foil and the model number of the detector cell or of the oven assembly in which the foils will be used.

Sincerely yours.

Nathan Bassin Isotopes Branch Division of Materials Licensing

bcc: Compliance, Region II

OFFICE DME IB

SURNAME NBSSin/jls

DATE 1/6/67

Form AEC-318 (Rev. 9-53)