NO.: TX-1153-S-101-S (Supercedes:TX-1068-S-102-S) DATE: December 12, 2006

PAGE 1 OF 7

High Ex-2

<u>DEVICE TYPE</u>: Sealed source

 MODEL:
 BM0l Series (BM01-10, BM01-15, BM01-20, BM01-99)
 BM02 Series (BM02-05, BM02-10, BM02-15)
 BM04 Series (BM04-10, BM04-15, BM04-99)
 BM05 Series (BM05-10, BM05-05)
 BM09 Series (BM09-10, BM09-75, BM09-99)
 BM09 Series (BM09-10, BM09-75, BM09-99)

MANUFACTURER:

International Isotopes Idaho, Inc. 4137 Commerce Circle Idaho Falls, ID 83401

DISTRIBUTOR:

International Isotopes Idaho, Inc. 4137 Commerce Circle Idaho Falls, ID 83401

ISOTOPE:

Co-57

(b)(2)High

MODEL NUMBER

MAXIMUM ACTIVITY

BM01 Series BM02 Series BM04 Series BM05 Series BM09 Series

6 months

PRINCIPAL USE: Medical reference source

CUSTOM SOURCE: ____ YES X NO

CUSTOM USER: No

LEAK TEST FREQUENCY:

B-4

<u>NO.</u>: TX-1153-S-101-S (Supercedes TX-1068-S-102-S) <u>DATE</u>: December 12, 2006

<u>PAGE 2 OF 7</u>

<u>DEVICE TYPE</u>: Sealed source

Dhusiaal Die

DESCRIPTION:

These sources consist of Co-57 as cobaltous chloride uniformly dispersed in high impact epoxy casting resin (E & C Stycast 1264 or equivalent) which is cured and placed in a 1/8 inch thick ABS encapsulation. Dimensions are given below. The encapsulation consists of two **pre-formed pieces of ABS plastic** that interlock upon assembly and are chemically welded so that disassembly without destruction of the encapsulation is not possible. Each source is supplied in a shielded storage case.

Model Mulloel	r nysicar Dunchsions		
BM01 Series	25.28 in. (64.21 cm) X 17.89 in. (45.44 cm) X 0.7 in. (1.78 cm)		
BM02 Series	20.28 in. (51.51 cm) X 0.7 in. (1.78 cm)		
BM04 Series	19.62 in. (49.83 cm) X 15.56 in. (39.52 cm) X 0.71 in. (1.8 cm)		
BM05 Series	10 in. (25.4 cm) X 10 in. (25.4 cm) X 0.71 in. (1.8 cm)		
BM09 Series	16.42 in. (41.71 cm) X 10.87 in. (27.61 cm) X 0.71 in. (1.8 cm)		

LABELING:

Each source and storage case is conspicuously labeled "Co-57 Flood Source" and bears the warning "CAUTION: RADIOACTIVE MATERIAL" as well as the trefoil radiation symbol in magenta on a yellow background, the manufacturer's name and logo, and the instruction "See Use and Handling Instructions". The radioactive content, in millicuries, and the date manufacture is shown, as well as model number and serial number.

DIAGRAM:

See Attachments 1 - 7. Attachment 1 is a drawing for Models BM01-10, BM01-15, BM01-20 and BM01-99. Attachment 2 is a drawing for Models BM02-05, BM02-10 and BM02-99. Attachment 3 is the label to be affixed to each source. Attachment 4 is a drawing for Models BM04-10, BM04-15 and BM04-99. Attachment 5 is a drawing for Models BM05-10 and BM05-99. Attachment 6 is a drawing for Models BM09-10, BM09-75, and BM09-99. Attachment 7 shows details of the source construction for all models.

CONDITIONS OF NORMAL USE:

The sources are designed for use in a medical environment and are not expected to experience extreme environmental factors. The flood sources provide a uniform field of radiation for evaluation of nuclear medicine gamma camera performance, allowing detection and correction of any camera malfunction prior to diagnostic use. The useful life of the flood sources is approximately 2 years.

<u>NO.</u>: TX-1153-S-101-S (Supercedes TX-1068-S-102-S)

<u>DATE</u>: December 12, 2006

<u>PAGE 3 OF 7</u>

DEVICE TYPE: Sealed source

PROTOTYPE TESTING:

The BM01 and BM02 Series sources were tested in accordance with ANSI N542-1977 "Sealed Radiation Sources, Classification" and achieved a sealed source classification of ANSI 77C22312. The BM04 and BM05 Series sources were tested in accordance with ANSI N43.6-1997 "Sealed Radioactive Source - Classification" and achieved a sealed source classification of ANSI 97C22312. Leak testing of prototypes for models BM01 through BM05 was performed in accordance with ISO 9978. No prototype testing was performed for the model BM09 Series. The ANSI classification of 97C22312 was assigned to this Series by the manufacturer based on the similarity of the construction materials and methods with sources previously tested and currently in production. The manufacturer has also elected not to perform leak testing of prototype sources in this Series for the same reasons of similarity of materials and construction to existing sources.

EXTERNAL RADIATION LEVELS:

 $100 \text{ cm} - 0.1 \text{ mrem/hr} (1 \ \mu \text{Sv/hr})$

Measured radiation dose rates for a nomina	iource are as follows: Huh Ex-2
BM01 Series	BM02 Series
Contact - 35 mrem/hr (350 μ Sv/hr) 5 cm - 35 mrem/hr (350 μ Sv/hr) 30 cm - 6 mrem/hr (60 μ Sv/hr) 100 cm - 1.1 mrem/hr (11 μ Sv/hr)	Contact - 45 mrem/hr (450 μ Sv/hr) 5 cm - 30 mrem/hr (300 μ Sv/hr) 30 cm - 6 mrem/hr (60 μ Sv/hr) 100 cm - 1.1 mrem/hr (11 μ Sv/hr)
BM04 Series	BM05 Series
Contact - 45 mrem/hr (450 μ Sv/hr) 5 cm - 36 mrem/hr (360 μ Sv/hr) 30 cm - 6 mrem/hr (60 μ Sv/hr)	Contact - 90 mrem/hr (900 μ Sv/hr) 5 cm - 60 mrem/hr (600 μ Sv/hr) 30 cm - 8 mrem/hr (80 μ Sv/hr)

 $100 \text{ cm} - 0.2 \text{ mrem/hr} (2 \mu \text{Sv/hr})$

<u>NO.</u>: TX-1153-S-101-S (Supercedes TX-1068-S-102-S) DATE: December 12, 2006

BM09-10^{(b)(2)High}

 $100 \text{ cm} - 0.2 \text{ mrem/hr} (2 \ \mu \text{Sv/hr})$

PAGE 4 OF

DEVICE TYPE: Sealed source

EXTERNAL RADIATION LEVELS:

(Continued)

Calculated radiation dose rates for sources in the BM09 Series are as follows:

BM09-75^{(b)(2)High}

Contact - 54 mrem/hr (540 μ Sv/hr) 5 cm - 36 mrem/hr (360 μ Sv/hr) 30 cm - 4.8 mrem/hr (48 μ Sv/hr) 100 cm - 0.2 mrem/hr (2 μ Sv/hr)

BM09-99

Contact - 108 mrem/hr (1.08 mSv/hr) 5 cm - 72 mrem/hr (720 μ Sv/hr) 30 cm - 9.6 mrem/hr (96 μ Sv/hr) 100 cm - 0.3 mrem/hr (3 μ Sv/hr)

QUALITY ASSURANCE AND CONTROL:

International Isotopes, Inc. and/or RadQual, LLC quality assurance staff will monitor the product quality by performing periodic quality assurance audits of the manufacturer.

LIMITATIONS AND/OR OTHER CONSIDERATIONS OF USE

• The sources shall be distributed to persons specifically licensed by the NRC, an Agreement State or a Licensing State.

- Handling, storage, use, transfer and disposal to be determined by the licensing authority but should be, at a minimum, in accordance with the product information pamphlet provided by the distributor.
- The sources shall be leak tested at intervals not to exceed 6 months using techniques capable of detecting 0.005 microcurie (185 Bq) of removable contamination.
- The sources shall not be subjected to conditions that exceed their ANSI N542-1977 or ANSI N43.6-1997 classifications of ANSI 77C22312 or ANSI 97C22312.

Contact - 72 mrem/hr (720 μ Sv/hr) 5 cm - 48 mrem/hr (480 μ Sv/hr) 30 cm - 6.4 mrem/hr (64 μ Sv/hr)

<u>NO.</u>: TX-1153-S-101-S (Supercedes TX-1068-S-102-S) DATE: December 12, 2006

PAGE 5 OF 7

DEVICE TYPE: Sealed source

LIMITATIONS AND/OR OTHER CONSIDERATIONS OF USE:

(Continued)

- The probable effect of severe environmental conditions, such as accidents and fire, would be minimal release of radioactivity since the Cobalt-57 is incorporated into the cured epoxy matrix. Although the ABS encapsulation, even as it commenced to melt, would prevent dispersion of radioactivity, temperatures exceeding 200 degrees C would result in combustion of the resin in which the Cobalt-57 is dispersed, as well as the plastic encapsulation.
- Care should be taken to avoid contact with organic solvents, hot surfaces, and excessive mechanical stress.
- The sources shall be leak tested at intervals not to exceed 6 months using techniques capable of detecting the presence of 0.005 microcurie (185 Bq) of removable contamination.
- The sources should be stored and transported in the manufacturer's shielded case.
- This registration sheet and the information contained within the references shall not be changed without the written consent of the Texas Department of Health, Bureau of Radiation Control.

SAFETY ANALYSIS SUMMARY:

The BM01, BM02, BM04, BM05 and BM09 Series sources are intended for use as quality control and reference sources for medical imaging devices and are expected to maintain their integrity throughout the useful life of the source if not subjected to extreme conditions of handling or environment. Rupture of the encapsulating material would still not reasonably be expected to allow dispersion of radioactive material due to the epoxy matrix with which the radioactive material is mixed.

Based on review of the information and test data submitted for the BM01, BM02, BM04 and BM05 Series sources, the references cited below, and the assurance of International Isotopes Idaho, Incorporated concerning the manufacturing processes for the BM09 Series, we conclude the these sources are acceptable for licensing purposes as described within this certificate.

Furthermore, we conclude that the sources would be expected to maintain their containment integrity for normal conditions of use and accidental conditions that might occur during uses specified in this certificate.

MANUFACTURING AND DISTRIBUTION CONTROLS:

International Isotopes Idaho, Inc. provides for design control, procurement control, process quality control, and final quality assurance with oversight by International Isotopes, Inc. and/or RadQual, LLC. Distribution will be by drop shipment directly from the manufacturer to the customer.

<u>NO.</u>: TX-1153-S-101-S (Supercedes TX-1068-S-102-S)

DATE: December 12, 2006

PAGE 6 OF 7

DEVICE TYPE: Sealed source

Licensees in possession of sources that have decayed below their useful range of activities should contact International Isotopes Idaho, Inc. for instructions regarding return to the manufacturer. In most cases this will be as limited quantity of radioactive material as specified in 49 CFR 173.421.

MANUFACTURER'S SAFETY ANALYSIS OF SEALED SOURCE REVIEW:

Prototype sources for models BM01 and BM02 were constructed and subjected to environmental testing as provided in ANSI N542-1977, Table 1, and achieved a classification of 77C22312. Prototype sources for models BM04 and BM05 were constructed and subjected to environmental testing as provided in ANSI N43.6-1997 and achieved a classification of 97C22312. No environmental testing was performed on prototype sources of the BM09 Series. The manufacturer has judged the BM09 Series sources to have an ANSI classification of 97C2212 based on the similarity of materials and construction methods with sources already in production from this same manufacturer.

Maximum external radiation levels for the BM01 through BM09 Series sources are as previously identified in this certificate.

REFERENCES:

The following supporting documents for the Series BM01 through BM09 flood sources are hereby incorporated by reference and are made a part of this registry document.

- International Isotopes, Inc. application dated December 19, 2000, signed by Bruce Sanza, CHP, Radiation Safety Officer, with enclosures thereto.
- International Isotopes, Inc. letter dated March 13, 2001, signed by Bruce Sanza, CHP, Radiation Safety Officer, with enclosures thereto.
- International Isotopes, Inc. letters dated April 13, 2001, April 18, 2001, with enclosures thereto.
- International Isotopes Idaho, Inc. Quality Assurance Manual, dated October 1, 1999, received April 16, 2001.
- International Isotopes, Inc. letter dated June 21, 2001, signed by Bruce Sanza, CHP, Radiation Safety Officer.
- International Isotopes, Inc. letter dated October 25, 2001, signed by Bruce Sanza, CHP, Radiation Safety Officer.

<u>NO.</u>: TX-1153-S-101-S (Supercedes TX-1068-S-102-S) DATE: December 12, 2006

<u>PAGE 7 OF 7</u>

DEVICE TYPE: Sealed source

REFERENCES:

(Continued)

- NRX Acquisition Corp. letter dated October 25, 2001, signed by Bruce Sanza, CHP, Radiation Safety Officer.
- International Isotopes, Inc. letter dated November 13, 2001, signed by Bruce Sanza, CHP, Radiation Safety Officer.
- International Isotopes Idaho, Inc. letters dated January 8, 2002 and January 22, 2002, signed by John J. Miller, CHP, Radiation Safety Officer.
 - International Isotopes Idaho, Inc. letters dated March 29, 2002 and April 11, 2002, signed by John J. Miller, CHP, Radiation Safety Officer.
- International Isotopes Idaho, Inc. letter dated February 25, 2003, signed by John J. Miller, CHP, Radiation Safety Officer and RadQual letter Dated August 5, 2003, signed by Keith C. Allberg.
- International Isotopes Inc. letter dated October 10, 2006, signed by John J. Miller, CHP, Radiation Safety Officer.

ISSUING AGENCY:

Texas Department of Health Bureau of Radiation Control

Date:	December 12, 2006	Reviewer: A. fut lac
Data	December 12, 2006	Concurrence:
Dale:	December 12, 2000	David B. Fogle
		()



<u>NO.</u>: TX-1153-S-101-S (Supercedes TX-1068-S-102-S)

<u>DATE</u>: December 12, 2006

ATTACHMENT 1

BM01 Series (Rectanglar)

· .



<u>NO.</u>: TX-1153-S-101-S (Supercedes TX-1068-S-102-S) DATE: December 12, 2006

ATTACHMENT 2

BM02 Series (Circular)



APPIX SOURCE LABEL IN RECESS

NO.: TX-1153-S-101-S (Supercedes TX-1068-S-102-S) DATE: December 12, 2006

ATTACHMENT 3

SOURCE LABEL.

Bench/mark

Cobalt-57 Flood Source



The Bureau of Radiation Control, Texas Department of Health, has approved this sealed source for distribution to persons licensed pursuant to 25 TAC 289.252 or under equivalent licenses of the USNRC, Agreement State, Licensing State, and outside the United States, to persons authorized by the appropriate authority.

See Use and Handling Instructions

Manufactured and distributed for RadQual, LLC, Aurora, OH, by International Isotopes Idaho, Inc., Idaho falis, 10

FL0001-1201







.



<u>NO.</u>: TX-1153-S-101-S <u>DAT</u> (Supercedes TX-1068-S-102-S)

DATE: December 12, 2006

ATTACHMENT 7

Source Assembly Detail (all models)



ENCAPSULATION BASE