



SOUTHWEST RESEARCH INSTITUTE®

6220 Culebra Road, P.O. Drawer 28510
Institute Quality Systems
Institute Calibration Laboratory
Phone: 210-522-5215 Fax 210-522-4834



Certificate of Calibration

0972-01

Submitted By: DIV20

Address: B57

Contact: DON BANNON

Manufacturer / Model: OMEGA / Type K

Description: THERMOCOUPLE

Serial No: 11117

Asset No: 011117

Procedure: TEMPERATURE PROBES - 5 JUNE, 2006

Work Order: 303072416

Date Issued: Jan 3, 2007

Calibration Date: Jan 3, 2007

***Calibration Due:** Jul 3, 2007

Calibration Location: Bldg. 64

Environment: Temp. 73.0°F Hum. 40 %RH

****Data Type:** FOUND-LEFT

DivID/Location: N/A

This certificate documents traceability to the National Institute of Standards and Technology (NIST) and the International System of Units (SI). The Laboratory quality system conforms to ISO/IEC 17025, 2005, ANSI/NC SL Z540-1-1994 and relevant requirements of the ISO 9000-2000 standard. This certificate shall not be reproduced, except in full, without the written approval of the Southwest Research Institute Calibration Laboratory. This certificate shall not be used to claim product endorsement by Southwest Research Institute, American Association for Laboratory Accreditation (A2LA) or any agency of the U. S. Government. Results of this calibration relate only to the instrument described above at the time of calibration and does not imply any long term stability of the instrument.

*Determined by the customer, does not imply the instrument will remain within tolerance as any number of factors may cause an out-of-tolerance condition before this date. **Found/Left = adjustment and/or repair was not required, As Left = adjusted and/or repaired was required. The client has sole responsibility for determination of in-/out-of-tolerance or compliance/noncompliance. See Remarks or attached Measurement Report with the same Work Order number for data.

Reported uncertainty calculated in accordance with the ISO "Guide to the Expression of Uncertainty in Measurement" (GUM) and represents an expanded uncertainty with a coverage factor of k=2 to approximate a 95% confidence level.

Remarks: Cal'd at 900,1000,and1100 °C

Standards Used

Asset No.	Serial No.	Manufacturer	Model	Description	Cal Due
009137	A21208	HART SCIENTIFIC	1575	SUPER THERMOMETER	Jun 27, 07
010281	0421	HART SCIENTIFIC	5628	SPRT	Jun 24, 08
010814	A44625	HART SCIENTIFIC	1529	THERMOCOUPLE THERMOMETER	Oct 17, 07
011449	11449	OMEGA	TJ36-CAXL-14U-2	THERMOCOUPLE	Mar 16, 07

Reviewed by: blt () jrg () pwc (*) wgh ()

Metrology Technician

m:\a2la1.rpt Rev date August 15, 2005

Measurements by: Bob Trollinger
Metrology Technician

Southwest Research Institute
 Calibration Laboratory
 Measurement Report

Work Order: 303072416	Mfr: OMEGA	Technician: blt
Asset No: 011117	Model: TYPE K	
Serial No: 11117	Type: THERMOCOUPLE	Cal Date: 03-Jan-07
Remarks: Limits are based on ASTM E230-02		

Function/Range	Test Point	TI Reading	Difference	+/-Limits	+/-Uncertainty	Found/Left
TEMPERATURE	°C	°C	°C	°C	°C	Result
	898.287	899.90	-1.61	6.7	1.7	Pass
	997.287	995.90	1.39	7.5	1.7	Pass
	1097.778	1097.47	0.31	8.2	2.0	Pass

END OF REPORT