



SOUTHWEST RESEARCH INSTITUTE®

6220 Culebra Road, P.O. Drawer 28510
Institute Quality Systems
Institute Calibration Laboratory
Phone: 210-522-5215 Fax 210-522-4834



Certificate of Calibration

0972-01

Submitted By: DIV20

Address: B57

Contact: DON BANNON

Manufacturer / Model: FISHER SCIENTIFIC / 15-166A

Description: THERMOMETER, GLASS

Serial No: C96-637

Asset No: 004991

Procedure: THERMOMETERS, GLASS - 11 SEP 2006

Work Order: 303072868

Date Issued: Feb 6, 2007

Calibration Date: Feb 6, 2007

***Calibration Due:** Aug 6, 2007

Calibration Location: Bldg. 64

Environment: Temp. 73.0°F Hum. 40 %RH

****Data Type:** FOUND-LEFT

DivID/Location: N/A

This certificate documents traceability to the National Institute of Standards and Technology (NIST) and the International System of Units (SI). The Laboratory quality system conforms to ISO/IEC 17025, 2005, ANSI/NCSL Z540-1-1994 and relevant requirements of the ISO 9000-2000 standard. This certificate shall not be reproduced, except in full, without the written approval of the Southwest Research Institute Calibration Laboratory. This certificate shall not be used to claim product endorsement by Southwest Research Institute, American Association for Laboratory Accreditation (A2LA) or any agency of the U. S. Government. Results of this calibration relate only to the instrument described above at the time of calibration and does not imply any long term stability of the instrument.

*Determined by the customer, does not imply the instrument will remain within tolerance as any number of factors may cause an out-of-tolerance condition before this date. **Found/Left = adjustment and/or repair was not required, As Left = adjusted and/or repaired was required. The client has sole responsibility for determination of in-/out-of-tolerance or compliance/noncompliance. See Remarks or attached Measurement Report with the same Work Order number for data.

Reported uncertainty calculated in accordance with the ISO "Guide to the Expression of Uncertainty in Measurement" (GUM) and represents an expanded uncertainty with a coverage factor of k=2 to approximate a 95% confidence level.

Remarks: None

Standards Used

| Asset No. | Serial No. | Manufacturer | Model | Description | Cal Due |
|-----------|------------|-----------------|-------|-------------------|------------|
| 009137 | A21208 | HART SCIENTIFIC | 1575 | SUPER THERMOMETER | Jun 27, 07 |
| 010281 | 0421 | HART SCIENTIFIC | 5628 | SPRT | Nov 30, 08 |

[Handwritten Signature]
Reviewed by: blt () pwc () wgh ()
Metrology Technician

[Handwritten Signature]
Measurements by: Bob Trollinger
Metrology Technician

Southwest Research Institute
 Calibration Laboratory
 Measurement Report

| | | | | | |
|-------------|-------------------------------------|--------|-------------------|-------------|-----------|
| Work Order: | 303072868 | Mfr: | FISHER SCIENTIFIC | Technician: | blt |
| Asset No: | 004991 | Model: | 15-166A | Cal Date: | 06-Feb-07 |
| Serial No: | C96-637 | Type: | THERMOMETER | | |
| Remarks: | Accuracy Reference ANSI Z236.1-1983 | | | | |

| Function/Range | Test Point | TI Reading | Difference | +/-Limit | +/-Uncertainty | Found/Left |
|----------------|------------|------------|------------|----------|----------------|------------|
| | °C | °C | °C | °C | °C | Result |
| Temperature | -19.83 | -20.0 | -0.17 | 0.5 | 0.29 | Pass |
| | 0.01 | 0.5 | 0.5 | 0.5 | 0.29 | Pass |
| | 50.02 | 50.5 | 0.5 | 0.5 | 0.29 | Pass |
| | 99.93 | 100.0 | 0.1 | 0.5 | 0.29 | Pass |
| | 150.06 | 150.0 | -0.1 | 0.5 | 0.29 | Pass |

END OF REPORT