INSTRUMENT CALIBRATION RECORD

Inst. Type	Inst. Model	Asset No.	S/N
Ludlum Survey Meter	3	003536	100939
Range	**Year**	**User Name**	
X0.1 - X100	1993	Paul Bartetti	

Date Calibrated	Calibrated by		Calculated Readings	Actual Readings	Remarks
2-6-07	Diane Hubbs	X0.1	100, 200	100, 200	Because of scale-using 200, 2000, 20K, 200K
		X1	1000, 2000	1000, 2000	
		X10	10K, 20K	10K, 20K	Replaced batteries
		X100	100K, 200K	100K, 200K	