



UNITED STATES
NUCLEAR REGULATORY COMMISSION
REGION I
475 ALLENDALE ROAD
KING OF PRUSSIA, PENNSYLVANIA 19406-1415

March 16, 2007

Docket No. 030-37421
Control No. 140159

License No. 37-31226-01

Marcela Bohm-Velez, M.D.
President
Weinstein Imaging Assoc.
5850 Centre Avenue
Pittsburgh, PA 15206

SUBJECT: WEINSTEIN IMAGING ASSOC., REQUEST FOR ADDITIONAL INFORMATION
CONCERNING APPLICATION FOR NEW LICENSE, CONTROL NO. 140159

Dear Dr. Bohm-Vélez:

This is in reference to your application dated February 26, 2007 applying for a Nuclear Regulatory Commission license. In order to continue our review, we need the following additional information:

1. In support of your request to authorize Dr. Barbara Ward as the Radiation Safety Officer (RSO) for materials permitted under 10 CFR 35.200, please provide one of the following, as applicable:
 - a. Previous license number (if issued by NRC) or a copy of the license (if issued by an Agreement State) on which the physician was specifically named as an RSO for the uses requested; or
 - b. Written attestation, signed by a preceptor RSO, that the training and experience specified in 10 CFR 35.50 has been satisfactorily completed and that a level of competency sufficient to function independently as an RSO for the medical uses authorized has been achieved.

NRC Form 313A(RSO) (enclosed) may be used to document this information. Please note that if the training and experience, was received more than 7 years ago, evidence of recentness of training in accordance with 10 CFR 35.59 must also be submitted.

2. Please specify the shielding that will be used for radioactive material preparation, storage and handling, e.g. L-Block, lead fort, etc.

Current NRC regulations and guidance are included on the NRC's website at www.nrc.gov; select **Nuclear Materials; Medical, Industrial, and Academic Uses of Nuclear Material**; then **Toolkit Index Page**. Or you may obtain these documents by contacting the Government Printing Office (GPO) toll-free at 1-888-293-6498. The GPO is open from 7:00 a.m. to 9:00 p.m. EST, Monday through Friday (except Federal holidays).

M. Bohm-Velez
Weinstein Imaging Assoc.

2

We will continue our review upon receipt of this information. Please reply to my attention at the Region I Office and refer to Mail Control No. 140159. If you have any technical questions regarding this deficiency letter, please call Lizette Roldán at (610) 337-5237.

If we do not receive a reply from you within 30 calendar days from the date of this letter, we will assume that you do not wish to pursue your application.

Sincerely,

Original signed by Michelle Beardsley

Michelle Beardsley
Senior Health Physicist
Commercial and R&D Branch
Division of Nuclear Materials Safety

Encl: 313A(RSO)

DOCUMENT NAME: C:\FileNet\ML070780419.wpd

SUNSI Review Complete: LRoldán

After declaring this document "An Official Agency Record" it will be released to the Public.

To receive a copy of this document, indicate in the box: "C" = Copy w/o attach/encl "E" = Copy w/ attach/encl "N" = No copy

OFFICE	DNMS/RI	N	DNMS/RI	N	DNMS/RI			
NAME	LRoldán /LXR3/		MBeardsley /MRB/					
DATE	3/16/07		3/16/07					

OFFICIAL RECORD COPY