

From: Michelle Simmons
To: DRice@SRHS-pa.org; JEsposito@SRHS-pa.org
Date: Thu, Mar 29, 2007 1:47 PM
Subject: Sharon Regional Health Systems Deficiency Letter

License No.: 37-01626 -01
Docket No: 03013670
Control No: 140024

Dear Sir or Madam,

This is in reference to your amendment requesting that Mr. Perna be named Radiation Safety Officer (RSO) on your license. Please be advised, Mr. Perna presently has obligations at 15 NRC licensed facilities.

1. In support of your request, please confirm that the individual fulfilling the responsibilities of the Radiation Safety Officer to
be an individual who:

a. Has obtained written attestation, signed by a preceptor Radiation Safety Officer, that the individual has
satisfactorily completed the requirements in paragraph (e) and in paragraphs
(a)(1)(i) and (a)(1)(ii) or (a)(2)(i)
and (a)(2)(ii) or (b)(1) or (c)(1) or (c)(2) of 10 CFR 35.50, and has achieved a level
of radiation safety
knowledge sufficient to function independently as a Radiation Safety Officer for a
medical use licensee; and

b. Has training in the radiation safety, regulatory issues, and emergency procedures for the types of use for which
a licensee seeks approval. This training requirement may be satisfied by
completing training that is supervised
by a Radiation Safety Officer, authorized medical physicist, authorized nuclear
pharmacist, or authorized user,
as appropriate, who is authorized for the type(s) of use for which the licensee is
seeking approval.

2. In support of your request to authorize Drs. Potts and Sharma for materials permitted under 10 CFR 35.500, please
provide a previous license number (if issued by NRC) or a copy of the license (if issued by an Agreement State) on which
the physician was specifically named as an authorized user for the use requested, or

a. Has completed 8 hours of classroom and laboratory training in basic radionuclide handling techniques specifically
applicable to the use of the device. The training must include:

- (1) Radiation physics and instrumentation;
- (2) Radiation protection;
- (3) Mathematics pertaining to the use and measurement of radioactivity; and
- (4) Radiation biology; and

b. Has completed training in the use of the device for the uses requested
In order to continue our review, we request that you submit a response as soon as possible.

Thank you for your help. If you have any questions, you may send an e-mail or call me at 610-337-6921. Please send an e-mail to confirm receipt of this message.

Please note that you may not reply to this letter by return e-mail. Your reply must be in writing by letter or facsimile (610-337-5269). Please submit the requested information as soon as possible. Please include the mail control number 140024 in your response.

Michelle

CC: LAT

Mail Envelope Properties (460BFB9A.1C8 : 13 : 26225)

Subject: Sharon Regional Health Systems Deficiency Letter
Creation Date Thu, Mar 29, 2007 1:47 PM
From: Michelle Simmons

Created By: MRS5@nrc.gov

Recipients

nrc.gov
kp1_po.KP_DO
LAT CC (LAT)

SRHS-pa.org
DRice (DRice@SRHS-pa.org)
JEsposito (JEsposito@SRHS-pa.org)

Post Office
kp1_po.KP_DO

Route
nrc.gov
SRHS-pa.org

Files	Size	Date & Time
MESSAGE	4971	Thursday, March 29, 2007 1:47 PM
Michelle Simmons.vcf	129	Thursday, March 29, 2007 1:47 PM

Options

Expiration Date: None
Priority: Standard
ReplyRequested: No
Return Notification: None

Concealed Subject: No
Security: Standard

Junk Mail Handling Evaluation Results

Message is not eligible for Junk Mail handling
Message is from an internal sender

Junk Mail settings when this message was delivered

Junk Mail handling disabled by User
Junk List is not enabled
Junk Mail using personal address books is not enabled
Block List is not enabled

Document Availability: ☒ Publicly Available ☐ Non-Publicly Available

☒ Non-Sensitive ☐ Non-Sensitive Copyright ☐ Sensitive ☐ Sensitive Copyright

☐ Immediate Release ☐ Normal Release Date ☐ Delay Release Date

Prepared & SUNSI Review Completed Michelle Simmons / RA / 3/28/07
By: _____

(Name)

(Date)