

**U.S. NUCLEAR REGULATORY COMMISSION**

Date: 4/12/2007

**TELEPHONE CONVERSATION RECORD**

Time: 3:30 PM

Mail Control 140317 License No(s). 29-28041-01 Docket No(s). 03030030  
 or Report No(s).

Name of Licensee: PIC, Physicians Imaging Center

Name of Participant(s): Charles Rose, CEO  
 Dennis Lawyer, NRC

Telephone No. 303-541-0044

Subject: RAI, AU Quals and New Authorization

(NOTE: This will be used as the Documents Title in ADAMS)

Summary: The preceptor attestation of the proposed Authorized User did not include the experience element of Using administrative controls to prevent a medical event and the generator experience did not state that it was performed under the supervision of an authorized user. I also verified that PIC intends to perform leak test counting for other than their facilities as a service.

Action Required: Mr. Rose will get the documentation for the AU.

Document Availability:  Publicly Available  Non-Publicly Available  
 Non-Sensitive  Non-Sensitive Copyright  Sensitive  Sensitive Copyright  
 Immediate Release  Normal Release  Delay Release Date

Prepared & SUNSI Review Completed By: Denis Lawyer / RA / Date: 4/12/2007