



WEST VIRGINIA SCHOOL OF OSTEOPATHIC MEDICINE

OFFICE OF THE DEAN



April 13, 2007

NMSB 2

Licensing Assistance Team
Division of Nuclear Materials Safety
U.S. Nuclear Regulatory Commission, Region I
475 Allendale Rd.
King of Prussia, PA 19406-1415

47-19315-02
03037121

Dear NRC:

Enclosed please find a proposed amendment to the NRC license for the West Virginia School of Osteopathic Medicine (WVSOM). The amendment is very small and its purpose is to approve the calibration of some of our survey instruments by our RSO in accordance with the instrument manufacturers' instructions. Our original license currently allows calibration by manufacturer technicians only.

We respectfully request that this small amendment be approved for several reasons. First, the RSO is qualified to perform the calibration procedures on the survey instruments involved. Second, the procedures are simple and typically performed in-house at most institutions. Finally, the flexibility of having the RSO perform these duties would save WVSOM significant sums of money in maintenance costs that primarily go towards travel time for manufacturer technicians.

Any questions should be directed towards Dr. Christopher M. Butt, RSO. He can be reached at 800.356.7836 x. 242, and his email is cbutt@wvsom.edu. Thank you for your time and consideration.

Best regards,

[Handwritten signature of Michael D. Adelman]

Michael D. Adelman, D.O., J.D.
Vice President for Academic Affairs and Dean

RECEIVED
REGION 1
2007 APR 19 PM 1:47

140411
NMSS/RGN1 MATERIALS-002

NRC FORM 313
(4-2004)
10 CFR 30, 32, 33,
34, 35, 36, 39, and 40

U.S. NUCLEAR REGULATORY COMMISSION

APPROVED BY OMB: NO. 3150-0120

EXPIRES: 10/31/2005

Estimated burden per response to comply with this mandatory collection request: 7 hours. Submittal of the application is necessary to determine that the applicant is qualified and that adequate procedures exist to protect the public health and safety. Send comments regarding burden estimate to the Records and FOIA/Privacy Services Branch (T-5 F52), U.S. Nuclear Regulatory Commission, Washington, DC 20555-0001, or by internet e-mail to infocollects@nrc.gov, and to the Desk Officer, Office of Information and Regulatory Affairs, NEOB-10202, (3150-0120), Office of Management and Budget, Washington, DC 20503. If a means used to impose an information collection does not display a currently valid OMB control number, the NRC may not conduct or sponsor, and a person is not required to respond to, the information collection.

APPLICATION FOR MATERIAL LICENSE

INSTRUCTIONS: SEE THE APPROPRIATE LICENSE APPLICATION GUIDE FOR DETAILED INSTRUCTIONS FOR COMPLETING APPLICATION. SEND TWO COPIES OF THE ENTIRE COMPLETED APPLICATION TO THE NRC OFFICE SPECIFIED BELOW.

APPLICATION FOR DISTRIBUTION OF EXEMPT PRODUCTS FILE APPLICATIONS WITH:

DIVISION OF INDUSTRIAL AND MEDICAL NUCLEAR SAFETY
OFFICE OF NUCLEAR MATERIALS SAFETY AND SAFEGUARDS
U.S. NUCLEAR REGULATORY COMMISSION
WASHINGTON, DC 20555-0001

ALL OTHER PERSONS FILE APPLICATIONS AS FOLLOWS:

IF YOU ARE LOCATED IN:

ALABAMA, CONNECTICUT, DELAWARE, DISTRICT OF COLUMBIA, FLORIDA, GEORGIA, KENTUCKY, MAINE, MARYLAND, MASSACHUSETTS, MISSISSIPPI, NEW HAMPSHIRE, NEW JERSEY, NEW YORK, NORTH CAROLINA, PENNSYLVANIA, PUERTO RICO, RHODE ISLAND, SOUTH CAROLINA, TENNESSEE, VERMONT, VIRGINIA, VIRGIN ISLANDS, OR WEST VIRGINIA, SEND APPLICATIONS TO:

LICENSING ASSISTANCE TEAM
DIVISION OF NUCLEAR MATERIALS SAFETY
U.S. NUCLEAR REGULATORY COMMISSION, REGION I
475 ALLENDALE ROAD
KING OF PRUSSIA, PA 19406-1415

IF YOU ARE LOCATED IN:

ILLINOIS, INDIANA, IOWA, MICHIGAN, MINNESOTA, MISSOURI, OHIO, OR WISCONSIN, SEND APPLICATIONS TO:

MATERIALS LICENSING BRANCH
U.S. NUCLEAR REGULATORY COMMISSION, REGION III
2443 WARRENVILLE ROAD, SUITE 210
LISLE, IL 60532-4352

ALASKA, ARIZONA, ARKANSAS, CALIFORNIA, COLORADO, HAWAII, IDAHO, KANSAS, LOUISIANA, MONTANA, NEBRASKA, NEVADA, NEW MEXICO, NORTH DAKOTA, OKLAHOMA, OREGON, PACIFIC TRUST TERRITORIES, SOUTH DAKOTA, TEXAS, UTAH, WASHINGTON, OR WYOMING, SEND APPLICATIONS TO:

NUCLEAR MATERIALS LICENSING BRANCH
U.S. NUCLEAR REGULATORY COMMISSION, REGION IV
611 RYAN PLAZA DRIVE, SUITE 400
ARLINGTON, TX 76011-4005

03037121

PERSONS LOCATED IN AGREEMENT STATES SEND APPLICATIONS TO THE U.S. NUCLEAR REGULATORY COMMISSION ONLY IF THEY WISH TO POSSESS AND USE LICENSED MATERIAL IN STATES SUBJECT TO U.S. NUCLEAR REGULATORY COMMISSION JURISDICTIONS.

<p>1. THIS IS AN APPLICATION FOR (Check appropriate item)</p> <p><input type="checkbox"/> A. NEW LICENSE</p> <p><input checked="" type="checkbox"/> B. AMENDMENT TO LICENSE NUMBER <u>47-19315-02</u></p> <p><input type="checkbox"/> C. RENEWAL OF LICENSE NUMBER _____</p>	<p>2. NAME AND MAILING ADDRESS OF APPLICANT (Include ZIP code)</p> <p>West Virginia School of Osteopathic Medicine 400 North Lee Street Lewisburg, WV 24901</p>
<p>3. ADDRESS WHERE LICENSED MATERIAL WILL BE USED OR POSSESSED</p> <p>West Virginia School of Osteopathic Medicine 400 North Lee Street Lewisburg, WV 24901</p>	<p>4. NAME OF PERSON TO BE CONTACTED ABOUT THIS APPLICATION</p> <p>Christopher M. Butt, Ph.D., RSO</p> <p>TELEPHONE NUMBER</p> <p>(304) 647-6242</p>

SUBMIT ITEMS 5 THROUGH 11 ON 8-1/2 X 11" PAPER. THE TYPE AND SCOPE OF INFORMATION TO BE PROVIDED IS DESCRIBED IN THE LICENSE APPLICATION GUIDE.

<p>5. RADIOACTIVE MATERIAL</p> <p>a. Element and mass number; b. chemical and/or physical form; and c. maximum amount which will be possessed at any one time.</p>	<p>6. PURPOSE(S) FOR WHICH LICENSED MATERIAL WILL BE USED.</p>				
<p>7. INDIVIDUAL(S) RESPONSIBLE FOR RADIATION SAFETY PROGRAM AND THEIR TRAINING EXPERIENCE.</p>	<p>8. TRAINING FOR INDIVIDUALS WORKING IN OR FREQUENTING RESTRICTED AREAS.</p>				
<p>9. FACILITIES AND EQUIPMENT.</p>	<p>10. RADIATION SAFETY PROGRAM.</p>				
<p>11. WASTE MANAGEMENT.</p>	<p>12. LICENSE FEES (See 10 CFR 170 and Section 170.31)</p> <table border="1"> <tr> <td>FEE CATEGORY</td> <td>Exempt 170.11</td> <td>AMOUNT ENCLOSED</td> <td>\$ 0.00</td> </tr> </table>	FEE CATEGORY	Exempt 170.11	AMOUNT ENCLOSED	\$ 0.00
FEE CATEGORY	Exempt 170.11	AMOUNT ENCLOSED	\$ 0.00		

13. CERTIFICATION. (Must be completed by applicant) THE APPLICANT UNDERSTANDS THAT ALL STATEMENTS AND REPRESENTATIONS MADE IN THIS APPLICATION ARE BINDING UPON THE APPLICANT.

THE APPLICANT AND ANY OFFICIAL EXECUTING THIS CERTIFICATION ON BEHALF OF THE APPLICANT, NAMED IN ITEM 2, CERTIFY THAT THIS APPLICATION IS PREPARED IN CONFORMITY WITH TITLE 10, CODE OF FEDERAL REGULATIONS, PARTS 30, 32, 33, 34, 35, 36, 39, AND 40, AND THAT ALL INFORMATION CONTAINED HEREIN IS TRUE AND CORRECT TO THE BEST OF THEIR KNOWLEDGE AND BELIEF.

WARNING: 18 U.S.C. SECTION 1001 ACT OF JUNE 25, 1948 62 STAT. 749 MAKES IT A CRIMINAL OFFENSE TO MAKE A WILLFULLY FALSE STATEMENT OR REPRESENTATION TO ANY DEPARTMENT OR AGENCY OF THE UNITED STATES AS TO ANY MATTER WITHIN ITS JURISDICTION.

CERTIFYING OFFICER -- TYPED/PRINTED NAME AND TITLE	SIGNATURE	DATE
Michael D. Adelman, D.O./VP Academic Affairs and Dean	<i>Michael D. Adelman</i>	4/13/07

FOR NRC USE ONLY

TYPE OF FEE	FEE LOG	FEE CATEGORY	AMOUNT RECEIVED	CHECK NUMBER	COMMENTS
			\$		
APPROVED BY				DATE	

Amended Application; West Virginia School of Osteopathic Medicine

Proposed Amendment

10. Radiation Safety Program

- Radiation Monitoring Instruments

Our radiation monitoring instruments are:

1. Bicron GM Surveyor (Model 2000) with PGM pancake probe and check source. The meter will be calibrated by Physics Associates of Roanoke, VA (NRC License #45-17344-01);
2. Wallac Microbeta 1450-021 96-well liquid scintillation counter, maintained and calibrated by the RSO or by Perkin-Elmer in accordance with the manufacturer's instructions; and
3. an LKB Minigamma 1275 counter, maintained and calibrated by the RSO or by Perkin-Elmer in accordance with the manufacturer's instructions.

These instruments meet the radiation monitoring instrument specifications published in Appendix M to NUREG 1556, volume 7. We reserve the right to upgrade these instruments as necessary to meet regulatory requirements.

Previously approved (Page 6 of 15 in original application)

10. Radiation Safety Program

- Radiation Monitoring Instruments

Our radiation monitoring instruments are:

1. Bicron GM Surveyor (Model 2000) with PGM pancake probe and check source. The meter will be calibrated by Physics Associates of Roanoke, VA (NRC License #45-17344-01);
2. Wallac Microbeta 1450-021 96-well liquid scintillation counter, maintained and calibrated by Perkin-Elmer; and
3. an LKB Minigamma 1275 counter, maintained and calibrated by Perkin-Elmer.

These instruments meet the radiation monitoring instrument specifications published in Appendix M to NUREG 1556, volume 7. We reserve the right to upgrade these instruments as necessary to meet regulatory requirements.

This is to acknowledge the receipt of your letter/application dated

4/13/2007, and to inform you that the initial processing which includes an administrative review has been performed.

Allowed. 47-19315-02 There were no administrative omissions. Your application was assigned to a technical reviewer. Please note that the technical review may identify additional omissions or require additional information.

Please provide to this office within 30 days of your receipt of this card

A copy of your action has been forwarded to our License Fee & Accounts Receivable Branch, who will contact you separately if there is a fee issue involved.

Your action has been assigned **Mail Control Number** 140411.
When calling to inquire about this action, please refer to this control number.
You may call us on (610) 337-5398, or 337-5260.