



SOUTHWEST RESEARCH INSTITUTE®

6220 Culebra Road, P.O. Drawer 28510
Institute Quality Systems
Institute Calibration Laboratory
Phone: 210-522-5215 Fax 210-522-4834



Certificate of Calibration

0972-01

Submitted By: DIV20

Address: B57

Contact: DON BANNON

Manufacturer / Model: SWRI / 100K-10M OHMS

Description: RESISTOR BOX

Serial No: 7243

Asset No: 007243

Procedure: DECADE RESISTOR TO 100 MOHM-STEP - 21 MAR 2006

Work Order: 303073374

Date Issued: Mar 7, 2007

Calibration Date: Mar 7, 2007

*Calibration Due: Sep 7, 2007

Calibration Location: Bldg. 64

Environment: Temp. 73.0°F Hum. 40 %RH

**Data Type: FOUND-LEFT

DivID/Location: N/A

This certificate documents traceability to the National Institute of Standards and Technology (NIST) and the International System of Units (SI). The Laboratory quality system conforms to ISO/IEC 17025, 2005, ANSI/NCSL Z540-1-1994 and relevant requirements of the ISO 9000-2000 standard. This certificate shall not be reproduced, except in full, without the written approval of the Southwest Research Institute Calibration Laboratory. This certificate shall not be used to claim product endorsement by Southwest Research Institute, American Association for Laboratory Accreditation (A2LA) or any agency of the U. S. Government. Results of this calibration relate only to the instrument described above at the time of calibration and does not imply any long term stability of the instrument.

*Determined by the customer, does not imply the instrument will remain within tolerance as any number of factors may cause an out-of-tolerance condition before this date. **Found/Left = adjustment and/or repair was not required, As Left = adjusted and/or repaired was required. The client has sole responsibility for determination of in-/out-of-tolerance or compliance/noncompliance. See Remarks or attached Measurement Report with the same Work Order number for data.

Reported uncertainty calculated in accordance with the ISO "Guide to the Expression of Uncertainty in Measurement" (GUM) and represents an expanded uncertainty with a coverage factor of k=2 to approximate a 95% confidence level.

Remarks: None

Standards Used

| Asset No. | Serial No. | Manufacturer | Model | Description | Cal Due |
|-----------|------------|-----------------|---------------|-------------|------------|
| 007001 | 2823A21362 | HEWLETT-PACKARD | 3458A/OPT 002 | MULTIMETER | Feb 26, 08 |

Reviewed by: blt () eg () pwc () wgh ()

Metrology Technician

m:\a2la1.rpt Rev date August 15, 2005

Measurements by: Bob Trollinger

Metrology Technician

Southwest Research Institute
Calibration Laboratory
Measurement Report

| | | | | | |
|-------------|-----------|-------|--------------|------------|----------|
| Work Order: | 303073374 | Mfr. | SWRI | Technician | bit |
| Asset No. | 007243 | Model | 100k-10M OHM | | |
| Serial No. | 7243 | Type. | Resistors | Cal Date. | 7-Mar-07 |
| Remarks: | | | | | |

| Function/Range | Test Point | TI Reading | Difference | +/-Limit | +/-Uncertainty | Found/Left |
|----------------|------------|------------|------------|----------|----------------|------------|
| Resistance | kOhm | kOhm | kOhm | kOhm | kOhm | Result |
| | 100.00 | 99.73 | -0.27 | 1.00 | 0.0018 | Pass |
| | MOhm | MOhm | MOhm | MOhm | MOhm | |
| | 10.00 | 9.94 | -0.06 | 0.10 | 0.0010 | Pass |
| END OF REPORT | | | | | | |