

APPENDIX A

NOTICE OF VIOLATION

Paul S. Friedman, M.D.
Philadelphia, Pennsylvania
License No. 37-00147-02

Docket No. 30-2940

Based on the results of an NRC inspection conducted July 3, 1979, it appears that certain of your activities were not conducted in full compliance with NRC regulations and the conditions of your license as indicated below.

- A. Condition 10 of your license limits the use of licensed material to your office at 1422 Chestnut Street, in Philadelphia, Pennsylvania.

Contrary to this requirement, as of July 3, 1979, you have routinely had licensed materials delivered to your residence, an address other than that authorized.

- B. 10 CFR 20.201(b) requires that you make such surveys as may be necessary for you to comply with all sections of Part 20. As defined in 10 CFR 20.101(a), "survey" means an evaluation of the radiation hazards incident to the production, use, release, disposal, or presence of radioactive materials or other sources of radiation under a specific set of conditions.

1. Contrary to this requirement, you failed to make such surveys, of the quantity and distribution of removable contamination in the room where radiomedicines were handled, as were necessary for you to assure that the occupants of this room were not exposed in excess of the limits specified in 10 CFR 20.103, "Exposure of individuals to concentrations of radioactive material in restricted areas". Specifically, it was noted that your GM meter was inoperable on the day of the inspection, July 3, 1979.
2. Contrary to this requirement, as of July 3, 1979, you failed to perform such surveys (evaluations) as were necessary to assure compliance with 10 CFR 20.101, "Exposure of individuals to radiation in restricted areas." Specifically, you failed to provide the necessary personnel monitoring to evaluate the dose to the extremities of individuals preparing and administering radiopharmaceuticals at your office.

These are infractions.

B/3
2200 543

8001070 192