

4399 Corporate Road
Charleston, SC 29405
Phone: (843) 747-7800
Fax: (843) 202-6278



Fax

To: Mr. Bill Reichhold From: Jason Barnett

cc: Date: 5-16-07

Fax: 630-515-1259 or 630-629-9772 Phone: 843-202-6234

Total Pages: 11 Sent by: JAB

Re: request for Radiactive Materials License Amendment - license # 39-32137-01
control # 315266

● Comments:



May 11, 2007

U.S. NRC – Region III
2443 Warrenville Road
Suite 210
Lisle, Illinois 60532-4352

License Number : 39-32137-01
Control Number : 315266

Dear Mr. Bill Reichhold:

This letter is to serve as a request for an amendment to our license regarding the items below;

1. We wish to remove G. Michael Brown as the RSO and add Jason Barnett as the RSO. He has successfully completed the NDC Systems training program – certificate is attached.
2. All gauges are stored within a locked unit that only our custodian has access to. No member of family or public has access to this locked unit. All gauges are stored so that the radiation dose to the public is within the regulatory limits.
3.
 - a. Gauge 2195 was in the possession of Bob Bilodeau at 138 Asher lane, Spring Mills, PA 16875. Bob has left the company so the gauge responsibilities are being transitioned to Mike Brown at 564 Shanelly Lane, Port Matilda, PA 16870. Attached are Mike Brown's gauge training certificate and a storage sketch of the new location.
 - b. Gauge 2569 is in possession of Bill Robertson and the storage location has changed from 3701 Liberty Point, Midlothian, VA to 4475 Pouncey Tract Rd, Unit 1001, Glenn Allen, VA 23059. Attached are Bill Robertson's gauge training certificate and a storage sketch of the new location.
 - c. Leak test for all gauges are attached.
4. We do not have any history of any leaking sealed source.

5. All radionuclide's have been removed from their prior storage locations and transferred to their new storage locations.

If you need additional information to process these changes, please call me at 843.202.6234 or email me at jason.barnett@astenjohnson.com.

Sincerely,



Jason Barnett

Radiation Safety Officer

Certificate of Completion

awarded to

Jason Barnett

for participation in

Radiation Safety Seminar

January 30 - February 1, 2006 - Las Vegas
presented by Engelhardt & Associates, Inc.

Susan Engelhardt

Susan J. Engelhardt, M.S.

Ralph Grunewald

Ralph Grunewald, Ph.D.

Joshua Walkowicz


Joshua Walkowicz, M.S., CHP

Judith Grunewald

Judith Grunewald, R.N., M.S.

CERTIFICATE OF TRAINING

This is to certify that Jason Barnett of Asten Johnson has successfully completed a course in the use, maintenance and repair of the NDC measurement sensors - Models 104, 104P and 104PD. This course was completed at Irwindale, California on November 06, 2006.



*Frank Aguirre
Radiation Safety Officer
NDC Infrared Engineering*

Dated: November 10, 2006

CERTIFICATE OF RADIATION SAFETY TRAINING

This is to certify that G. Michael Brown of Asten Inc. has successfully completed a course in radiation, radiation safety and use of the NDC Systems nuclear measurement device. This course, Number 104, was given at Augusta, Georgia on August 30, 1995.



Daniel Fishman
Radiation Safety Officer
NDC Systems

Dated: December 11, 1995




ASTENJOHNSON

Shoe Press Users Conference

APRIL 25, 2007

G. MICHAEL BROWN
564 SHANELLY DR.
PORT MATILDA, PA 16870
814.482.0253



Palmetto Counting Systems
1416 Drexell Drive
Charleston, S.C. 29407
(843) 763-1255

AstenJohnson Corporation
4399 Corporate Road
Charleston, S.C. 29405

P.O # Verbal
Date of Leak Test: 4/30/07
Date of Next Leak Test: 10/30/07
Date of Report: 5/10/07

Ref: Leak Test Report

Attn: Mr. Jason Barnett
Corporate R.S.O.


Summary: Less than 0.005 uci of removable activity from any of the below listed sources.

Radionuclide	Quantity	Manuf.	Model No.	SN:
Am-241	25 mci	NDC	104PWE	2195

All wipes were counted using a Nucleus-500 system with a NaI detector.

All standard sources have their calibration traceable to the National Institute of Standards and Technology.

The methods used for these tests are capable of detecting less than 0.005 uci of activity.



J. A. Seay Jr.
S.C. License #423

CERTIFICATE OF RADIATION SAFETY TRAINING

This is to certify that Bill Robertson of JWI Group has successfully completed a course in radiation, radiation safety and use of the NDC Systems nuclear measurement device.

This course, Number 104, was given at Atlanta, Georgia on February 20, 1997.


Eugene Murakami
Radiation Safety Officer
NDC Systems

Dated: April 14, 1997


Richmond City

05/16/2007 08:37 8432026278

ASTENJOHNSON

PAGE 10/11

4475
Pouncey Tract Road



1007	1006	1005	1004	1003	1002	1001 my STORAGE UNIT	1000
------	------	------	------	------	------	-------------------------------	------

Page 2569
Robertson, Bill

Palmetto Counting Systems
1416 Drexell Drive
Charleston, S.C. 29407
(843) 763-1255

AstenJohnson Corporation
4399 Corporate Road
Charleston, S.C. 29405

P.O # Verbal
Date of Leak Test: 1/22/07
Date of Next Leak Test: 7/22/07
Date of Report: 1/29/07

Ref: Leak Test Report

Attn: Mr. Jason Barnett
Corporate R.S.O.


Summary: Less than 0.005 uci of removable activity from any of the below listed sources.

Radionuclide	Quantity	Manuf.	Model No.	SN:
Am-241	80 mci	NDC	104PWE	2569

All wipes were counted using a Nucleus-500 system with a NaI detector.

All standard sources have their calibration traceable to the National Institute of Standards and Technology.

The methods used for these tests are capable of detecting less than 0.005 uci of activity.



J. A. Seay, Jr.
S.C. License #423