U.S. NUCLEAR REGULATORY COMMISSION

TELEPHONE CONVERSATION RECORD

Date: 5/31/07

Time:

Mail Control or Report No(s). 140306 License No(s). 29-01698-02 Docket No(s). 03002441

Name of Licensee:	Capital Health System
Name of Participant(s):	Janet Bryant, RSO
Telephone No.	609-394-4000 X1998
Subject: (NOTE: This will be used as the Documents Title in ADAMS)	Voidance of Control 140306
Summary:	In a letter dated 3/22/07, Capital Health System requested authorization of a Cs-137 source (IPL Model HEG-137) for attenuation correction. The requested activity was for 20 millicuries per source. Based on this information, it was determined that the source was covered under 10 CFR 35.65 and did not need to be listed on the license. Therefore the action should be voided.
Action Required:	Void the action

Document Availability: x **Publicly Available** ____ Non-Publicly Available

x Non-Sensitive ____ Non-Sensitive Copyright ____ Sensitive ____ Sensitive Copyright

____ Immediate Release x Normal Release ____ Delay Release Date

Prepared & SUNSI Review Completed By: TLWeidner ***/ RA /*** Date: 5/31/07