

U. S. ATOMIC ENERGY COMMISSION
BYPRODUCT MATERIAL LICENSE

Supplementary Sheet

47-260-2

License Number (160)

AMENDMENT NO. 3

DUPLICATED
FOR DIV. OF INSP.

Union Carbide Olefins Company
Special Instrumentation Division
437 MacCorkle Avenue, S. W.
South Charleston, West Virginia

Attention: L. J. Rogers

Byproduct Material License Number 47-260-2 is hereby amended under Condition 12 to read as follows:

- 15. Sealed sources containing Cesium 137 which are utilized in Ghmart Corporation Model SHRM and Model SHRN gauges shall be tested for leakage and contamination at intervals of not more than three (3) years and records of test results shall be maintained by the licensee. Leak testing shall be performed by H. H. Ketcham as specified in letter of September 5, 1958, from L. J. Rogers.

*Amend 4
12-31-59
JMB*

Date April 8, 1959

For the U. S. Atomic Energy Commission

Original Signed By
James R. Mason

by Chief, Isotopes Branch
Division of Licensing and Regulation
Washington 25, D. C.

meq

REC 4-9-59

A-7