

U.S. NUCLEAR REGULATORY COMMISSION

Date: June 5, 2007

TELEPHONE CONVERSATION RECORD

Time: 12:20

Mail Control 2007-001 License No(s). 29-30646-01 Docket No(s). 03035748
 or Report No(s).

Name of Licensee: Rudy Rezzadeh, M.D.

Name of Participant(s): Ray Quinn, Nuclear Medicine Technologist
 Steven R. Courtemanche, Health Physicist, NRC RI

Telephone No. 201-767-3666
 610-337-5075

Subject: Continuation of Inspection

(NOTE: This will be used as the Documents Title in ADAMS)

Summary: I spoke with Mr. Quinn about the licensee's program. He informed me that he uses a syringe containing between 200 and 300 microcuries of Technetium-99m (Tc-99m) for performing floods of the camera since their Co-57 flood source is so old. The licensee does not provide him with ring badges to measure the exposure to his hands. The whole body badges are collected by Dr. Rezzadeh who sometimes forgets to send them in to the radiation monitor processor. Mr. Quinn stated that he does not receive a ring dosimeter because operations are conducted only one night a week, 4 to 6 patients are administered licensed material, and Dr. Rezzadeh performs the injections. I asked Mr. Quinn if the doses administered to patients were split from the doses received from the radiopharmacy. He stated that they were. I informed him that my review of the dose records showed between 15 and 30 millicuries of activity was administered to patients.

Action Required:

Document Availability: Publicly Available Non-Publicly Available
 Non-Sensitive Non-Sensitive Copyright Sensitive Sensitive Copyright
 Immediate Release Normal Release Delay Release Date
 Prepared & SUNSI Review Completed By: Steven R. Courtemanche / RA / Date: 6/11/2007