

### SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED: <b>Associated Physicians and Surgeons Clinic 221 South 6<sup>th</sup> Street Terre Haute, IN 47807-4214</b>	2. NRC/REGIONAL OFFICE <b>U.S. Nuclear Regulatory Commission Region III 2443 Warrenville Road Suite 210 Lisle, Illinois 60532-4351</b>
--	---

REPORT 2007-001

3. DOCKET NUMBER(S) 030-34580	4. LICENSEE NUMBER(S) 13-32039-01	5. DATE(S) OF INSPECTION June 4, 2007
----------------------------------	--------------------------------------	--

**LICENSEE:**

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:

- 1. Based on the inspection findings, no violations were identified.
- 2. Previous violation(s) closed.
- 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, NUREG-1600, to exercise discretion, were satisfied.

\_\_\_\_\_ Non-Cited Violation(s) was/were discussed involving the following requirement(s) and Corrective Action(s):

- 4. During this inspection certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.

(Violations and Corrective Actions)

**Licensee's Statement of Corrective Actions for Item 4, above.**

I hereby state that, within 30 days, the actions described by me to the inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

Title	Printed Name	Signature	Date
LICENSEE'S REPRESENTATIVE			
NRC INSPECTOR	Deborah A. Piskura		6/04/2007

**Docket File Information**

**SAFETY INSPECTION REPORT  
AND COMPLIANCE INSPECTION**



1. LICENSEE <b>Associated Physicians and Surgeons Clinic</b> REPORT <b>2007-001</b>		2. NRC/REGIONAL OFFICE <b>Region III</b> <b>2443 Warrenville Road, Suite 210</b> <b>Lisle, IL 60532</b>	
3. DOCKET NUMBER(S) 030-34580	4. LICENSE NUMBER(S) 13-32039-01	5. DATE(S) OF INSPECTION June 04, 2007	
6. INSPECTION PROCEDURES USED 87130 & 87131		7. INSPECTION FOCUS AREAS 03.01, 03.02, 03.03, 03.04, 03.05, 03.06, 03.07, and 03.08	

**SUPPLEMENTAL INSPECTION INFORMATION**

1. PROGRAM CODE(S) 02200	2. PRIORITY G 3	3. LICENSEE CONTACT Krishna Mohan Namburi, M.D., RSO	4. TELEPHONE NUMBER 812-232-0564
-----------------------------	--------------------	---	-------------------------------------

Main Office Inspection      Next Inspection Date: June 2010

Field \_\_\_\_\_

Temporary Job Site \_\_\_\_\_

**PROGRAM SCOPE**

This licensee was a private practice physicians group office and authorized to use licensed material permitted by Sections 35.100, 35.200, and 35.300 (limited to I-131 for treatment of hyperthyroidism). The nuclear medicine department was staffed with three full-time technologists who performed approximately 220 diagnostic nuclear cardiology procedures per month. The licensee received unit doses from a licensed radiopharmacy. In March 2006, the clinic resumed its use of I-131 for hyperthyroid treatments. Radioiodine was obtained from the pharmacy in capsule form. In 2006, the clinic administered 16 treatments; 9 I-131 treatments were administered YTD 2007.

This inspection consisted of interviews with licensee personnel, a review of select records, tour of the nuclear medicine department, and independent measurements. The inspector observed licensee personnel prepare, assay and administer several cardiolite unit dosages for cardiac rest and stress testing procedures.