



SOUTHWEST RESEARCH INSTITUTE®

6220 Culebra Road, P.O. Drawer 28510
Institute Quality Systems
Institute Calibration Laboratory
Phone: 210-522-5215 Fax 210-522-4834



Calibration Laboratory
Certificate #0972-01

Certificate of Calibration

Submitted By: DIV20

Address: B57

Contact: DON BANNON

Manufacturer / Model: OMEGA / HH22

Description: THERMOCOUPLE THERMOMETER

Serial No: T-94140

Asset No: 001997

Procedure: DIGITAL THERMOMETERS - 17 APR 07

Work Order: 303074245

Date Issued: Apr 26, 2007

Calibration Date: Apr 26, 2007

***Calibration Due:** Oct 26, 2007

Calibration Location: Bldg. 64

Environment: Temp. 68.0°F Hum. 52 %RH

****Data Type:** FOUND-LEFT

DivID/Location: N/A

This certificate documents traceability to the National Institute of Standards and Technology (NIST) and the International System of Units (SI). The Laboratory quality system conforms to ISO/IEC 17025, 2005, ANSI/NC SL Z540-1-1994 and relevant requirements of the ISO 9000-2000 standard. This certificate shall not be reproduced, except in full, without the written approval of the Southwest Research Institute Calibration Laboratory. This certificate shall not be used to claim product endorsement by Southwest Research Institute, American Association for Laboratory Accreditation (A2LA) or any agency of the U. S. Government. Results of this calibration relate only to the instrument described above at the time of calibration and does not imply any long term stability of the instrument.

*Determined by the customer, does not imply the instrument will remain within tolerance as any number of factors may cause an out-of-tolerance condition before this date. **Found/Left = adjustment and/or repair was not required, As Left = adjusted and/or repaired was required. The client has sole responsibility for determination of in-/out-of-tolerance or compliance/noncompliance. See Remarks or attached Measurement Report with the same Work Order number for data.

Reported uncertainty calculated in accordance with the ISO "Guide to the Expression of Uncertainty in Measurement" (GUM) and represents an expanded uncertainty with a coverage factor of k=2 to approximate a 95% confidence level.

Remarks: None

Standards Used

Asset No.	Serial No.	Manufacturer	Model	Description	Cal Due
004164	6380025	FLUKE	5500A/SC300	CALIBRATOR	Jul 21, 07

Scott Kistner

Reviewed by: () wgh (x) srk () jrg () blt () pwc

Metrology Technician

m:\a2la1.rpt Rev date August 15, 2005

Joe Greagrey

Measurements by: Joe Greagrey

Metrology Technician

Southwest Research Institute
Calibration Laboratory
Measurement Report

Work Order	303074245	Mfr.	OMEGA	Technician	JRG
Asset No.	001997	Model	HH22		
Serial No.	T-94140	Type	Digital Thermometer	Cal Date	26-Apr-07

Remarks:

Function/Range	Test Point	TI Reading	Difference	+/-Limit	+/-Uncertainty	Found/Left
T1 Input	°F	°F	°F	°F	°F	Result
Type J	-300	-300.1	0.1	1.3	0.57	Pass
	110	109.5	0.5	1.1	0.31	Pass
	525	525.0	0.0	1.5	0.37	Pass
	940	940.1	0.1	1.9	0.37	Pass
	1350	1349.7	0.3	2.4	0.37	Pass
Type J	°C	°C	°C	°C	°C	
	-200	-200.1	-0.1	0.8	0.33	Pass
	40	39.6	-0.4	0.6	0.21	Pass
	275	274.9	-0.1	0.9	0.32	Pass
	510	510.0	0.0	1.1	0.32	Pass
	750	749.7	-0.3	1.4	0.32	Pass
Type K	°F	°F	°F	°F	°F	
	-300	-300.3	0.3	1.3	0.70	Pass
	390	390.1	0.1	1.4	0.55	Pass
	1075	1075.6	0.6	2.1	0.55	Pass
	1760	1760.8	0.8	2.8	0.55	Pass
	2450	2451.1	1.1	3.5	0.84	Pass
Type K	°C	°C	°C	°C	°C	
	-150	-150.0	0.0	0.8	0.40	Pass
	210	210.0	0.0	0.8	0.32	Pass
	575	575.3	0.3	1.2	0.32	Pass
	940	940.4	0.4	1.5	0.32	Pass
	1300	1300.6	0.6	1.9	0.47	Pass
T1 Input -T2 Input		0.1	0.1	1.0	0.40	Pass
T2 Input	°F	°F	°F	°F	°F	
Type J	-300	-301.2	1.2	1.3	0.57	Pass
	110	109.6	0.4	1.1	0.31	Pass
	525	525.1	0.1	1.5	0.37	Pass
	940	940.3	0.3	1.9	0.37	Pass
	1350	1349.9	0.1	2.4	0.37	Pass
Type J	°C	°C	°C	°C	°C	
	-200	-200.5	-0.5	0.8	0.33	Pass
	40	39.7	-0.3	0.6	0.21	Pass
	275	275.0	0.0	0.9	0.32	Pass
	510	510.1	0.1	1.1	0.32	Pass
	750	750.0	0.0	1.4	0.32	Pass

Southwest Research Institute
Calibration Laboratory
Measurement Report

Work Order	303074245	Mfr.	OMEGA	Technician	JRG
Asset No.	001997	Model	HH22		
Serial No.	T-94140	Type	Digital Thermometer	Cal Date	26-Apr-07

Function/Range	Test Point	TI Reading	Difference	+/-Limit	+/-Uncertainty	Found/Left
Type K	°F	°F	°F	°F	°F	Result
	-300	-300.6	0.6	1.3	0.70	Pass
	390	389.9	0.1	1.4	0.55	Pass
	1075	1075.7	0.7	2.1	0.55	Pass
	1760	1760.9	0.9	2.8	0.55	Pass
	2450	2451.3	1.3	3.5	0.84	Pass
Type K	°C	°C	°C	°C	°C	
	-150	-150.1	-0.1	0.8	0.40	Pass
	210	210.0	0.0	0.8	0.32	Pass
	575	575.4	0.4	1.2	0.32	Pass
	940	940.5	0.5	1.5	0.32	Pass
	1300	1300.8	0.8	1.9	0.47	Pass

END OF REPORT