NRC FORM 591M PART 1
(10-2003)
10 CFR 2.201

U.S. NUCLEAR REGULATORY COMMISSION

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED:	2. NRC/REGIONAL OFFICE
Saleem Husain, M.D. 1314 Park Avenue, Suite 9 Plainfield, New Jersey 07060 REPORT Nos 03035846/2007-001	**U.S. Nuclear Regulatory Commission** **Region I, 475 Allendale Road** **King of Prussia, Pennsylvania 19406-1415**

3. DOCKET NUMBER(S)	4. LICENSE NUMBER(S)	5. DATE(S) OF INSPECTION
03035846	29-30684-01	5/30/2007

LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:

☐ 1. Based on the inspection findings, no violations were identified.

☐ 2. Previous violation(s) closed.

☐ 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, NUREG-1600, to exercise discretion, were satisfied.

☐ Non-Cited Violation(s) was/were discussed involving the following requirement(s) and Corrective Action(s):

☒ 4. During this inspection certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.

10 CFR 20.1101(a) requires each licensee to develop, document and implement a radiation protection program commensurate with the scope and extent of licensed activities and sufficient to ensure compliance with this part.
The licensee committed, in an application dated October 10, 2001, to the Model ~~procedures~~ Rules for Safe Use of Radiopharmaceuticals in Regulatory Guide 10.8, Rev. 2. These rules required, in part, that laboratory coats or other protective clothing be worn when in areas where radioactive materials are used.
Contrary to the above, on May 30, 2007, the nuclear medicine tech used unsealed radioactive materials without a lab coat or other protective clothing. The tech indicated lab coats would be worn. The RSO stated he would periodically observe the tech.

Licensee's Statement of Corrective Actions for Item 4, above.

I hereby state that, within 30 days, the actions described by me to the inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

Title	Printed Name	Signature	Date
LICENSEE'S REPRESENTATIVE	SALEEM HUSAIN MD RSO		5/30/2007
NRC INSPECTOR	Steven R. Courtemanche	Steven R. Courtemanche	5/30/2007

SUNSI Review Completed By: /RA/ Steven R. Courtemanche ☒ Public ☒ Non-Sensitive