


## Telephone Record

Regional Cardiology Associates, P.C.  
21-32298-01, 030-35626, CN 316246, ML 071340227

Amendment request dated March 19, 2007, to remove a  
place of use.

The licensee performed a closeout survey of their Nerredia  
Street East facility which was a cardiology nuclear  
medicine clinic. The survey, did not include calibration  
dates on the survey equipment or counting equipment.

Also, leak test results were not included in the  
amendment. Laura Smith, the licensee's consultant, was  
contacted by the reviewer on 6/14/07, and was asked to  
provide this information which was faxed on June 14,  
2007, however, calibration dates were not included.  
Therefore, the reviewer called Laura Smith on 6/18/07 and  
learned that the survey instruments were calibrated on  
August 3, 2006, September 24, 2006, and the counting  
equipment efficiency was determined to be 76% for Tc-  
99m. According to Laura, the counting equipment  
internally determines the efficiency every time the system  
is calibrated. It now appears all the information is needed  
to finish this decommissioning request.

  
James Mullauer, M.H.S.