


Telephone Record

Regional Cardiology Associates, P.C.
21-32298-01, 030-35626, CN 316246, ML 071340227

Amendment request dated March 19, 2007, to remove a
place of use.

The licensee performed a closeout survey of their Nerredia
Street East facility which was a cardiology nuclear
medicine clinic. The survey, did not include calibration
dates on the survey equipment or counting equipment.

Also, leak test results were not included in the
amendment. Laura Smith, the licensee's consultant, was
contacted by the reviewer on 6/14/07, and was asked to
provide this information which was faxed on June 14,
2007, however, calibration dates were not included.
Therefore, the reviewer called Laura Smith on 6/18/07 and
learned that the survey instruments were calibrated on
August 3, 2006, September 24, 2006, and the counting
equipment efficiency was determined to be 76% for Tc-
99m. According to Laura, the counting equipment
internally determines the efficiency every time the system
is calibrated. It now appears all the information is needed
to finish this decommissioning request.


James Mullauer, M.H.S.