SOUTHWEST RESEARCH INSTITUTE®

6220 Culebra Road, P.O. Drawer 28510
Institute Quality Systems
Institute Calibration Laboratory
Phone: 210-522-5215 Fax 210-522-4834

Certificate of Calibration

ACCREDITED
Calibration Laboratory
Certificate #0972-01

Submitted By: DIV20
Address: B57
Contact: DON BANNON
Manufacturer / Model: DURO-SENSE / TYPE K
Description: THERMOCOUPLE
Serial No: 0003
Asset No: 011659
Procedure: TEMPERATURE PROBES - 5 JUNE, 2006

Work Order: 303074756
Date Issued: Jun 6, 2007
Calibration Date: Jun 6, 2007
***Calibration Due:** Dec 6, 2007
Calibration Location: Bldg. 64
Environment: Temp. 74.0°F Hum. 51 %RH
****Data Type:** FOUND-LEFT
DivID/Location: N/A

This certificate documents traceability to the National Institute of Standards and Technology (NIST) and the International System of Units (SI). The Laboratory quality system conforms to ISO/IEC 17025, 2005, ANSI/NCSL Z540-1-1994 and relevant requirements of the ISO 9000-2000 standard. This certificate shall not be reproduced, except in full, without the written approval of the Southwest Research Institute Calibration Laboratory. This certificate shall not be used to claim product endorsement by Southwest Research Institute, American Association for Laboratory Accreditation (A2LA) or any agency of the U. S. Government. Results of this calibration relate only to the instrument described above at the time of calibration and does not imply any long term stability of the instrument.

*Determined by the customer, does not imply the instrument will remain within tolerance as any number of factors may cause an out-of-tolerance condition before this date. **Found/Left = adjustment and/or repair was not required, As Left = adjusted and/or repaired was required. The client has sole responsibility for determination of in-/out-of-tolerance or compliance/noncompliance. See Remarks or attached Measurement Report with the same Work Order number for data.

Reported uncertainty calculated in accordance with the ISO "Guide to the Expression of Uncertainty in Measurement" (GUM) and represents an expanded uncertainty with a coverage factor of k=2 to approximate a 95% confidence level.

Remarks: Cal'd @ 0° and 150° C per customer request

Standards Used

Asset No.	Serial No.	Manufacturer	Model	Description	Cal Due
009137	A21208	HART SCIENTIFIC	1575	SUPER THERMOMETER	Jun 27, 07
010281	0421	HART SCIENTIFIC	5628	SPRT	Nov 30, 08
010814	A44625	HART SCIENTIFIC	1529	THERMOCOUPLE THERMOMETER	Oct 17, 07

Scott Kurt

Reviewed by: ( ) wgh (✓) srk ( ) jrg ( ) blt ( ) pwc
Metrology Technician

Measurements by: Joe Greagrey
Metrology Technician

Southwest Research Institute
Calibration Laboratory
Measurement Report

Work Order:	303074756	Mfr.	DURO-SENSE	Technician	JRG
Asset No:	011659	Model	Type K		
Serial No:	0003	Type	Thermocouple	Cal Date	6-Jun-07
Remarks:	Limits taken from ASTM E230-02 and are based on brand new unused thermocouples. Limited Cal - tested at 0.0 and 150 C				

Function/Range	Test Point	TI Read	Difference	+/-Limit	+/-Uncertainty	Found/Left
Temperature	° C	° C	° C	° C	° C	Result
	0.12	-0.08	0.20	2.2	0.19	Pass
	149.89	150.67	-0.78	2.2	0.30	Pass

END OF REPORT