



# SOUTHWEST RESEARCH INSTITUTE®

6220 Culebra Road, P.O. Drawer 28510  
Institute Quality Systems  
Institute Calibration Laboratory  
Phone: 210-522-5215 Fax 210-522-4834



Calibration Laboratory  
Certificate #0972-01

## Certificate of Calibration

**Submitted By:** DIV20

**Address:** B57

**Contact:** DON BANNON

**Manufacturer / Model:** TCC / S102C

**Description:** RESISTOR

**Serial No:** 2

**Asset No:** 011355

**Procedure:** DECADE RESISTOR TO 100 MOHM-STEP - 21 MAR 2006

**Work Order:** 303074758

**Date Issued:** Jun 6, 2007

**Calibration Date:** Jun 6, 2007

**\*Calibration Due:** Dec 6, 2007

**Calibration Location:** Bldg. 64

**Environment:** Temp. 68.0°F Hum. 42 %RH

**\*\*Data Type:** FOUND-LEFT

**DivID/Location:** N/A

This certificate documents traceability to the National Institute of Standards and Technology (NIST) and the International System of Units (SI). The Laboratory quality system conforms to ISO/IEC 17025, 2005, ANSI/NCSL Z540-1-1994 and relevant requirements of the ISO 9000-2000 standard. This certificate shall not be reproduced, except in full, without the written approval of the Southwest Research Institute Calibration Laboratory. This certificate shall not be used to claim product endorsement by Southwest Research Institute, American Association for Laboratory Accreditation (A2LA) or any agency of the U. S. Government. Results of this calibration relate only to the instrument described above at the time of calibration and does not imply any long term stability of the instrument.

\*Determined by the customer, does not imply the instrument will remain within tolerance as any number of factors may cause an out-of-tolerance condition before this date. \*\*Found/Left = adjustment and/or repair was not required, As Left = adjusted and/or repaired was required. The client has sole responsibility for determination of in-/out-of-tolerance or compliance/noncompliance. See Remarks or attached Measurement Report with the same Work Order number for data.

Reported uncertainty calculated in accordance with the ISO "Guide to the Expression of Uncertainty in Measurement" (GUM) and represents an expanded uncertainty with a coverage factor of k=2 to approximate a 95% confidence level.

**Remarks:** Unit value of 250.0124 ohms; Uncertainty of measurement is 0.0050 ohms

### Standards Used

Asset No.	Serial No.	Manufacturer	Model	Description	Cal Due
007001	2823A21362	HEWLETT-PACKARD	3458A/OPT 002	MULTIMETER	Feb 26, 08

*Scott Kuter*

Reviewed by: ( ) wgh (x) srk ( ) jrg ( ) blt ( ) pwc

Metrology Technician

m:\a2la1.rpt Rev date August 15, 2005

*Joe Greagrey*

Measurements by: Joe Greagrey

Metrology Technician

Southwest Research Institute  
 Calibration Laboratory  
 Measurement Report

Work Order:	303074758	Mfr:	TCC	Technician:	JRG
Asset No:	011355	Model:	S102C	Cal Date:	06-Jun-07
Serial No:	2	Type:	Resistor 250 Ohm		
Remarks:					

Function/Range	Test Point	TI Reading	Difference	+/-Limit	+/-Uncertainty	Found/Left
Resistance	Ohm 250.0000	Ohm 250.0124	Ohm 0.0124	Ohm 10	Ohm 0.0050	Result Pass
END OF REPORT						