



Office of Radiation Safety Services

June 28, 2007

United States Nuclear Regulatory Commission
475 Allendale Road
King of Prussia, PA 19406-1415

MS 16
P-6

Attn: Ms. Penny Lanzisera

**Reference: Radioactive Materials License No. 29-02957-13
Amendment to Include Ophthalmic Applicator Device
Docket Number: 030-09926
Mail Control Number: 140517**

Dear Ms. Lanzisera:

As per your e-mail dated June 22, 2007, the following additional information is submitted for your review regarding our recent amendment request to include Epi-Rad Device to our license number: 29-02957-13:

1. As indicated in the FDA documentation submitted that the device use is limited to research only.
2. We confirm that any future guidance developed by the NRC for use of this device will be reviewed and implemented by the UMDNJ-Radiation Safety Committee, as appropriate.
3. Since the use of this device is sufficiently different from previously used eye applicator services, we confirm that authorized users will meet the training and experience requirements in 10 CFR 35.490.
4. We confirm that the applicator device will be used in conjunction with the Multi-Channel Tester calibrated with the device prior to treatments and that calibrations will be conducted in accordance with the manufacturer's instructions.
5. We confirm that the applicator device will be returned to the storage container when not in use and that the container will be taken to the Office of Radiation Safety Services or Same Day Surgery in order to store the source expeditiously.


Docket Number: 030-09926
Mail Control Number: 140517
June 28, 2007
Page 2

6. We confirm that the applicator device will be returned to QSA Global no later than 1 year from the initial date of receipt (or no more than 25 uses after initial receipt) for service and maintenance.
7. We confirm that all Authorized Users, Physicists and the Radiation Safety Officer and his associates will receive the vendor training described in the application prior to use of the device.

We request that you execute an expedited review of this additional information provided for our license amendment request. Should you require any additional information, please contact me at (973) 972-5305.

Thank you for your cooperation in this regard.

Sincerely,



Venkata K. Lanka
Radiation Safety Officer

c: Brendan S. McCluskey, J.D., MPA
Roger W. Howell, Ph.D.
File