

.....

Northwestern Memorial
Hospital
251 East Huron Street
Chicago, Illinois

facsimile transmittal

To: JAMES MULLAUGH Fax: 630-829-9873

From: MIKE BANNER Date: 6/28/07

Re: _____ Pages: _____

CC: _____

Urgent For Review Please Comment Please Reply Please Recycle

Notes:

LICENS: 13-32585-01
CN 316264
ML 071420073

.....

**CARDIAC CARE CONSULTANTS P.C.
9116 COLUMBIA AVE
MUNSTER IN 46321**

ANURADHA DIVAKARUNI, M.D.

SHASHIDHAR DIVAKARUNI, M.D.

June 28, 2007

James R. Mullauer, MHS
Health Physicist
U.S. Nuclear Regulatory Commission, Region III
2443 Warrenville Road, Suite 210
Lisle, IL 60532


License No 13-32585-01
CN 316264
ML 071420073

Dear Mr. Mullauer;

Attached please find the latest leak test certificate for the sealed sources at Cardiac Care Consultants.

If further information is required, you can contact me at the contact number on the application or my fax number which is 312 926-0862.

Sincerely,



A. Michael Zimmer, Ph.D.
Radiation Safety Officer

Northwestern Memorial Hospital Leak Test Certificate

Name of Individual: Cardiac Care Consultants
Department:

Person Collecting Sample: A. M. Zimmer
Date Sample Collected: 11-Feb-07

Person Performing Analysis: A. M. Zimmer
Date Analysis Performed: 13-Feb-07
RSO Approval: *A. M. Zimmer* 2/13/2007

| Source (Nuclide) | Manufacturer | Model Number | Serial Number | Activity | Date | Wipe Test Results* |
|------------------|------------------------|--------------|---------------|-----------|----------|--------------------|
| Ca-137 | Northern Am Scientific | MED 3550 | 63898 | 237.6 uCi | 02/01/05 | NEG |
| Ba-133 | Northern Am Scientific | MED 3550 | 66518 | 281.1 uCi | 09/01/05 | NEG |
| Co-57 | Northern Am Scientific | MED 3550 | 66411 | 5.535 mCi | 07/01/05 | NEG |
| Co-57 | Northern Am Scientific | MED 372 | 66690 | 10 mCi | 09/11/05 | NEG |
| Co-57 | RadQual | BMO5-10 | BM05100277 | 10 mCi | 10/08/04 | NEG |
| | | | | | | |
| | | | | | | |
| | | | | | | |
| | | | | | | |
| | | | | | | |
| | | | | | | |

*NEGLIGIBLE-ACTIVITY IS LESS THAN 0.001 MICROCURIE