

From: Penny Lanzisera
To: dasika.rao@rwjuh.edu
Date: Fri, Jul 6, 2007 11:54 AM
Subject: Request for Additional Information for Amendment Request dated May 10, 2007

Licensee: Robert Wood Johnson University Hospital
License No. 29-10173-02
Docket No. 030-02525
Mail Control No. 140576

Dear Mr. Rao and Dr. Narra, please provide a description of the following regarding your new treatment room:

1. Warning systems and restricted area controls (e.g., locks on door and to secure device, signs, warning lights and alarms, interlock systems);
2. Area radiation monitoring equipment and an indication of where installed within the treatment room;
3. Viewing and intercom systems;
4. Steps that will be taken to ensure that no two units can be operated simultaneously, if other radiation-producing equipment (e.g., CT/simulator) are in the treatment room;
5. Methods to ensure that whenever the device is not in use or is unattended, the console keys will be inaccessible to unauthorized persons; and
6. Storage location for emergency response equipment.

Please reply to my attention. You may fax your response to 610-337-5269. Please refer to Mail Control No. 140576 in your response.

Penny Lanzisera
US NRC, Region 1

Mail Envelope Properties (468E65D2.22E : 17 : 55940)

Subject: Request for Additional Information for Amendment Request dated May 10, 2007

Creation Date: Fri, Jul 6, 2007 11:54 AM

From: Penny Lanzisera

Created By: PAN@nrc.gov

Recipients

rwjuh.edu

dasika.rao (dasika.rao@rwjuh.edu)

Post Office

Route

rwjuh.edu

Files

MESSAGE

Size

1966

Date & Time

Friday, July 6, 2007 11:54 AM

Options

Expiration Date: None

Priority: Standard

ReplyRequested: No

Return Notification: None

Concealed Subject: No

Security: Standard