



**LEWISTOWN
HOSPITAL**

RECEIVED
REGION 1

2007 JUL -6 AM 8:03

June 29, 2007

NMSB1

USNRC Region 1
Materials Licensing Section
475 Allendale Road
King of Prussia PA 19406-1415

03010676

RE: License No. 37-05453-03

To Whom It May Concern:

This letter is a request for an amendment to decommission the Cardio MD camera room on the fourth floor of the hospital. The room has not been used for the past two weeks. The camera is scheduled to be de-installed the week of July 9, 2007. The room will become an unrestricted area upon your approval. Attached are the area wipe tests and surveys performed on 6/29/07. I ask that the amendment be expedited so it can turned over for construction.

Thank you for your cooperation

Please contact Dwayne Eberle at (717) 242-7329 if you require any additional information.

Sincerely,

Kay Hamilton

Kay Hamilton RN, MSN, CNO
President/CEO

Enclosures

140768

NMSS/RGN1 MATERIALS-002

Wipe Tests for Cardio MD Room 6/29/07

Background=67 dpm
Cardio MD=22 dpm
Pegasys=31 dpm
Acquisition Computer=24 dpm
Trash=39 dpm
Computer=46 dpm
Desk=45 dpm
Floor=51 dpm
Chair=23 dpm

Surveys for Cardio MD Room 6/29/07

Background=0.03 mR/hr
Cardio MD=0.03 mR/hr
Pegasys=0.03 mR/hr
Acquisition Computer=0.03 mR/hr
Trash=0.03 mR/hr
Computer=0.03 mR/hr
Desk=0.03 mR/hr
Floor=0.03 mR/hr
Chair=0.03 mR/hr