



5475 Airport Blvd., Boulder, CO 80301
303-443-7100 1-800-236-1801 FAX: 303-448-6870
Customer Response Center: 1-877-475-3317
Instrument Service & Support: 1-800-557-6883
www.pmeasuring.com

Radioactive Source Leak Test Certificate

This certifies that the Ion Mobility Spectroscopy Cell listed below was found to have less than 0.005 microcuries (11,000 DPM) of removable radioactive material.

Date Tested:	May 11, 2006
Cell Serial Number:	1778
Isotope:	Ni-63
Analyzer S/N:	776 AN

Present Owner of Analyzer:	EAI Corporation / TBE
Location:	504 th Street, Bldg. 53-210 Pine Bluff Arsenal White Hall, AK 71612

Note: The next leak test is required six months after the date listed above.

Measuring Device Used:	LKB Wallac Liquid Scintillation Counter Model: 1219-003 S/N: 190489
-------------------------------	---

Signature



Bruce Jablonski
Radiation Safety Officer



5475 Airport Blvd., Boulder, CO 80301
303-449-7100 1-800-238-1801 FAX: 303-449-8870
Customer Response Center: 1-877-475-3317
Instrument Service & Support: 1-800-557-6363
www.pmeasuring.com

Radioactive Source Leak Test Certificate

This certifies that the Ion Mobility Spectroscopy Cell listed below was found to have less than 0.005 microcuries (11,000 DPM) of removable radioactive material.

Date Tested:	May 5, 2006
Cell Serial Number:	3007
Isotope:	Ni-63
Analyzer S/N:	1026 AN

Present Owner of Analyzer:	EAI Corporation / TBE
Location:	504 th Street, Bldg. 53-210 Pine Bluff Arsenal Whitehall, AR 71612

Note: The next leak test is required six months after the date listed above.

Measuring Device Used:	LKB Wallac Liquid Scintillation Counter Model: 1219-003 S/N: 190489
-------------------------------	---

Signature

Bruce Jablonski
Radiation Safety Officer