

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED: <i>Clinton Diagnostics</i> <i>43475 Salcoma Dr</i> <i>Clinton Township, MI 48038</i> REPORT <i>2007001</i>	2. NRC/REGIONAL OFFICE U.S. Nuclear Regulatory Commission Region III 2443 Warrenville Road Lisle, Illinois 60532-4351	
3. DOCKET NUMBER(S) <i>03036580</i>	4. LICENSEE NUMBER(S) <i>21-32514-01</i>	5. DATE(S) OF INSPECTION <i>July 20, 2007</i>

LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:

- 1. Based on the inspection findings, no violations were identified.
- 2. Previous violation(s) closed.
- 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, NUREG-1600, to exercise discretion, were satisfied.



_____ Non-Cited Violation(s) was/were discussed involving the following requirement(s) and Corrective Action(s):

- 4. During this inspection certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.

(Violations and Corrective Actions)

Licensee's Statement of Corrective Actions for Item 4, above.

I hereby state that, within 30 days, the actions described by me to the inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

Title	Printed Name	Signature	Date
LICENSEE'S REPRESENTATIVE			
NRC INSPECTOR	George Parker		<i>7/20/07</i>

MP

Docket File Information**SAFETY INSPECTION REPORT
AND COMPLIANCE INSPECTION**

1. LICENSEE Clinton Diagnostics REPORT NUMBER(S) 2007-001		2. NRC/REGIONAL OFFICE Region III 2443 Warrenville Road Lisle, IL 60532	
3. DOCKET NUMBER(S) 030-36580	4. LICENSE NUMBER(S) 21-32514-01	5. DATE(S) OF INSPECTION 7/20/07	
6. INSPECTION PROCEDURES USED 87130	7. INSPECTION FOCUS AREAS 03.01-03.07		

SUPPLEMENTAL INSPECTION INFORMATION

1. PROGRAM CODE(S) 2201	2. PRIORITY 5	3. LICENSEE CONTACT J Antonio Bouff	4. TELEPHONE NUMBER 586/226-4147
<input checked="" type="checkbox"/> Main Office Inspection		Next Inspection Date: <u>7/2008</u>	
<input type="checkbox"/> Field		_____	
<input type="checkbox"/> Temporary Job Site		_____	

PROGRAM SCOPE

License is for a small cardiology clinic located in Clinton Township, Michigan. This licensee has not began to use material as of the date of this inspection. Conversations with the Office Manager indicates they plan to start using material within the next six months.

This license should be scheduled for inspection one year from the current date.