



******Facsimile Request******
Date: May 16, 2007

Message For: Susan Weiss

Of: Consultants in Medical Physics

Facsimile Number: 651-464-8036

Number of Pages (including this form): 2

From
James R. Mullauer, M.H.S.
Health Physicist
United States Nuclear Regulatory Commission
2443 Warrenville Road
Lisle, IL 60532-4351

Current Telephone Number: (623) 214-5213 Fax Number: (630) 829-9873

E-mail: jrm1@nrc.gov

Dear Susan, Please call with any questions.

Thanks, Jim

A handwritten signature in cursive script that reads "James R. Mullauer".

In accordance with 10 CFR 2.390 of the NRC's "Rules of Practice," a copy of this letter and its enclosure will be available electronically for public inspection in the NRC Public Document Room or from the Publicly Available Records (PARS) component of NRC's document system (ADAMS). The NRC's document system is accessible from the NRC Web site at <http://www.nrc.gov/reading-rm/adams.html> (the Public Electronic Reading Room).

See Attached
SENT 5/16/07

Indiana Institute of Cardiology

13-32652-01, 030-37464, CN 316218, ML 071220222
New license application dated 4/26/07.

Initial Review Status

- ✓ 1. Please provide a letter signed by both managements (DigiRad Imaging Solutions and Indiana Institute of Cardiology (IIC)) that DigiRad releases the facility as is to IIC and IIC accepts full responsibility for the facility, i.e. any radioactive contamination that may be present and all sources of radiation, i.e. check sources, etc.
2. Please provide the following concerning your appointment of RSO for IIC.
 - ✓ a. How you will divide your time between your consultation duties and onsite availability at IIC to conduct the duties of RSO.
 - ✓ b. Please specify the amount of time it will take you to respond to an emergency involving radioactive materials when you are not present.
 - ✓ c. Please describe the mechanisms for reporting to management of unsafe practices and incidents and management role in responding to such circumstances.
 - ✓ d. Please state the minimum frequency that you will provide radiation safety reports to management.
 - ✓ e. Please list all other RSO positions you currently hold at NRC as well as Agreement States licensees facilities if any.
- ✓ 4. The checklist in your application references Appendix J in 10 CFR Part 35 for the training and experience of George J. Grevich, M.D. Please make the proper current reference to 10 CFR 35.290.
- ✓ 5. Please confirm that you reserve the right to upgrade survey instruments as necessary as long as they are adequate to measure the type and level of radiation for which they are used. Also, the checklist states that the efficiency test results are attached, however, no results were attached. (Side note. I am not sure why this statement is in the application?)
- ✓ 6. Preceptor form on Page 4 is not dated. Please date provide page 4.