U.S. NU	CLEAR REGULATORY COMMISSION	Amendment No. 41	
M	ATERIALS LICENSE		
Pursuant to the Atomic Energy Act of 1954, as amended, the Energy Reorganization Act of 1974 (Public Law 93-438), and Title 10, Code of Federal Regulations, Chapter I, Parts 30, 31, 32, 33, 34, 35, 36, 39, 40, and 70, and in reliance on statements and representations heretofore made by the licensee, a license is hereby issued authorizing the licensee to receive, acquire, possess, and transfer byproduct, source, and special nuclear material designated below; to use such material for the purpose(s) and at the place(s) designated below: to deliver or transfer such material to persons authorized to receive it in accordance with the regulations of the applicable Part(s). This license shall be deemed to contain the conditions specified in Section 183 of the Atomic Energy Act of 1954, as amended, and is subject to all applicable rules, regulations, and orders of the Nuclear Regulatory Commission now or hereafter in effect and to any conditions specified below.			
Licensee	In accordance	with letter dated	
	July 30, 2007,		
1. LaPorte Hospital and Health Services	3. License numbe	er 13-15151-01 is amended in	
Department of Nuclear Medicine	its entirety to re	its entirety to read as follows:	
2 1007 Lincolnway	4. Expiration date	March 31, 2014	
LaPorte IN 46350-0250	5. Docket No. 03	30-08653	
	Reference No.		
 Byproduct, source, and/or special nuclear material 	emical and/or physical form	 Maximum amount that licensee may possess at any one time under this license 	
A. Any byproduct material A. permitted by 10 CFR 35.100	Any	A. As needed	
B. Any byproduct material B. permitted by 10 CFR 35 .200	Any	B. As needed	
C. Any byproduct material C. permitted by 10 CFR 35.300	Any think 3	C. B curies	
D. Any byproduct materi a l D. permitted by 10 CFR 35.400	Sealed sources (Bard Bradhytherapy, Inc. Model	D. 2 curies	
	Theragenics Corp. Model No. 200 and Seed Model 125.S06		
E. Any byproduct material E. permitted by 10 CFR 35.500	Sealed sources identified in 10 GFR 35,500, (North American Scientific Model No. MED 3601 or DuPont Pharma Model No. NES 8412 and Isotope Products Model HEG-137)	E. 400 millicuries	
F. Depleted uranium F.	Stainless steel covered metal	F. 4 shields, not to exceed 12 kilograms each	

FOR	M 374A U.S. NUCLEAR REGULATORY	COMMISSION	PAGE 2 of 4 PAGES
			License Number 13-15151-01
MATERIALS LICENSE			Docket or Reference Number 030-08653
			Amendment No. 41
<u> </u>			<u>i</u>
Aut	norized Use:		
Α.	Any uptake, dilution and excretion study p	ermitted by	10 CFR 35.100.
B. Any imaging and localization study permitted by 10 CFR 35.200.			
C. Any diagnostic study or therapy procedure permitted by 10 CFR 35.300.			
D.	Any manual brachytherapy procedure per	mitted by 10) CFR 35.400.
E. Medical use described in 10 CFR 35.500 in devices which have been evaluated and approved for licensing purposes by the U.S. Nuclear Regulatory Commission or an Agreement State.			
F. Shielding in ADAC Laboratories MCD-AC attenuation correction system.			
Lice	ensed material may be used at the licensee	's facilities I	ocated at 1007 Lincolnway, LaPorte, Indiana.
Rac	liation Safety Officer: James C. Hatten.		
Lice	nsed material is only authorized for use by	Huuud Huuud	supervision of:
A.	Individuals permitted to work as an author	ized user in	accordance with 10 CFR 35.13 and 35.14.
Β.	The following individuals are autorized u	sers for me	icatuse as indicated:
	Authorized Users	Mat	erial and Use
	Hester Muller, M.D.	10 CFR 35	.192, 35.200, 35.300 and 35.500.
	Russell Johnson, M.D.	10 CFR 35	.100, 35.200, 35.300, 35.400 and 35.500.
	Smari Thordarson, M.D.	10 CFR 35	.100, 35.200, 35.300 and 35.500.
	Douglas J. Van Putten, M.D.	10 CFR 35	.400.
	John G. McGue, M.D.	10 CFR 35	.100, 35.200, 35.300 and 35.500.
	David A. Hornback, M.D.	10 CFR 35	.300, 35.400 and 35.500.
	Nina Fukunaga-Johnson, M.D.	10 CFR 35	.300, 35.400 and 35.500.
	Walter Edward Wojcicki, M.D.	10 CFR 35	.100 and 35.200.
	Aut A. B. C. D. E. F. Lice A. B.	Authorized Use: A. Any uptake, dilution and excretion study permit C. Any diagnostic study or therapy procedure D. Any manual brachytherapy procedure per E. Medical use described in 10 CFR 35.500 licensing purposes by the U.S. Nuclear Re F. Shielding in ADAC Laboratories MCD-AC C. Licensed material may be used at the licensee Radiation Safety Officer: James C. Hatten. Licensed material is only authorized for use per A. Individuals permitted to work as an author B. The following individuals are authorized use Authorized Users Hester Muller, M.D. Russell Johnson, M.D. Douglas J. Van Putten, M.D. John G. McGue, M.D. Nina Fukunaga-Johnson, M.D. Walter Edward Wojcicki, M.D.	INTREMENTARY SHEET MATERIALS LICENSE SUPPLEMENTARY SHEET Authorized Use: Any uptake, dilution and excretion study permitted by B. Any imaging and localization study permitted by 10 C C. C. Any diagnostic study or therapy procedure permitted by 10 C D. Any manual brachytherapy procedure permitted by 10 C E. Medical use described in 10 CFR 35.500 in devices w licensing purposes by the U.S. Nuclear Regulatory Cr F. Shielding in ADAC Laboratories MCD-AC attenuation CONDITIONS Licensed material may be used at the licensee's facilities I Radiation Safety Officer: James C. Hatten Licensed material is only authorized for use but A. Individuals permitted to work as an authorized user in B. The following individuals are authorized user in B. Douglas J. Van Putten, M.D. 10 CFR 35 Douglas J. V

NRC	FOR	M 374A	U.S. NUCLEAR REGULATORY	COMMISSION	PAGE 3 of 4 PAGES
				_,	License Number 13-15151-01
MATERIALS LICENSE SUPPLEMENTARY SHEET		Docket or Reference Number 030-08653			
					Amendment No. 41
					L
		Toby S. Kram	er, M.D.	10 CFR 35	j.400.
		John E. DePe	rsio, M.D.	10 CFR 35 administra than or eq	5.100, 35.200, and 35.300 (for iodine-131, oral ation of sodium iodide-131 in quantities less jual to 33 millicuries).
		Vivek Mishra, M.D. 10 CFR 35.100 and 35.200.		5.100 and 35.200.	
		Richard Dob	ben, M.D.	. 10 CFR 35.100, 35.200, iodine-131 for use under 35.300 and 35.500.	
13.	For	sealed source	s not associated with 10 CFI	R Part 35 us	se, the following conditions apply:
	A. Sealed sources shall be tested for leakage and/or contamination at intervals not to exceed the intervals specified in the certificate of registration issued by the U.S. Nuclear Regulatory Commission under 10 CFR 32.210 or under equivalent regulations of an Agreement State.				
	B. In the absence of a certificate from a transferor indicating that a leak test has been made within the intervals specified in the certificate of registration issued by the U.S. Nuclear Regulatory Commission under 10 CFR 32.210 or under equivalent regulations of an Agreement State, prior to the transfer, a sealed source received from another person shall not be put into use until tested and the test results received.				
	C. Sealed sources need not be tested if they contain not more than 100 microcuries of beta and/or gamma emitting material or not more than 10 microcuries of alpha-emitting material.				
	D. Sealed sources need not be tested if they are in storage and are not being used; however, when they are removed from storage for use of transferred to another person and have not been tested within the required leak test interval, they shall be tested before use or transfer. No sealed source shall be stored for a period of more than 10 years without being tested for leakage and/or contamination.				
	E. The leak test shall be capable of detecting the presence of 0.005 microcurie (185 becquerels) of radioactive material on the test sample. If the test reveals the presence of 0.005 microcurie (185 becquerels) or more of removable contamination, a report shall be filed with the U.S. Nuclear Regulatory Commission in accordance with 10 CFR 30.50(c)(2), and the source shall be removed immediately from service and decontaminated, repaired, or disposed of in accordance with Commission regulations.				
	F. Tests for leakage and/or contamination, including leak test sample collection and analysis, shall be performed by the licensee or by other persons specifically licensed by the U.S. Nuclear Regulatory Commission or an Agreement State to perform such services.				
14. Sealed sources containing licensed material shall not be opened or sources removed from source holders by the licensee.					

NRC FORM 374A U.S. NUCLEAR REGULATORY COMMISSION		M 374A U.S. NUCLEAR REGULATORY COMMISSION	PAGE 4 of 4 PAGES	
			License Number	
	MATERIALS LICENSE SUPPLEMENTARY SHEET		Docket or Reference Number 030-08653	
			Amendment No. 41	
15.	The U.S und	licensee shall conduct a physical inventory every six r . Nuclear Regulatory Commission, to account for all so er the license.	nonths, or at other intervals approved by the purces and/or devices received and possessed	
10.	10 CFR Part 71, "Packaging and Transportation of Radioactive Material."			
17.	17. Except as specifically provided otherwise in this license, the licensee shall conduct its program in accordance with the statements, representations, and procedures contained in the documents, including any enclosures, listed below. This license condition applies only to those procedures that are required to be submitted in accordance with the regulations. Additionally, this license condition does not limit the licensee's ability to make changes to the radiation protection program as provided for in 10 CFR 35.26. The U.S. Nuclear Regulatory Commission's regulations shall govern unless the statements, representations, and procedures in the licensee's application and correspondence are more restrictive than the regulations.			
	Α.	Application received March 5, 2004;		
	Β.	Letters dated June 9, 2005, October 14, 2005, Noven	nber 1, 2006 and December 27, 2006;	
	C.	Facsimile letters dated September 6, 2005, September	er 8, 2005; and	
	D.	Facsimile (of e-mail) received February 15, 2007.		
		FOR THE U	U.S. NUCLEAR REGULATORY COMMISSION	
Date	e	AUG 0 7 2007 James F Material Region	R. Mullauer, M.H.S. Is Licensing Branch	