



Cardiac, Thoracic and Peripheral Vascular Surgery
 Charles B. Beckman, M.D.
 Viswa B. Nathan, M.D.
General Vascular Surgery
 Walter Kwass, M.D.
Executive Director
 William E. Appicelli

Cardiology
 John P. Chandler, M.D.
 Ajoy Kapoor, M.D.
 Inku K. Lee, M.D.
 Robert F. Morrison, M.D.
 Arumbakam Purushotham, M.D.
 Edgardo P. Ragaza, M.D.
 Nathan L. Valin, M.D.
 Henry N. Ward, M.D.
 Donald S. Dock, M.D.
 (1927-1996)

August 17, 2007

H-ø

ATTN: LAT
 Additional Info for MC140554
 FAX #610-337-5269

06-30090-01

0303309

To Whom It May Concern:

Here is the leak tests requested for Dr. Nathan Valin.

If you need anything else, please do not hesitate to contact my office.

Sincerely,

Jamie Miller
 Nathan Valin, M.D.

140554

NMSS/RGN1 MATERIALS-002

Lakewood Center
 251 West Main Street
 Branford, CT 06405
 (203) 315-5300
 Fax (203) 315-5312

281 Seaside Avenue
 Milford, CT 06460
 (203) 783-5889
 Fax (203) 783-6665

Prince Professional Center
 46 Prince Street
 New Haven, CT 06519
 (203) 867-5300
 Fax (203) 867-5312

Spring Glen Medical Center
 Suite 220
 2200 Whitney Avenue
 Hamden, CT 06518
 (203) 287-3740
 Fax (203) 287-3744

55 Meriden Avenue
 Southington, CT 06489
 (860) 628-0391
 (860) 628-0703
 Fax (860) 276-1940

Broadway Square
 2 Broadway
 North Haven, CT 0647
 (203) 234-3761
 Fax (203) 234-3765

CONNECTICUT HEART GROUP, Cheshire Ofc

Year: 2007

Sealed Sources - Inventory Record & Leak Test Record

Qtr. Period	Source	Model	Serial #	Activity (uCi)	Location*	Activity (uCi) [^]	Leakage (uCi)	Method*
I.	Cs-137	SRV-137-200U	788-21-11	185.3	HL	180.6	< 0.005	AW
3.06	Co-57	SRV-057-5M	1145-34-22	5410	HL	2178	< 0.005	AW
2.06	Co-57 fld	SFL18C-057-10M	1153-097	10000	HL	3675	< 0.005	AW

Initials: P J M Date: 3/11/07 RSO: _____

Qtr. Period	Source	Model	Serial #	Activity (uCi) [^]	Location*
II.	Cs-137	SRV-137-200U	788-21-11	185.3	HL
3.06	Co-57	SRV-057-5M	1145-34-22	5410	HL
2.06	Co-57 fld	SFL18C-057-10M	1153-097	10000	HL

Initials: P J M Date: 6/25/07 RSO: _____

Qtr. Period	Source	Model	Serial #	Activity (uCi) [^]	Location*	Activity (uCi) [^]	Leakage (uCi)	Method*
III.	Cs-137	SRV-137-200U	788-21-11	185.3	HL		< 0.005	AW
3.06	Co-57	SRV-057-5M	1145-34-22	5410	HL		< 0.005	AW
2.06	Co-57 fld	SFL18C-057-10M	1153-097	10000	HL		< 0.005	AW

RSO: _____

Qtr. Period	Source	Model	Serial #	Activity (uCi) [^]	Location*
IV.	Cs-137	SRV-137-200U	788-21-11	185.3	HL
3.06	Co-57	SRV-057-5M	1145-34-22	5410	HL
2.06	Co-57 fld	SFL18C-057-10M	1153-097	10000	HL

RSO: _____

NOTE: Required for gamma emitters of > 100 uCi activity.

[^] Present estimated activity.

Leakage must be < 0.005 uCi (< 11,000 dpm).

* Abbreviations: HL-hot lab, IR-imaging room, GC-gamma camera, PC-PET camera, absorption, or, ID-immersion in detergents. LS if liquid scintillation is used.

Nal probe if: AW-alcohol wipe, CA-charcoal

CONNECTICUT HEART GROUP, PC

46 Prince Street, 3rd floor, New Haven, CT 06519

To: USNRC Region I
Commercial and R&D Branch
475 Allendale Road
King of Prussia, PA 19406

May 24, 2007

Re: Decommissioning of 366 South Main St., Route 10, Cheshire, CT site

Dear Sir or Miss:

We performed a final closeout survey at the office site of 366 South Main Street, Cheshire, CT. Sealed sources were removed and transferred to the new office of 1157 Highland Avenue Cheshire, CT 06410. Wastes were allowed to decay for greater than 10 half-lives and then removed as normal or biomedical trash, as appropriate.

The following data summarizes the final survey for ambient radiation (mR/hr) levels and removable contamination (dpm) from selected areas of the closed office. There were no areas demonstrating contamination and the final survey results were:

<u>Area</u>	<u>mR/hr</u>	<u>dpm</u>
Hot Lab (including waste storage)	0.03	< 200
Office Area	0.05	< 200
Stress Lab	0.04	< 200
Imaging Area	0.03	< 200
Hallway	0.03	< 200
Bathroom	0.03	< 200

The background radiation levels were: 0.03 - 0.04 mR/hr (ambient survey), 50 dpm (contamination wipe). The survey instruments used were: Ludlum Model 14C GM Survey Meter, and a Ludlum Model 2200 Scaler/Ratemeter and Shielded NaI crystal with a well.

Respectfully submitted,

William Appicelli, CEO