



Andrea D. Browne, Ph.D.
Radiation Safety Officer
Medical Imaging Dept.
Community Hospital - East
1500 N. Ritter Avenue
Indianapolis, IN 46219
(317) 355-5865

Community Hospitals Indianapolis

Community Hospital East

16th Street and Ritter Avenue
P.O. Box 19805
Indianapolis, Indiana 46219-0805
Telephone (317) 355-1411

September 25, 2001

U.S. Nuclear Regulatory Commission
Region III
801 Warrenville Road
Lisle, IL 60532-4351

Dear Sir or Madam:

I am reporting the possible loss of radioactive material, an I-125 seed of 0.327 mCi activity. An incorrect post-procedure count indicated that a loss may have occurred. Since no verification count was performed on the order prior to handling we are unable to substantiate if a loss did or did not occur. All surveys at post-handling steps were negative for radiation. A copy of the report compiled by the physicist performing the procedure is attached. In order to prevent future events the procedure for verifying and handling the seeds has been changed. A copy of the revised procedure is also attached. If any further information regarding this event becomes available I will forward it to you. If you have any questions, please contact me at the phone number listed above.

Sincerely,

Andrea D. Browne
Radiation Safety Officer



Printed on recycled paper.

B-18

SEP 28 2001