


Cumberland Cardiology, P.S.C.



613 - 23rd Street
Suite 230
P.O. Box 2380
Ashland, Kentucky 41105-2380
Phone: (606) 324-4745
Fax: (606) 324-4941

Richard E. Paulus, M.D., F.A.C.C.
John M. Van Deren, III, M.D., F.A.C.C.
Zane A. Darnell, M.D.
Terence C. Ross, M.D., F.A.C.C.
Richard A. Ansinelli, M.D., F.A.C.C.
Christopher P. Epling, D.O.
Robert C. Touchon, M.D., F.A.C.C.
Vaughn W. Payne, M.D., F.A.C.C.
Arley A. Peter, M.D.

J-6
47-31272-01
03037536

9/6/2007

Dear Sir,

Cumberland Cardiology has reviewed and approved the change in our license application for the Barboursville, West Virginia office. NRC Form 313. Item 9 Dose Calibrator Calibration.

Thank you,


Dave Legenza
Office Manager

140981

**CUMBERLAND CARDIOLOGY
6007 U. S. ROUTE 60 EAST - SUITE 308
BARBOURSVILLE, WEST VIRGINIA 25504**

NRC LICENSE APPLICATION

NRC FORM 313

Item 9 Radiation Monitoring Instruments

Survey Meters will be calibrated accordance with Ohio License No. 03223480006 for Universal Consultants, Inc., North Ridgeville, Ohio or by a person qualified to perform survey meter calibrations.

Item 9 Dose Calibrator Calibration

Equipment used to measure dosages will be calibrated in accordance with nationally recognized standards or the manufacturer's instructions.

Item 9 Occupational Dose

Either we will perform a prospective evaluation demonstrating that unmonitored individuals are not likely to receive, in one year, a radiation dose in excess of 10% of the allowable limits in 10 CFR Part 20 or we will provide dosimetry that meets the requirements listed under "Criteria" in NUREG-1556, Vol. 9, "Consolidated Guidance About Materials Licenses; Program-Specific Guidance About Medical Use Licensees," dated October, 2002.

Film badges/TLD will be obtained and exchanged on a monthly basis from R. S. Landauer, Inc.

Item 10 RADIATION PROTECTION PROGRAM

Item 10 Leak Testing

Leak test of sealed sources will be done in accordance the procedures in Appendix Q to NUREG-1556, Vol. 9 and with Ohio License #03223480006 for Universal Consultants, Inc., North Ridgeville, Ohio or other licensed service.

Item 10 Safe Use of Unsealed License Material

We have developed and will implement and maintain procedures for safe use of unsealed byproduct material that meet the requirements of Appendix T of NUREG-1556, Vol. 8.