

**From:** "Peterson, Steven" <Steven.Peterson@goodrich.com>  
**To:** <czg@nrc.gov>  
**Date:** Tue, Sep 4, 2007 12:17 PM  
**Subject:** Goodrich Danbury - NRC License 06-20581-01

Craig,

As we discussed last Wednesday during your inspection the Goodrich Facility at 100 Wooster Heights Road in Danbury, CT (NRC License 06-20581-01), I am sending you the leak test results conducted on May 30, 2007 for our Cd-109 source and the 2007 audit report of the radiation protection program. The leak test and the audit were conducted by Radiation Safety Associates (RSA) of Hebron, CT.

Based on your inspection last week and the RSA audit report, the following actions have been taken to date.

- 1 I'm registering for the RSO training course for November 12 -16, 2007 by RSA. Mike Toth will be the RSO until I have been trained. The license needs to be amended to reflect these changes and I will look to your recommendations as to whether we should amend it now for Mike and in Nov for me, or just me at this time (this issue is discussed finding #1 in the RSA report).
2. The dosimetry monitoring has been reinstated in July 2007 and I am working with Global Dosimetry (our vendor) to add names to the rings and to include additional employees who are new to the operations in the lab (this issue is discussed in finding #6 of the RSA report).
3. To ensure completion of the 6-month inventory, leak test and annual audit, these items have been added to an electronic calendar which is viewed monthly for actions to be taken in the upcoming 3 months. (These issues are discussed in findings #2 and #8 of the RSA report.)
4. RSA will be here in October to ship the old source back the to vendor (this issue is discussed in findings #3 and # 10 of the RSA report).
5. We are in the process of organizing the records (finding #12) and the old NRC Form 3 has been replaced with the revised 5-2005 edition (finding # 4).
6. Training will be provided in October by RSA to the employees working in the lab (finding # 5 in the RSA report).

You mentioned that you will call me prior to sending your letter to which we will officially respond. I will be traveling on Sept 11 - Sept 14 and will be difficult to reach during that time.

Thanks,

Steve

Steven W. Peterson, P.E.  
EHS Manager  
ISR Systems  
Goodrich Corporation  
100 Wooster Heights Road

Danbury, CT 06810-7589  
203-797-5571  
203-797-6499 (fax)

**Mail Envelope Properties** (46DD92EE.E73 : 17 : 56947)

**Subject:** Goodrich Danbury - NRC License 06-20581-01  
**Creation Date** Tue, Sep 4, 2007 12:16 PM  
**From:** "Peterson, Steven" <Steven.Peterson@goodrich.com>

**Created By:** Steven.Peterson@goodrich.com

**Recipients**

nrc.gov  
kp1\_po.KP\_DO  
CZG (Craig Gordon)

**Post Office**

kp1\_po.KP\_DO

**Route**

nrc.gov

<b>Files</b>	<b>Size</b>	<b>Date &amp; Time</b>
MESSAGE	2237	Tuesday, September 4, 2007 12:16 PM
TEXT.htm	5048	
Cd-109 05-2007 leak test.pdf	204353	
2007 Goodrich Audit Report.pdf		40498
Mime.822	345908	

**Options**

**Expiration Date:** None  
**Priority:** Standard  
**ReplyRequested:** No  
**Return Notification:** None

**Concealed Subject:** No  
**Security:** Standard

**Junk Mail Handling Evaluation Results**

Message is eligible for Junk Mail handling  
This message was not classified as Junk Mail

**Junk Mail settings when this message was delivered**

Junk Mail handling disabled by User  
Junk List is not enabled  
Junk Mail using personal address books is not enabled  
Block List is not enabled